


LB126687

SFAM01.1

/ INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9123e-004 slope

C= -1.0631e-001 y-intercept

Rho= 0.9998379

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/O N
0.00	0.000	-0.067	-0.067	203	0.000	203					4
0.05	0.050					504	0				0
0.20	0.200	0.207	0.007	1636	0.0 %	1636					3
2.50	2.500	2.496	-0.004	13608	0.0 %	13608					0
5.00	5.000	5.126	0.126	27360	0.0 %	27360					1
7.50	7.500	7.515	0.015	39856	0.0 %	39856					
10.0	10.000	9.924	-0.076	52450	0.0 %	52450					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	203	-		0 SFAM01.1	PPB	8/1/2023 15:22	Std
	0.2 S01	1636	-		0.2 SFAM01.1	PPB	8/1/2023 15:24	Std
	2.5 S02	13608	-		2.5 SFAM01.1	PPB	8/1/2023 15:27	Std
	5 S03	27360	-		5 SFAM01.1	PPB	8/1/2023 15:29	Std
	7.5 S04	39856	-		7.5 SFAM01.1	PPB	8/1/2023 15:31	Std
	10 S05	52450	-		10 SFAM01.1	PPB	8/1/2023 15:33	Std
ICV094	ICV094	20866	3.8839 -		SFAM01.1	PPB	8/1/2023 15:36	SMPL
ICB094	ICB094	170	-0.0738 -		SFAM01.1	PPB	8/1/2023 15:39	SMPL
CCV040	CCV040	27515	5.1554 -		SFAM01.1	PPB	8/1/2023 15:41	SMPL
CCB040	CCB040	159	-0.0759 -		SFAM01.1	PPB	8/1/2023 15:43	SMPL
PB154570BL	PBS570	43	-0.0981 -		SFAM01.1	PPB	8/1/2023 15:45	SMPL
O3725-01	E45F4	62768	11.8969 -		SFAM01.1	PPB	8/1/2023 15:48	SMPL
O3725-02	E45F5	144280	27.4846 -		SFAM01.1	PPB	8/1/2023 15:50	SMPL
O3725-03	E45F6	112109	21.3325 -		SFAM01.1	PPB	8/1/2023 15:52	SMPL
O3725-04	E45F7	9061	1.6264 -		SFAM01.1	PPB	8/1/2023 15:55	SMPL
O3725-05	E45F9	1646	0.2085 -		SFAM01.1	PPB	8/1/2023 15:57	SMPL
O3725-06	E45G1	949	0.0752 -		SFAM01.1	PPB	8/1/2023 15:59	SMPL
O3725-07	E45G1D	1181	0.1195 -		SFAM01.1	PPB	8/1/2023 16:01	SMPL
O3725-08	E45G1S	15152	2.7912 -		SFAM01.1	PPB	8/1/2023 16:04	SMPL
O3725-09	E45G2	929	0.0713 -		SFAM01.1	PPB	8/1/2023 16:06	SMPL
O3725-10	E45G3	1151	0.1138 -		SFAM01.1	PPB	8/1/2023 16:08	SMPL
O3725-11	E45G7	780	0.0429 -		SFAM01.1	PPB	8/1/2023 16:10	SMPL
O3725-12	E45G9	405	-0.0289 -		SFAM01.1	PPB	8/1/2023 16:13	SMPL
PB154571BL	PBS571	63	-0.0943 -		SFAM01.1	PPB	8/1/2023 16:15	SMPL
O3726-01	E45F8	17601	3.2596 -		SFAM01.1	PPB	8/1/2023 16:17	SMPL
O3726-02	E45G0	3585	0.5793 -		SFAM01.1	PPB	8/1/2023 16:19	SMPL
O3726-03	E45G4	3287	0.5223 -		SFAM01.1	PPB	8/1/2023 16:22	SMPL
O3726-04	E45G4D	3369	0.538 -		SFAM01.1	PPB	8/1/2023 16:24	SMPL
O3726-05	E45G4S	17506	3.2414 -		SFAM01.1	PPB	8/1/2023 16:26	SMPL
O3726-06	E45G5	3285	0.5219 -		SFAM01.1	PPB	8/1/2023 16:29	SMPL
CCV041	CCV041	27686	5.1881 -		SFAM01.1	PPB	8/1/2023 16:31	SMPL
CCB041	CCB041	167	-0.0744 -		SFAM01.1	PPB	8/1/2023 16:33	SMPL
O3726-07	E45G6	4093	0.6764 -		SFAM01.1	PPB	8/1/2023 16:35	SMPL
O3726-08	E45G8	1090	0.1021 -		SFAM01.1	PPB	8/1/2023 16:38	SMPL
O3726-09	E45H0	1696	0.218 -		SFAM01.1	PPB	8/1/2023 16:40	SMPL
PB154572BL	PBS572	311	-0.0468 -		SFAM01.1	PPB	8/1/2023 16:42	SMPL
O3728-01	MBH518	2474	0.3668 -		SFAM01.1	PPB	8/1/2023 16:44	SMPL
O3728-02	MBH520	330	-0.0432 -		SFAM01.1	PPB	8/1/2023 16:47	SMPL
O3728-03	MBH521	615	0.0113 -		SFAM01.1	PPB	8/1/2023 16:49	SMPL
O3728-04	MBH522	5522	0.9497 -		SFAM01.1	PPB	8/1/2023 16:51	SMPL
O3728-05	MBH523	373	-0.035 -		SFAM01.1	PPB	8/1/2023 16:53	SMPL
O3728-06	MBH524	370	-0.0356 -		SFAM01.1	PPB	8/1/2023 16:56	SMPL
O3728-07	MBH525	2344	0.3419 -		SFAM01.1	PPB	8/1/2023 16:58	SMPL
O3728-08	MBH526	2740	0.4177 -		SFAM01.1	PPB	8/1/2023 17:00	SMPL
O3728-09	MBH527	1903	0.2576 -		SFAM01.1	PPB	8/1/2023 17:03	SMPL
O3728-10	MBH528	1875	0.2523 -		SFAM01.1	PPB	8/1/2023 17:05	SMPL

LB126687

INSTRUMENT ID : CV1

O3728-11	MBH531	2260	0.3259 -	SFAM01.1	PPB	8/1/2023 17:07 SMPL
O3728-12	MBH532	1299	0.1421 -	SFAM01.1	PPB	8/1/2023 17:09 SMPL
O3728-13	MBH533	1661	0.2113 -	SFAM01.1	PPB	8/1/2023 17:12 SMPL
O3728-14	MBH535	3620	0.5859 -	SFAM01.1	PPB	8/1/2023 17:14 SMPL
O3728-15	MBH536	1746	0.2276 -	SFAM01.1	PPB	8/1/2023 17:16 SMPL
O3728-16	MBH537	1856	0.2486 -	SFAM01.1	PPB	8/1/2023 17:18 SMPL
CCV042	CCV042	27653	5.1818 -	SFAM01.1	PPB	8/1/2023 17:21 SMPL
CCB042	CCB042	171	-0.0736 -	SFAM01.1	PPB	8/1/2023 17:23 SMPL
O3728-17	MBH538	995	0.084 -	SFAM01.1	PPB	8/1/2023 17:25 SMPL
O3728-18	MBH539	245	-0.0595 -	SFAM01.1	PPB	8/1/2023 17:28 SMPL
O3728-19	MBH540	222	-0.0639 -	SFAM01.1	PPB	8/1/2023 17:30 SMPL
O3728-20	MBH548	1827	0.2431 -	SFAM01.1	PPB	8/1/2023 17:32 SMPL
O3728-21	MBH548D	1707	0.2201 -	SFAM01.1	PPB	8/1/2023 17:34 SMPL
O3728-22	MBH548S	16329	3.0163 -	SFAM01.1	PPB	8/1/2023 17:37 SMPL
CCV043	CCV043	27724	5.1954 -	SFAM01.1	PPB	8/1/2023 17:39 SMPL
CCB043	CCB043	195	-0.069 -	SFAM01.1	PPB	8/1/2023 17:42 SMPL
O3725-01DLX2		32790	6.1642 -	SFAM01.1	PPB	8/1/2023 17:44 SMPL
O3725-02DLX5		32576	6.1232 -	SFAM01.1	PPB	8/1/2023 17:47 SMPL
O3725-03DLX5		24697	4.6165 -	SFAM01.1	PPB	8/1/2023 17:49 SMPL
CCV044	CCV044	28046	5.257 -	SFAM01.1	PPB	8/1/2023 17:51 SMPL
CCB044	CCB044	176	-0.0727 -	SFAM01.1	PPB	8/1/2023 17:53 SMPL