

Clear table

Instrument ID: IC-2

Analyst: IZ

Method: 300.0 / 9056A

ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-18_	7/18/23 11:54		IZ/MA
STD2	0.5190	0.6440	0.6590	2.1930	0.5650	1.2310	3.4630	2023-07-18_	7/18/23 12:25		IZ/MA
STD3	0.7210	1.2220	1.2040	3.9830	1.0030	1.9630	6.0460	2023-07-18_	7/18/23 12:55		IZ/MA
STD4	1.0180	1.5110	1.5050	5.1180	1.2620	2.5080	7.5420	2023-07-18_	7/18/23 13:26		IZ/MA
STD5	1.9840	2.9150	2.9260	9.6980	2.4210	4.8360	14.4150	2023-07-18_	7/18/23 13:57		IZ/MA
STD6	3.7660	5.9300	5.9290	19.6870	4.9030	9.6670	29.4280	2023-07-18_	7/18/23 14:27		IZ/MA
STD7	5.1680	7.5780	7.5770	25.3210	6.3470	12.7960	38.1050	2023-07-18_	7/18/23 14:58		IZ/MA
ICV	2.0250	2.9710	2.9980	9.8960	2.5030	4.9290	14.7030	2023-07-18_	7/18/23 15:29		IZ/MA
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-18_	7/18/23 16:00		IZ/MA
CCV	1.9300	3.1160	2.9580	10.0770	2.4930	4.6950	14.8920	2023-07-31_	7/31/23 10:12		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-31_	7/31/23 10:43		IZ/MA
LB126690BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-31_	7/31/23 11:14		IZ/MA
LB126690BSW	1.9160	3.0340	2.9140	10.0510	2.4800	4.9270	14.8110	2023-07-31_	7/31/23 11:45		IZ/MA
O3572-09	0.2420	0.4350	0.3860	1.3380	0.3460	0.8570	2.1870	2023-07-31_	7/31/23 12:22		IZ/MA
O3572-09RE 0.25ML	0.1750	0.2950	0.2640	0.9240	0.2410	0.7320	1.5690	2023-07-31_	7/31/23 12:53		IZ/MA
CCV	1.9380	3.1900	2.9890	10.2210	2.5450	5.0860	15.1850	2023-07-31_	7/31/23 13:25		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-31_	7/31/23 13:55		IZ/MA

Report date: 8/1/2023 4:02:12 PM
Printed by: wet


Ident: CCV
Analysis from: 7/31/2023 10:12:01 AM
File: _2023-07-31_

Last save: 7/31/2023 10:41:13 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34999

Last save: 7/19/2023 10:12

SAMPLE:
: IZ/MA
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.65	1.198	4.93	2.08	396.860	6.06
2	9.80	0.228	30.07	12.72	461.721	7.05
3	11.88	0.278	52.21	22.09	986.403	15.07
4	15.55	0.331	29.86	12.63	661.643	10.11
5	17.87	0.403	34.78	14.71	944.555	14.43
6	22.67	0.584	29.65	12.54	1361.792	20.81
7	26.37	0.478	54.87	23.21	1732.156	26.46
7	28.00	0.500	236.36	100.00	6545.131	100.00

This report has been created by IC Net
METROHM LTD


Report date: 8/1/2023 4:02:19 PM
Printed by: wet
Ident: CCB
Analysis from: 7/31/2023 10:43:58 AM
File: _2023-07-31_

Last save: 7/31/2023 11:11:59 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35000

Last save: 7/19/2023 10:12

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD


Report date: 8/1/2023 4:02:33 PM
Printed by: wet
Ident: LB126690BLW
Analysis from: 7/31/2023 11:14:37 AM
File: _2023-07-31_

Last save: 8/1/2023 4:02:06 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35001

Last save: 7/19/2023 10:12

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD


Report date: 8/1/2023 4:02:42 PM
 Printed by: wet
 Ident: LB126690BSW
 Analysis from: 7/31/2023 11:45:15 AM
 File: _2023-07-31_

Last save: 8/1/2023 4:02:06 PM

Method: AnionsIC2-071823.mtw
 Run operator: wet
 Analysis number: 35002

Last save: 7/19/2023 10:12

SAMPLE:
 : IZ/MA
 Vial number: 14
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.58	1.130	5.18	2.13	393.769	5.99
2	9.64	0.224	29.56	12.15	449.102	6.83
3	11.66	0.270	53.67	22.07	971.111	14.77
4	15.21	0.319	31.00	12.75	659.841	10.04
5	17.44	0.386	36.11	14.85	939.457	14.29
6	22.54	0.569	32.15	13.22	1439.222	21.89
7	26.22	0.469	55.55	22.84	1722.213	26.19
7	28.00	0.481	243.21	100.00	6574.715	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 8/1/2023 4:02:49 PM
Printed by: wet


Ident: 03572-09
Analysis from: 7/31/2023 12:22:26 PM
File: _2023-07-31_

Last save: 8/1/2023 3:55:19 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35003

Last save: 7/19/2023 10:12

SAMPLE:
: IZ/MA
Vial number: 20
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.45	0.979	0.33	1.58	19.320	3.56
2	9.56	0.224	2.85	13.60	47.354	8.71
3	11.57	0.264	4.91	23.45	86.338	15.89
4	15.10	0.320	2.67	12.78	55.844	10.28
5	17.38	0.390	3.23	15.44	84.160	15.49
6	22.28	0.608	1.82	8.71	77.502	14.26
7	25.94	0.503	5.10	24.38	172.850	31.81
7	28.00	0.470	20.91	99.95	543.367	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/1/2023 4:02:57 PM
Printed by: wet


Ident: O3572-09RE 0.25ML
Analysis from: 7/31/2023 12:53:04 PM
File: _2023-07-31_

Last save: 8/1/2023 3:53:37 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35004

Last save: 7/19/2023 10:12

SAMPLE:
: IZ/MA
Vial number: 21
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.41	0.500	0.12	1.12	4.298	1.56
2	9.52	0.226	1.51	14.30	25.637	9.32
3	11.54	0.266	2.46	23.35	43.742	15.89
4	15.07	0.320	1.31	12.43	27.158	9.87
5	17.35	0.389	1.62	15.35	41.912	15.23
6	22.25	0.565	0.78	7.35	35.438	12.88
7	25.94	0.525	2.74	25.95	97.028	35.26
7	28.00	0.399	10.53	99.85	275.212	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/1/2023 4:03:04 PM
Printed by: wet


Ident: CCV
Analysis from: 7/31/2023 1:25:05 PM
File: _2023-07-31_

Last save: 7/31/2023 1:53:06 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35005

Last save: 7/19/2023 10:12

SAMPLE:
: IZ/MA
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.53	1.107	5.35	2.10	398.823	5.89
2	9.53	0.220	31.79	12.51	473.233	6.99
3	11.51	0.267	55.54	21.85	997.245	14.74
4	14.98	0.310	32.49	12.78	671.573	9.92
5	17.14	0.376	38.12	15.00	965.325	14.27
6	22.32	0.567	33.66	13.24	1492.748	22.06
7	25.96	0.467	57.24	22.52	1768.103	26.13
7	28.00	0.474	254.17	100.00	6767.050	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/1/2023 4:03:12 PM
Printed by: wet


Ident: CCB
Analysis from: 7/31/2023 1:55:46 PM
File: _2023-07-31_

Last save: 7/31/2023 2:23:46 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35006

Last save: 7/19/2023 10:12

SAMPLE:
:
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD