



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126691

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	08/01/2023	14:15
2	O3843-01		1	12	100.00	>212.0	>212.0	08/01/2023	14:45
3	O3843-01DUP		1	11	100.00	>212.0	>212.0	08/01/2023	15:15
4	O3855-03		1	14	100.00	>212.0	>212.0	08/01/2023	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB126691

WorkList Name : fp-7-31 WorkList ID : 172323 Department : Wet-Chemistry Date : 07-31-2023 08:04:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3843-01E	NWB-1935	Water	Flash Point	Cool 4 deg C	PSEG03	L11	07/31/2023	1010B
O3855-03C	34607-08	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/01/2023	1010B

Date/Time 08/01/2023 14:00
 Raw Sample Received by: RM CP Sin
 Raw Sample Relinquished by: RM CP Sin

Date/Time 08/01/2023 16:30
 Raw Sample Received by: CP Sin
 Raw Sample Relinquished by: RM CP Sin