



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126695

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	08/02/2023	08:20
2	CAL2	1	Water	NA	NA	20.4	7.00	08/02/2023	08:21
3	CAL3	1	Water	NA	NA	20.4	10.02	08/02/2023	08:22
4	ICV	1	Water	NA	NA	20.7	7.02	08/02/2023	08:25
5	CCV1	1	Water	NA	NA	20.5	2.01	08/02/2023	08:30
6	O3845-01	1	Water	NA	NA	23.2	5.02	08/02/2023	08:35
7	O3845-03	1	Water	NA	NA	22.4	5.34	08/02/2023	08:40
8	O3845-04	1	Water	NA	NA	20.1	3.37	08/02/2023	08:44
9	O3855-01	1	Water	NA	NA	20.4	3.49	08/02/2023	08:47
10	O3855-02	1	Water	NA	NA	20.3	3.62	08/02/2023	08:50
11	O3855-02DUP	1	Water	NA	NA	20.4	3.63	08/02/2023	08:51
12	CCV2	1	Water	NA	NA	20.7	12.02	08/02/2023	09:00

VB 21665

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 03845 WorkList ID : 172392 Department : Wet-Chemistry Date : 08-02-2023 08:09:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3845-01	TH-430-DI-1	Water	pH	Cool 4 deg C	NYUL01	L23	07/28/2023	9040C
O3845-03	TH-430-DI-3	Water	pH	Cool 4 deg C	NYUL01	L23	07/28/2023	9040C
O3845-04	TH-430-DI-4	Water	pH	Cool 4 deg C	NYUL01	L23	07/28/2023	9040C
O3855-01	34607	Water	pH	Cool 4 deg C	PSEG03	L11	08/01/2023	9040C
O3855-02	34608	Water	pH	Cool 4 deg C	PSEG03	L11	08/01/2023	9040C

Date/Time 08/02/23 08:15
Raw Sample Received by: JDLWC
Raw Sample Relinquished by: CP SM

Date/Time 08/02/23 10:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JDLWC