


LB126696

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2529e-004

C= 3.7137e-002

Rho= 0.9998400

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/11
0.0	0.000	0.095	0.095	259	0.000	259	0				1
0.2	0.200	0.212	0.012	774	0.0 %	774					6
2.5	2.500	2.384	-0.116	10415	0.0 %	10415					5
5.0	5.000	4.957	-0.043	21838	0.0 %	21838					1
7.5	7.500	7.519	0.019	33208	0.0 %	33208					0
10.0	10.000	10.034	0.034	44372	0.0 %	44372					0

LB126696 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Aug 2023 10:11:43

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	259	-	0.0000	7470A	PPB	02 Aug 2023 10:24:05	S	Std
0.2 - 1	50.2	774	-	0.2000	7470A	PPB	02 Aug 2023 10:26:22	S	Std
2.5 - 1	52.5	10415	-	2.5000	7470A	PPB	02 Aug 2023 10:28:38	S	Std
5.0 - 1	55	21838	-	5.0000	7470A	PPB	02 Aug 2023 10:30:55	S	Std
7.5 - 1	57.5	33208	-	7.5000	7470A	PPB	02 Aug 2023 10:35:52	S	Std
10.0 - 1	510	44372	-	10.0000	7470A	PPB	02 Aug 2023 10:38:08	S	Std
ICV95 - 1	ICV95	16645	3.7871	-	7470A	PPB	02 Aug 2023 10:41:13	U	SMPL
ICB95 - 1	ICB95	133	0.0671	-	7470A	PPB	02 Aug 2023 10:43:27	U	SMPL
CCV44 - 1	CCV44	21385	4.8550	-	7470A	PPB	02 Aug 2023 10:45:42	U	SMPL
CCB44 - 1	CCB44	147	0.0703	-	7470A	PPB	02 Aug 2023 10:47:57	U	SMPL
CRA - 1	CRA	675	0.1892	-	7470A	PPB	02 Aug 2023 10:50:12	U	SMPL
HighStd - 1	HighStd	44184	9.9914	-	7470A	PPB	02 Aug 2023 10:52:27	U	SMPL
ChkStd - 1	ChkStd	30909	7.0007	-	7470A	PPB	02 Aug 2023 10:54:42	U	SMPL
PB154565BL - 1	PBW	68	0.0525	-	7470A	PPB	02 Aug 2023 10:56:58	U	SMPL
PB154565BS - 1	LCSW	16846	3.8324	-	7470A	PPB	02 Aug 2023 10:59:14	U	SMPL
O3833-03 - 1	82ND-1	95	0.0585	-	7470A	PPB	02 Aug 2023 11:01:30	U	SMPL
O3849-02 - 1	GROVE-SOIL-0731	155	0.0721	-	7470A	PPB	02 Aug 2023 11:03:46	U	SMPL
O3849-02DUP - 1	GROVE-SOIL-0731DUP	160	0.0732	-	7470A	PPB	02 Aug 2023 11:06:03	U	SMPL
O3849-02MS - 1	GROVE-SOIL-0731MS	16347	3.7200	-	7470A	PPB	02 Aug 2023 11:08:26	U	SMPL
O3849-02MSD - 1	GROVE-SOIL-0731MSD	16013	3.6447	-	7470A	PPB	02 Aug 2023 11:10:42	U	SMPL
CCV45 - 1	CCV45	22162	5.0300	-	7470A	PPB	02 Aug 2023 11:12:57	U	SMPL
CCB45 - 1	CCB45	128	0.0660	-	7470A	PPB	02 Aug 2023 11:15:12	U	SMPL
PB154573BL - 1	PBW	56	0.0498	-	7470A	PPB	02 Aug 2023 11:17:27	U	SMPL
PB154573BS - 1	LCSW	18332	4.1672	-	7470A	PPB	02 Aug 2023 11:19:42	U	SMPL
O3806-02 - 1	ORA-1695	324	0.1101	-	7470A	PPB	02 Aug 2023 11:21:57	U	SMPL
O3806-02DUP - 1	ORA-1695DUP	202	0.0826	-	7470A	PPB	02 Aug 2023 11:24:12	U	SMPL
O3806-02MS - 1	ORA-1695MS	12033	2.7481	-	7470A	PPB	02 Aug 2023 11:29:00	U	SMPL
O3806-02MSD - 1	ORA-1695MSD	14424	3.2867	-	7470A	PPB	02 Aug 2023 11:31:16	U	SMPL
O3806-03 - 1	ORA-1693	2085	0.5069	-	7470A	PPB	02 Aug 2023 11:36:11	U	SMPL
O3808-01 - 1	1709	164	0.0741	-	7470A	PPB	02 Aug 2023 11:38:27	U	SMPL
O3808-09 - 1	1710-1715	141	0.0689	-	7470A	PPB	02 Aug 2023 11:40:45	U	SMPL
O3855-03 - 1	34607-08	496	0.1489	-	7470A	PPB	02 Aug 2023 11:43:00	U	SMPL
CCV46 - 1	CCV46	21885	4.9676	-	7470A	PPB	02 Aug 2023 11:45:16	U	SMPL
CCB46 - 1	CCB46	194	0.0808	-	7470A	PPB	02 Aug 2023 11:47:31	U	SMPL
PB154533TB - 1	PB154533TB	163	0.0739	-	7470A	PPB	02 Aug 2023 11:49:47	U	SMPL
O3849-02LX5 - 1		8	0.0389	-	7470A	PPB	02 Aug 2023 11:52:02	U	SMPL
O3849-02A - 1		18120	4.1194	-	7470A	PPB	02 Aug 2023 11:54:17	U	SMPL
O3806-02LX5 - 1		41	0.0464	-	7470A	PPB	02 Aug 2023 11:56:33	U	SMPL
O3806-02A - 1		19655	4.4652	-	7470A	PPB	02 Aug 2023 11:58:48	U	SMPL
CCV47 - 1	CCV47	22646	5.1391	-	7470A	PPB	02 Aug 2023 12:01:04	U	SMPL
CCB47 - 1	CCB47	1	0.0374	-	7470A	PPB	02 Aug 2023 12:03:22	U	SMPL