

LB126699


7471B

INSTRUMENT 10:CV1

Linear

 μ Abs.:

67104



A= 0.0000e+000

B= 1.5095e-004

C= -4.4415e-003

Rho= 0.9997447

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.025	0.025	198	0.000	198					0/61)
0.2	0.200	0.223	0.023	1509	0.0 %	1509					12
2.5	2.500	2.524	0.024	16748	0.0 %	16748					1
5.0	5.000	4.950	-0.050	32820	0.0 %	32820					1
7.5	7.500	7.352	-0.148	48736	0.0 %	48736					2
10.0	10.000	10.125	0.125	67104	0.0 %	67104					1

LB126699 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Aug 2023 13:20:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	198	-	0.0000	7471B	PPB	02 Aug 2023 13:22:26	S	Std
0.2 - 1	50.2	1509	-	0.2000	7471B	PPB	02 Aug 2023 13:24:43	S	Std
2.5 - 1	52.5	16748	-	2.5000	7471B	PPB	02 Aug 2023 13:27:01	S	Std
5.0 - 1	55	32820	-	5.0000	7471B	PPB	02 Aug 2023 13:29:17	S	Std
7.5 - 1	57.5	48736	-	7.5000	7471B	PPB	02 Aug 2023 13:31:33	S	Std
10.0 - 1	511	67104	-	10.0000	7471B	PPB	02 Aug 2023 13:36:22	S	Std
ICV97 - 1	ICV97	26693	4.0250	-	7471B	PPB	02 Aug 2023 13:39:11	U	SMPL
ICB97 - 1	ICB97	110	0.0122	-	7471B	PPB	02 Aug 2023 13:41:26	U	SMPL
CCV50 - 1	CCV50	32825	4.9506	-	7471B	PPB	02 Aug 2023 13:43:40	U	SMPL
CCB50 - 1	CCB50	116	0.0131	-	7471B	PPB	02 Aug 2023 13:45:56	U	SMPL
CRA - 1	CRA	1498	0.2217	-	7471B	PPB	02 Aug 2023 13:48:10	U	SMPL
HighStd - 1	HighStd	67521	10.1881	-	7471B	PPB	02 Aug 2023 13:50:31	U	SMPL
ChkStd - 1	ChkStd	45245	6.8255	-	7471B	PPB	02 Aug 2023 13:52:47	U	SMPL
PB154595BL - 1	PBS	187	0.0238	-	7471B	PPB	02 Aug 2023 13:55:04	U	SMPL
PB154595BS - 1	LCSS	22874	3.4485	-	7471B	PPB	02 Aug 2023 13:57:19	U	SMPL
O3840-01 - 1	BUILDING-B-13-(0-5)	1136	0.1670	-	7471B	PPB	02 Aug 2023 13:59:36	U	SMPL
O3840-02 - 1	BUILDING-B-5-(5-10)	842	0.1227	-	7471B	PPB	02 Aug 2023 14:01:53	U	SMPL
O3840-03 - 1	BUILDING-B-19-(0-5)	33964	5.1226	-	7471B	PPB	02 Aug 2023 14:04:11	U	SMPL
O3842-02 - 1	342	4379	0.6566	-	7471B	PPB	02 Aug 2023 14:06:28	U	SMPL
O3842-04 - 1	350	25276	3.8111	-	7471B	PPB	02 Aug 2023 14:08:43	U	SMPL
CCV51 - 1	CCV51	33013	4.9790	-	7471B	PPB	02 Aug 2023 14:10:58	U	SMPL
CCB51 - 1	CCB51	159	0.0196	-	7471B	PPB	02 Aug 2023 14:13:13	U	SMPL
O3842-04DUP - 1	350DUP	24389	3.6772	-	7471B	PPB	02 Aug 2023 14:15:30	U	SMPL
O3842-04MS - 1	350MS	47342	7.1420	-	7471B	PPB	02 Aug 2023 14:17:45	U	SMPL
O3842-04MSD - 1	350MSD	49773	7.5090	-	7471B	PPB	02 Aug 2023 14:20:01	U	SMPL
O3848-01 - 1	SU-02-073123	1621	0.2403	-	7471B	PPB	02 Aug 2023 14:22:16	U	SMPL
O3849-01 - 1	GROVE-SOIL-0731	16794	2.5307	-	7471B	PPB	02 Aug 2023 14:24:32	U	SMPL
O3842-04LX5 - 1		5417	0.8133	-	7471B	PPB	02 Aug 2023 14:26:50	U	SMPL
O3842-04A - 1		49684	7.4955	-	7471B	PPB	02 Aug 2023 14:29:06	U	SMPL
CCV52 - 1	CCV52	33361	5.0315	-	7471B	PPB	02 Aug 2023 14:31:25	U	SMPL
CCB52 - 1	CCB52	153	0.0187	-	7471B	PPB	02 Aug 2023 14:33:42	U	SMPL