

LB126700

SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51631

A= 0.0000e+000

B= 1.9143e-004 slope

C= -5.6432e-002 $\neq$ intercept

Rho= 0.9993172

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% of 1
0.00	0.000	-0.035	-0.035	113	0.000	113					=
0.05	0.050					504	0				-17
0.20	0.200	0.167	-0.033	1167	0.0 %	1167					✓ 3
2.50	2.500	2.584	0.084	13791	0.0 %	13791					-2 3
5.00	5.000	4.905	-0.095	25921	0.0 %	25921					-2 3
7.50	7.500	7.752	0.252	40790	0.0 %	40790					3 81112
10.0	10.000	9.827	-0.173	51631	0.0 %	51631					-2

LB126700
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	113	-		0 SFAM01.1	PPB	8/2/2023 15:23	Std
	0.2 S01	1167	-		0.2 SFAM01.1	PPB	8/2/2023 15:25	Std
	2.5 S02	13791	-		2.5 SFAM01.1	PPB	8/2/2023 15:28	Std
	5 S03	25921	-		5 SFAM01.1	PPB	8/2/2023 15:30	Std
	7.5 S04	40790	-		7.5 SFAM01.1	PPB	8/2/2023 15:35	Std
	10 S05	51631	-		10 SFAM01.1	PPB	8/2/2023 15:37	Std
ICV098	ICV098	22725	4.2937 -		SFAM01.1	PPB	8/2/2023 15:40	SMPL
ICB098	ICB908	49	-0.0471 -		SFAM01.1	PPB	8/2/2023 15:42	SMPL
CCV053	CCV053	25940	4.9091 -		SFAM01.1	PPB	8/2/2023 15:44	SMPL
CCB053	CCB053	48	-0.0472 -		SFAM01.1	PPB	8/2/2023 15:47	SMPL
PB154599BL	PBW599	83	-0.0405 -		SFAM01.1	PPB	8/2/2023 15:49	SMPL
O3727-01	MBH519	58	-0.0453 -		SFAM01.1	PPB	8/2/2023 15:51	SMPL
O3727-02	MBH519D	79	-0.0413 -		SFAM01.1	PPB	8/2/2023 15:53	SMPL
O3727-03	MBH519S	5960	1.0845 -		SFAM01.1	PPB	8/2/2023 15:56	SMPL
O3727-04	MBH529	24	-0.0518 -		SFAM01.1	PPB	8/2/2023 15:58	SMPL
O3727-05	MBH530	43	-0.0482 -		SFAM01.1	PPB	8/2/2023 16:00	SMPL
O3727-06	MBH534	54	-0.0461 -		SFAM01.1	PPB	8/2/2023 16:02	SMPL
O3727-07	MBH542	56	-0.0457 -		SFAM01.1	PPB	8/2/2023 16:05	SMPL
O3727-08	MBH543	54	-0.0461 -		SFAM01.1	PPB	8/2/2023 16:07	SMPL
O3727-09	MBH544	58	-0.0453 -		SFAM01.1	PPB	8/2/2023 16:09	SMPL
O3727-10	MBH552	78	-0.0415 -		SFAM01.1	PPB	8/2/2023 16:11	SMPL
O3727-11	MBH553	26	-0.0515 -		SFAM01.1	PPB	8/2/2023 16:14	SMPL
O3727-12	MBH554	47	-0.0474 -		SFAM01.1	PPB	8/2/2023 16:16	SMPL
O3727-13	MBH555	58	-0.0453 -		SFAM01.1	PPB	8/2/2023 16:18	SMPL
O3727-14	MBH556	63	-0.0444 -		SFAM01.1	PPB	8/2/2023 16:20	SMPL
O3727-15	MBH557	69	-0.0432 -		SFAM01.1	PPB	8/2/2023 16:23	SMPL
O3727-16	MBH558	89	-0.0394 -		SFAM01.1	PPB	8/2/2023 16:25	SMPL
O3727-17	MBH559	223	-0.0137 -		SFAM01.1	PPB	8/2/2023 16:27	SMPL
O3727-18	MBH560	103	-0.0367 -		SFAM01.1	PPB	8/2/2023 16:29	SMPL
PB154600BL	PBW600	91	-0.039 -		SFAM01.1	PPB	8/2/2023 16:32	SMPL
CCV054	CCV054	25928	4.9068 -		SFAM01.1	PPB	8/2/2023 16:34	SMPL
CCB054	CCB054	74	-0.0423 -		SFAM01.1	PPB	8/2/2023 16:36	SMPL
O3799-01	MBH563	74	-0.0423 -		SFAM01.1	PPB	8/2/2023 16:39	SMPL
O3799-02	MBH564	39	-0.049 -		SFAM01.1	PPB	8/2/2023 16:41	SMPL
O3799-03	MBH563D	55	-0.0459 -		SFAM01.1	PPB	8/2/2023 16:43	SMPL
O3799-04	MBH563S	5434	0.9838 -		SFAM01.1	PPB	8/2/2023 16:45	SMPL
PB154601BL	PBW601	106	-0.0361 -		SFAM01.1	PPB	8/2/2023 16:48	SMPL
O3801-01	GCB93	191	-0.0199 -		SFAM01.1	PPB	8/2/2023 16:50	SMPL
O3801-02	GCB93D	45	-0.0478 -		SFAM01.1	PPB	8/2/2023 16:52	SMPL
O3801-03	GCB93S	5571	1.01 -		SFAM01.1	PPB	8/2/2023 16:54	SMPL
O3801-04	GCB94	66	-0.0438 -		SFAM01.1	PPB	8/2/2023 16:57	SMPL
O3801-05	GCB95	82	-0.0407 -		SFAM01.1	PPB	8/2/2023 16:59	SMPL
O3801-06	GCB96	86	-0.04 -		SFAM01.1	PPB	8/2/2023 17:01	SMPL
O3801-07	GCB97	126	-0.0323 -		SFAM01.1	PPB	8/2/2023 17:03	SMPL
O3801-08	GCB98	106	-0.0361 -		SFAM01.1	PPB	8/2/2023 17:06	SMPL
PB154602BL	PBW602	158	-0.0262 -		SFAM01.1	PPB	8/2/2023 17:08	SMPL

LB126700

INSTRUMENT ID : CV1

O3802-01	GCBA8	149	-0.0279 -	SFAM01.1	PPB	8/2/2023 17:10 SMPL
O3802-02	GCBA8D	271	-0.0046 -	SFAM01.1	PPB	8/2/2023 17:12 SMPL
O3802-03	GCBA8S	5439	0.9847 -	SFAM01.1	PPB	8/2/2023 17:15 SMPL
O3802-04	GCBA9	150	-0.0277 -	SFAM01.1	PPB	8/2/2023 17:17 SMPL
O3802-05	GCBB0	159	-0.026 -	SFAM01.1	PPB	8/2/2023 17:19 SMPL
O3802-06	GCBB1	170	-0.0239 -	SFAM01.1	PPB	8/2/2023 17:22 SMPL
CCV055	CCV055	25848	4.8915 -	SFAM01.1	PPB	8/2/2023 17:24 SMPL
CCB055	CCB055	68	-0.0434 -	SFAM01.1	PPB	8/2/2023 17:26 SMPL
O3802-07	GCBB2	60	-0.0449 -	SFAM01.1	PPB	8/2/2023 17:28 SMPL
O3802-08	GCBB3	91	-0.039 -	SFAM01.1	PPB	8/2/2023 17:31 SMPL
CCV056	CCV056	25605	4.845 -	SFAM01.1	PPB	8/2/2023 17:33 SMPL
CCB056	CCB056	71	-0.0428 -	SFAM01.1	PPB	8/2/2023 17:35 SMPL