



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB126715

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 98.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/03/2023	08:00
2	CAL2	1	Water	NA	NA	20.1	6.99	08/03/2023	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	08/03/2023	08:03
4	ICV	1	Water	NA	NA	20.7	7.02	08/03/2023	08:05
5	CCV1	1	Water	NA	NA	20.7	2.01	08/03/2023	08:10
6	O3872-01	1	Water	NA	NA	23.1	5.83	08/03/2023	08:15
7	O3872-02	1	Water	NA	NA	23.0	6.21	08/03/2023	08:20
8	O3873-01	1	Water	NA	NA	20.6	6.79	08/03/2023	08:25
9	O3876-01	1	Water	NA	NA	20.3	4.88	08/03/2023	08:30
10	O3876-02	1	Water	NA	NA	20.6	5.57	08/03/2023	08:35
11	O3876-03	1	Water	NA	NA	20.5	7.33	08/03/2023	08:40
12	O3881-01	1	Water	NA	NA	20.1	7.24	08/03/2023	08:45
13	O3881-02	1	Water	NA	NA	20.1	7.24	08/03/2023	08:50
14	O3882-01	1	Water	NA	NA	20.1	7.27	08/03/2023	08:53
15	O3888-01	1	Water	NA	NA	20.3	6.68	08/03/2023	09:00
16	CCV2	1	Water	NA	NA	20.7	12.01	08/03/2023	09:05
17	O3888-01DUP	1	Water	NA	NA	20.4	6.69	08/03/2023	09:06
18	CCV3	1	Water	NA	NA	20.7	2.01	08/03/2023	09:15

W3 126715

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 03872 WorkList ID : 172428 Department : Wet-Chemistry Date : 08-03-2023 07:43:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3872-01	D 50107	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3872-02	C 50109	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3873-01	A A462	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3876-01	A AUD-1030	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3876-02	A AUD-1031	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3876-03	A AUD-1024	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3881-01	Q MH-1	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3881-02	Q MH-2	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3882-01	C TRE-1763	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C
O3888-01	A A6014	Water	pH	Cool 4 deg C	PSEG03	L11	08/02/2023	9040C

Date/Time 08/03/23 07:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08/03/23 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]