

LB126717


SFAM01.1

INSTRUMENT ID: CV1

Linear

$\mu$  Abs.:

52957



A= 0.0000e+000

B= 1.9128e-004 slope

C= -1.5564e-002 y-intercept

Rho= 0.9996547

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.00   | 0.000  | -0.009 | -0.009 | 36    | 0.000      | 36    |       |       |       |       | 2  |
| 0.05   | 0.050  |        |        |       |            | 504   | 0     |       |       |       | 8  |
| 0.20   | 0.200  | 0.215  | 0.015  | 1207  | 0.0 %      | 1207  |       |       |       |       | 5  |
| 2.50   | 2.500  | 2.619  | 0.119  | 13774 | 0.0 %      | 13774 |       |       |       |       | -2 |
| 5.00   | 5.000  | 4.880  | -0.120 | 25592 | 0.0 %      | 25592 |       |       |       |       | -2 |
| 7.50   | 7.500  | 7.380  | -0.120 | 38663 | 0.0 %      | 38663 |       |       |       |       | -2 |
| 10.0   | 10.000 | 10.114 | 0.114  | 52957 | 0.0 %      | 52957 |       |       |       |       | 1  |

LB126717

INSTRUM,ENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method       | Units | Date           | Type |
|------------|-------------|------------|-----------|----------|--------------|-------|----------------|------|
|            | 0 S0        | 36         | -         |          | 0 SFAM01.1   | PPB   | 8/3/2023 10:35 | Std  |
|            | 0.2 S01     | 1207       | -         |          | 0.2 SFAM01.1 | PPB   | 8/3/2023 10:42 | Std  |
|            | 2.5 S02     | 13774      | -         |          | 2.5 SFAM01.1 | PPB   | 8/3/2023 10:45 | Std  |
|            | 5 S03       | 25592      | -         |          | 5 SFAM01.1   | PPB   | 8/3/2023 10:47 | Std  |
|            | 7.5 S04     | 38663      | -         |          | 7.5 SFAM01.1 | PPB   | 8/3/2023 10:49 | Std  |
|            | 10 S05      | 52957      | -         |          | 10 SFAM01.1  | PPB   | 8/3/2023 10:57 | Std  |
| ICV099     | ICV099      | 22088      | 4.2095 -  |          | SFAM01.1     | PPB   | 8/3/2023 11:00 | SMPL |
| ICB099     | ICB099      | 47         | -0.0066 - |          | SFAM01.1     | PPB   | 8/3/2023 11:02 | SMPL |
| CCV057     | CCV057      | 25978      | 4.9536 -  |          | SFAM01.1     | PPB   | 8/3/2023 11:04 | SMPL |
| CCB057     | CCB057      | 52         | -0.0056 - |          | SFAM01.1     | PPB   | 8/3/2023 11:06 | SMPL |
| PB154621BL | PBW621      | 129        | 0.0091 -  |          | SFAM01.1     | PPB   | 8/3/2023 11:09 | SMPL |
| O3573-01   | MDL-WATER   | 404        | 0.0617 -  |          | SFAM01.1     | PPB   | 8/3/2023 11:16 | SMPL |
| CCV058     | CCV058      | 25439      | 4.8505 -  |          | SFAM01.1     | PPB   | 8/3/2023 11:18 | SMPL |
| CCB058     | CCB058      | 57         | -0.0047 - |          | SFAM01.1     | PPB   | 8/3/2023 11:21 | SMPL |