

LB126721


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54421



A= 0.0000e+000

B= 1.8496e-004 slope

C= -4.4727e-002
-4.4727e-002 = intercept

Rho= 0.9999820

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.017	-0.017	148	0.000	148					
0.05	0.050					504	0				
0.20	0.200	0.210	0.010	1375	0.0 %	1375					
2.50	2.500	2.518	0.018	13858	0.0 %	13858					
5.00	5.000	5.009	0.009	27325	0.0 %	27325					
7.50	7.500	7.459	-0.041	40570	0.0 %	40570					
10.0	10.000	10.021	0.021	54421	0.0 %	54421					

115-101-0

LB126721

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	148	-		0 SFAM01.1	PPB	8/3/2023 13:31	Std
	0.2 S01	1375	-		0.2 SFAM01.1	PPB	8/3/2023 13:38	Std
	2.5 S02	13858	-		2.5 SFAM01.1	PPB	8/3/2023 13:40	Std
	5 S03	27325	-		5 SFAM01.1	PPB	8/3/2023 13:43	Std
	7.5 S04	40570	-		7.5 SFAM01.1	PPB	8/3/2023 13:45	Std
	10 S05	54421	-		10 SFAM01.1	PPB	8/3/2023 13:52	Std
ICV001	ICV001	21536	3.9386 -		SFAM01.1	PPB	8/3/2023 13:55	SMPL
ICB001	ICB001	150	-0.017 -		SFAM01.1	PPB	8/3/2023 13:57	SMPL
CCV059	CCV059	27575	5.0555 -		SFAM01.1	PPB	8/3/2023 13:59	SMPL
CCB059	CCB059	158	-0.0155 -		SFAM01.1	PPB	8/3/2023 14:02	SMPL
PB154622BL	PBS622	179	-0.0116 -		SFAM01.1	PPB	8/3/2023 14:04	SMPL
O3573-03	MDL-SOIL	474	0.0429 -		SFAM01.1	PPB	8/3/2023 14:06	SMPL
CCV060	CCV060	27756	5.089 -		SFAM01.1	PPB	8/3/2023 14:08	SMPL
CCB060	CCB060	237	-0.0009 -		SFAM01.1	PPB	8/3/2023 14:16	SMPL