



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126722

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	08/03/2023	08:10
2	O3872-01		1	12	100.00	>212.0	>212.0	08/03/2023	08:40
3	O3872-02		1	11	100.00	>212.0	>212.0	08/03/2023	09:10
4	O3873-01		1	14	100.00	>212.0	>212.0	08/03/2023	09:40
5	O3881-01		1	11	100.00	>212.0	>212.0	08/03/2023	10:10
6	O3881-02		1	14	100.00	>212.0	>212.0	08/03/2023	10:40
7	O3882-01		1	12	100.00	>212.0	>212.0	08/03/2023	11:10
8	O3888-01		1	12	100.00	>212.0	>212.0	08/03/2023	11:40
9	O3888-01DUP		1	13	100.00	>212.0	>212.0	08/03/2023	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6126722

WorkList Name : fp-8-2

WorkList ID : 172391

Department : Wet-Chemistry

Date : 08-02-2023 08:01:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3872-01 C	50107	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3872-02 D	50109	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3873-01 C	A462	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3881-01 X	MH-1	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3881-02 X	MH-2	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3882-01 F	TRE-1763	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3888-01 C	A6014	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B

Date/Time 08/03/2023 08:00
 Raw Sample Received by: PM (WC)
 Raw Sample Relinquished by: 70160801

Date/Time 08/03/2023 12:45
 Raw Sample Received by: 70160801
 Raw Sample Relinquished by: 70160801