



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126725

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP103549
Turbidity Calibration std, 0NTU	WP103548
Turbidity Calibration std, 20NTU	WP103543
Turbidity Calibration std, 40NTU	WP103542
Turbidity Calibration std, 1NTU	WP103546
Turbidity Calibration std, 5NTU	WP103545
Turbidity LOQ std, 1.0NTU	WP103550
Turbidity Calibration - CCV std, 10 NTU	WP103544
10 NTU Standard 500 ml	W2971

Intercept: 0.3035

Slope: 1.0723

Regression: 0.9995

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.158	-0.14		08/01/2023	10:30
2	CAL2	1	1	1.240	0.87	-12.7	08/01/2023	10:34
3	CAL3	5	1	5.520	4.87	-2.7	08/01/2023	10:38
4	CAL4	10	1	12.000	10.91	9.1	08/01/2023	10:42
5	CAL5	20	1	21.100	19.39	-3	08/01/2023	10:45
6	CAL6	40	1	43.300	40.10	0.2	08/01/2023	10:49



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126725

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.200	10.162	08/01/2023	10:53
2	ICB		1	0.162	-0.132	08/01/2023	10:57
3	CCV1	10	1	10.600	9.602	08/01/2023	11:01
4	CCB1		1	0.160	-0.134	08/01/2023	11:06
5	LB126725BL		1	0.164	-0.130	08/01/2023	11:10
6	O3572-10		1	0.824	0.485	08/01/2023	11:14
7	O3572-11		1	1.410	1.032	08/01/2023	11:18
8	O3587-02		1	6.700	5.965	08/01/2023	11:22
9	O3587-02DUP		1	6.700	5.965	08/01/2023	11:26
10	CCV2	10	1	11.600	10.535	08/01/2023	11:30
11	CCB2		1	0.168	-0.126	08/01/2023	11:34