

LB126726


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

61315



A= 0.0000e+000

B= 1.6545e-004

C= 3.5339e-002

Rho= 0.9994580

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%)
0.0	0.000	0.059	0.059	141	0.000	141					19
0.2	0.200	0.238	0.038	1223	0.0 %	1223					2
2.5	2.500	2.539	0.039	15131	0.0 %	15131					-3
5.0	5.000	4.852	-0.148	29111	0.0 %	29111					-2
7.5	7.500	7.333	-0.167	44109	0.0 %	44109					2
10.0	10.000	10.180	0.180	61315	0.0 %	61315					

LB126726 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Aug 2023 14:40:30

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	141	-	0.0000	7471B	PPB	03 Aug 2023 14:44:32	S	Std
0.2 - 1	50.2	1223	-	0.2000	7471B	PPB	03 Aug 2023 14:46:49	S	Std
2.5 - 1	52.5	15131	-	2.5000	7471B	PPB	03 Aug 2023 14:49:05	S	Std
5.0 - 1	55	29111	-	5.0000	7471B	PPB	03 Aug 2023 14:51:22	S	Std
7.5 - 1	57.5	44109	-	7.5000	7471B	PPB	03 Aug 2023 14:53:38	S	Std
10.0 - 1	51.6	61315	-	10.0000	7471B	PPB	03 Aug 2023 14:55:53	S	Std
ICV02 - 1	ICV02	26113	4.3557	-	7471B	PPB	03 Aug 2023 14:58:40	U	SMPL
ICB02 - 1	ICB02	123	0.0557	-	7471B	PPB	03 Aug 2023 15:00:55	U	SMPL
CCV61 - 1	CCV61	29964	4.9929	-	7471B	PPB	03 Aug 2023 15:03:09	U	SMPL
CCB61 - 1	CCB61	125	0.0560	-	7471B	PPB	03 Aug 2023 15:05:24	U	SMPL
CRA - 1	CRA	1166	0.2283	-	7471B	PPB	03 Aug 2023 15:07:39	U	SMPL
HighStd - 1	HighStd	61223	10.1647	-	7471B	PPB	03 Aug 2023 15:09:54	U	SMPL
ChkStd - 1	ChkStd	41734	6.9402	-	7471B	PPB	03 Aug 2023 15:12:09	U	SMPL
PB154617BL - 1	PBS	-14	0.0330	-	7471B	PPB	03 Aug 2023 15:14:26	U	SMPL
PB154617BS - 1	LCSS	24117	4.0255	-	7471B	PPB	03 Aug 2023 15:16:42	U	SMPL
O3853-01 - 1	34369-70-71	2587	0.4634	-	7471B	PPB	03 Aug 2023 15:18:58	U	SMPL
O3854-01 - 1	BERGENFIELD-SOIL-COMP	5305	0.9131	-	7471B	PPB	03 Aug 2023 15:21:14	U	SMPL
O3854-04 - 1	BERGENFIELD-CONCRETE	525	0.1222	-	7471B	PPB	03 Aug 2023 15:23:31	U	SMPL
O3860-01 - 1	SOUTH-WALL	3235	0.5706	-	7471B	PPB	03 Aug 2023 15:25:47	U	SMPL
O3860-03 - 1	SP-1A	5041	0.8694	-	7471B	PPB	03 Aug 2023 15:28:03	U	SMPL
CCV62 - 1	CCV62	30312	5.0505	-	7471B	PPB	03 Aug 2023 15:30:18	U	SMPL
CCB62 - 1	CCB62	200	0.0684	-	7471B	PPB	03 Aug 2023 15:32:33	U	SMPL
O3860-05 - 1	SP-2B	8281	1.4054	-	7471B	PPB	03 Aug 2023 15:34:48	U	SMPL
O3862-01 - 1	BUILDING-B-11-(10-15)	401	0.1017	-	7471B	PPB	03 Aug 2023 15:37:02	U	SMPL
O3862-01DUP - 1	BUILDING-B-11-(10-15)DUP	370	0.0966	-	7471B	PPB	03 Aug 2023 15:39:18	U	SMPL
O3862-02 - 1	BUILDING-B-3-(5-10)	3624	0.6349	-	7471B	PPB	03 Aug 2023 15:41:33	U	SMPL
O3862-03 - 1	BUILDING-B-17-(0-5)	31578	5.2599	-	7471B	PPB	03 Aug 2023 15:43:48	U	SMPL
O3862-04 - 1	O3862-01MS	26358	4.3963	-	7471B	PPB	03 Aug 2023 15:46:04	U	SMPL
O3862-05 - 1	O3862-01MSD	22776	3.8036	-	7471B	PPB	03 Aug 2023 15:48:20	U	SMPL
O3862-06 - 1	BUILDING-B-9-(5-10)	34197	5.6932	-	7471B	PPB	03 Aug 2023 15:50:37	U	SMPL
O3862-07 - 1	BUILDING-B-15-(5-10)	19523	3.2654	-	7471B	PPB	03 Aug 2023 15:52:53	U	SMPL
O3862-08 - 1	BUILDING-B-1-(0-5)	21638	3.6154	-	7471B	PPB	03 Aug 2023 15:55:09	U	SMPL
CCV63 - 1	CCV63	30654	5.1071	-	7471B	PPB	03 Aug 2023 15:57:24	U	SMPL
CCB63 - 1	CCB63	216	0.0711	-	7471B	PPB	03 Aug 2023 15:59:39	U	SMPL
O3864-01 - 1	WC-1	44655	7.4235	-	7471B	PPB	03 Aug 2023 16:01:55	U	SMPL
PB154618BL - 1	PBS	92	0.0506	-	7471B	PPB	03 Aug 2023 16:04:10	U	SMPL
PB154618BS - 1	LCSS	21519	3.5957	-	7471B	PPB	03 Aug 2023 16:06:25	U	SMPL
O3877-01 - 1	COMP-1	866	0.1786	-	7471B	PPB	03 Aug 2023 16:08:41	U	SMPL
O3877-01DUP - 1	COMP-1DUP	679	0.1477	-	7471B	PPB	03 Aug 2023 16:10:56	U	SMPL
O3877-01MS - 1	COMP-1MS	22971	3.8359	-	7471B	PPB	03 Aug 2023 16:13:12	U	SMPL
O3877-01MSD - 1	COMP-1MSD	22856	3.8189	-	7471B	PPB	03 Aug 2023 16:15:28	U	SMPL
O3877-04 - 1	COMP-2	1130	0.2223	-	7471B	PPB	03 Aug 2023 16:17:45	U	SMPL
O3877-07 - 1	COMP-3	920	0.1876	-	7471B	PPB	03 Aug 2023 16:20:02	U	SMPL
O3886-01 - 1	500-1A-1B	1446	0.2749	-	7471B	PPB	03 Aug 2023 16:22:17	U	SMPL
CCV64 - 1	CCV64	30738	5.1210	-	7471B	PPB	03 Aug 2023 16:24:33	U	SMPL
CCB64 - 1	CCB64	-44	0.0281	-	7471B	PPB	03 Aug 2023 16:26:48	U	SMPL
O3887-01 - 1	MCLEAN-TP6	1399	0.2668	-	7471B	PPB	03 Aug 2023 16:29:04	U	SMPL
O3887-04 - 1	MCLEAN-TP7	1399	0.2668	-	7471B	PPB	03 Aug 2023 16:31:19	U	SMPL
O3887-07 - 1	MCLEAN-TP4	1626	0.3044	-	7471B	PPB	03 Aug 2023 16:33:35	U	SMPL
O3887-10 - 1	MCLEAN-TP2	382	0.0985	-	7471B	PPB	03 Aug 2023 16:35:51	U	SMPL
O3887-13 - 1	MCLEAN-TP9	685	0.1487	-	7471B	PPB	03 Aug 2023 16:38:06	U	SMPL
O3862-01LX5 - 1		25	0.0395	-	7471B	PPB	03 Aug 2023 16:40:22	U	SMPL
O3862-01A - 1		24737	4.1281	-	7471B	PPB	03 Aug 2023 16:42:38	U	SMPL
O3877-01LX5 - 1		102	0.0522	-	7471B	PPB	03 Aug 2023 16:44:55	U	SMPL
O3877-01A - 1		24324	4.0598	-	7471B	PPB	03 Aug 2023 16:47:11	U	SMPL
CCV65 - 1	CCV65	31131	5.1860	-	7471B	PPB	03 Aug 2023 16:49:27	U	SMPL
CCB65 - 1	CCB65	-62	0.0251	-	7471B	PPB	03 Aug 2023 16:51:42	U	SMPL