

LB126736


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44935



A= 0.0000e+000

B= 2.2395e-004

C= -1.7885e-002

Rho= 0.9998967

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.007	-0.007	50	0.000	50	0				
0.2	0.200	0.185	-0.015	907	0.0 %	907					-8
2.5	2.500	2.533	0.033	11392	0.0 %	11392					1
5.0	5.000	5.048	0.048	22623	0.0 %	22623					1
7.5	7.500	7.395	-0.105	33099	0.0 %	33099					1
10.0	10.000	10.045	0.045	44935	0.0 %	44935					0

LB126736 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Aug 2023 08:15:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	50	-	0.0000	7470A	PPB	04 Aug 2023 08:32:34	S	Std
0.2 - 1	50.2	907	-	0.2000	7470A	PPB	04 Aug 2023 08:34:50	S	Std
2.5 - 1	52.5	11392	-	2.5000	7470A	PPB	04 Aug 2023 08:37:07	S	Std
5.0 - 1	55	22623	-	5.0000	7470A	PPB	04 Aug 2023 08:39:24	S	Std
7.5 - 1	57.5	33099	-	7.5000	7470A	PPB	04 Aug 2023 08:41:40	S	Std
10.0 - 1	51.1	44935	-	10.0000	7470A	PPB	04 Aug 2023 08:49:16	S	Std
ICV03 - 1	ICV03	18841	4.2015	-	7470A	PPB	04 Aug 2023 08:51:59	U	SMPL
ICB03 - 1	ICB03	-56	-0.0309	-	7470A	PPB	04 Aug 2023 08:54:14	U	SMPL
CCV66 - 1	CCV66	21691	4.8397	-	7470A	PPB	04 Aug 2023 08:56:29	U	SMPL
CCB66 - 1	CCB66	-60	-0.0313	-	7470A	PPB	04 Aug 2023 08:58:46	U	SMPL
CRA - 1	CRA	890	0.1814	-	7470A	PPB	04 Aug 2023 09:01:01	U	SMPL
HighStd - 1	HighStd	45862	10.2528	-	7470A	PPB	04 Aug 2023 09:03:15	U	SMPL
ChkStd - 1	ChkStd	30490	6.8102	-	7470A	PPB	04 Aug 2023 09:05:31	U	SMPL
PB154623BL - 1	PBW	-93	-0.0387	-	7470A	PPB	04 Aug 2023 09:07:47	U	SMPL
PB154623BS - 1	LCSW	18823	4.1975	-	7470A	PPB	04 Aug 2023 09:10:03	U	SMPL
O3854-02 - 1	BERGENFIELD-SOIL-COMP	-52	-0.0295	-	7470A	PPB	04 Aug 2023 09:12:19	U	SMPL
O3854-05 - 1	BERGENFIELD-CONCRETE	0	-0.0179	-	7470A	PPB	04 Aug 2023 09:14:36	U	SMPL
O3877-02 - 1	COMP-1	25	-0.0123	-	7470A	PPB	04 Aug 2023 09:16:53	U	SMPL
O3877-05 - 1	COMP-2	36	-0.0098	-	7470A	PPB	04 Aug 2023 09:19:09	U	SMPL
O3877-08 - 1	COMP-3	37	-0.0096	-	7470A	PPB	04 Aug 2023 09:21:24	U	SMPL
CCV67 - 1	CCV67	21834	4.8718	-	7470A	PPB	04 Aug 2023 09:23:39	U	SMPL
CCB67 - 1	CCB67	-50	-0.0291	-	7470A	PPB	04 Aug 2023 09:25:54	U	SMPL
O3887-02 - 1	MCLEAN-TP6	17	-0.0141	-	7470A	PPB	04 Aug 2023 09:28:09	U	SMPL
O3887-05 - 1	MCLEAN-TP7	71	-0.0020	-	7470A	PPB	04 Aug 2023 09:30:24	U	SMPL
O3887-08 - 1	MCLEAN-TP4	12	-0.0152	-	7470A	PPB	04 Aug 2023 09:32:40	U	SMPL
O3887-11 - 1	MCLEAN-TP2	59	-0.0047	-	7470A	PPB	04 Aug 2023 09:34:55	U	SMPL
O3887-14 - 1	MCLEAN-TP9	46	-0.0076	-	7470A	PPB	04 Aug 2023 09:37:11	U	SMPL
O3887-14DUP - 1	MCLEAN-TP9DUP	54	-0.0058	-	7470A	PPB	04 Aug 2023 09:39:27	U	SMPL
O3887-14MS - 1	MCLEAN-TP9MS	15427	3.4369	-	7470A	PPB	04 Aug 2023 09:41:44	U	SMPL
O3887-14MSD - 1	MCLEAN-TP9MSD	16250	3.6212	-	7470A	PPB	04 Aug 2023 09:44:00	U	SMPL
PB154624BL - 1	PBW	-19	-0.0221	-	7470A	PPB	04 Aug 2023 09:46:17	U	SMPL
PB154624BS - 1	LCSW	16067	3.5803	-	7470A	PPB	04 Aug 2023 09:48:33	U	SMPL
CCV68 - 1	CCV68	21872	4.8803	-	7470A	PPB	04 Aug 2023 09:50:48	U	SMPL
CCB68 - 1	CCB68	-43	-0.0275	-	7470A	PPB	04 Aug 2023 09:53:03	U	SMPL
O3862-09 - 1	EQUIPMENT-BLANK-1	18	-0.0139	-	7470A	PPB	04 Aug 2023 09:55:18	U	SMPL
O3862-10 - 1	FIELD-BLANK-1	-28	-0.0242	-	7470A	PPB	04 Aug 2023 09:57:34	U	SMPL
O3873-01 - 1	A462	94	0.0032	-	7470A	PPB	04 Aug 2023 09:59:49	U	SMPL
O3876-01 - 1	AUD-1030	5542	1.2232	-	7470A	PPB	04 Aug 2023 10:02:05	U	SMPL
O3876-02 - 1	AUD-1031	460	0.0851	-	7470A	PPB	04 Aug 2023 10:04:21	U	SMPL
O3876-03 - 1	AUD-1024	579	0.1118	-	7470A	PPB	04 Aug 2023 10:06:38	U	SMPL
O3878-02 - 1	COMP	201	0.0271	-	7470A	PPB	04 Aug 2023 10:08:54	U	SMPL
O3881-01 - 1	MH-1	1	-0.0177	-	7470A	PPB	04 Aug 2023 10:11:11	U	SMPL
O3881-02 - 1	MH-2	28	-0.0116	-	7470A	PPB	04 Aug 2023 10:13:28	U	SMPL
O3882-01 - 1	TRE-1763	22	-0.0130	-	7470A	PPB	04 Aug 2023 10:15:44	U	SMPL
CCV69 - 1	CCV69	21436	4.7826	-	7470A	PPB	04 Aug 2023 10:17:59	U	SMPL
CCB69 - 1	CCB69	6	-0.0165	-	7470A	PPB	04 Aug 2023 10:20:14	U	SMPL
O3888-01 - 1	A6014	7498	1.6613	-	7470A	PPB	04 Aug 2023 10:22:29	U	SMPL
O3888-01DUP - 1	A6014DUP	6738	1.4911	-	7470A	PPB	04 Aug 2023 10:24:45	U	SMPL
O3888-01MS - 1	A6014MS	21654	4.8315	-	7470A	PPB	04 Aug 2023 10:27:01	U	SMPL
O3888-01MSD - 1	A6014MSD	23878	5.3295	-	7470A	PPB	04 Aug 2023 10:29:17	U	SMPL
PB154598TB - 1	PB154598TB	-24	-0.0233	-	7470A	PPB	04 Aug 2023 10:31:32	U	SMPL
O3887-14LX5 - 1		-24	-0.0233	-	7470A	PPB	04 Aug 2023 10:33:48	U	SMPL
O3887-14A - 1		18678	4.1650	-	7470A	PPB	04 Aug 2023 10:36:06	U	SMPL
O3888-01LX5 - 1		1453	0.3075	-	7470A	PPB	04 Aug 2023 10:38:22	U	SMPL
O3888-01A - 1		23746	5.3000	-	7470A	PPB	04 Aug 2023 10:40:39	U	SMPL
CCV70 - 1	CCV70	22692	5.0639	-	7470A	PPB	04 Aug 2023 10:42:55	U	SMPL
CCB70 - 1	CCB70	27	-0.0118	-	7470A	PPB	04 Aug 2023 10:45:11	U	SMPL