



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126740

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP103615
Turbidity Calibration std, 0NTU	WP103614
Turbidity Calibration std, 20NTU	WP103609
Turbidity Calibration std, 40NTU	WP103608
Turbidity Calibration std, 1NTU	WP103612
Turbidity Calibration std, 5NTU	WP103611
Turbidity Calibration - CCV std, 10 NTU	WP103610
10 NTU Standard 500 ml	W2971

Intercept: 0.2448

Slope: 1.0791

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.160	-0.08		08/04/2023	09:50
2	CAL2	1	1	1.210	0.89	-10.6	08/04/2023	09:54
3	CAL3	5	1	5.510	4.88	-2.4	08/04/2023	10:00
4	CAL4	10	1	11.800	10.71	7.1	08/04/2023	10:04
5	CAL5	20	1	21.300	19.51	-2.4	08/04/2023	10:08
6	CAL6	40	1	43.500	40.09	0.2	08/04/2023	10:12



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126740

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.800	9.781	08/04/2023	10:16
2	ICB		1	0.168	-0.071	08/04/2023	10:20
3	CCV1	10	1	11.100	10.059	08/04/2023	10:24
4	CCB1		1	0.166	-0.073	08/04/2023	10:28
5	LB126740BL		1	0.170	-0.069	08/04/2023	10:32
6	O3572-12		1	0.800	0.515	08/04/2023	10:36
7	CCV2	10	1	11.000	9.967	08/04/2023	10:40
8	CCB2		1	0.166	-0.073	08/04/2023	10:44