



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126741

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/04/2023	08:10
2	O3876-01		1	14	100.00	>212.0	>212.0	08/04/2023	08:40
3	O3876-02		1	12	100.00	>212.0	>212.0	08/04/2023	09:10
4	O3876-03		1	13	100.00	>212.0	>212.0	08/04/2023	09:40
5	O3896-03		1	14	100.00	>212.0	>212.0	08/04/2023	10:10
6	O3900-03		1	12	100.00	>212.0	>212.0	08/04/2023	10:40
7	O3900-03DUP		1	12	100.00	>212.0	>212.0	08/04/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126741

WorkList Name : fp-8-23

WorkList ID : 172462

Department : Wet-Chemistry

Date : 08-04-2023 07:52:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3876-01 C	AUD-1030	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3876-02 C	AUD-1031	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3876-03 C	AUD-1024	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/02/2023	1010B
O3896-03 E	50119	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/03/2023	1010B
O3900-03 A	1613	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/03/2023	1010B

Date/Time 08/04/2023 08.00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JN CWC

Date/Time 08/04/2023 11.45
 Raw Sample Received by: JN CWC
 Raw Sample Relinquished by: RM CWC