



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126742

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/04/2023	12:00
2	O3900-02		1	11	100.00	>212.0	>212.0	08/04/2023	12:30
3	O3900-02DUP		1	11	100.00	>212.0	>212.0	08/04/2023	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26126742

WorkList Name : FP S-8-04

WorkList ID : 172475

Department : Wet-Chemistry

Date : 08-04-2023 09:04:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3900-02 <u>B</u>	31921	Solid	Flash Point	Cool 4 deg C	PSEG03	L11	08/03/2023	1010B

Date/Time 08/04/2023 11:50
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: RM cwc

Date/Time 08/04/2023 13:40
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: RM cwc