


LB126746

SFAM01.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.013	0.013	101	0.000	101					=
0.05	0.050					504					10
0.20	0.200	0.220	0.020	1181	0.0 %	1181	0				5
2.50	2.500	2.615	0.115	13708	0.0 %	13708					-3
5.00	5.000	4.868	-0.132	25498	0.0 %	25498					-2
7.50	7.500	7.324	-0.176	38343	0.0 %	38343					2
10.0	10.000	10.160	0.160	53177	0.0 %	53177					

LB126746

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	101	-		0 SFAM01.1	PPB	8/4/2023 13:43	Std
	0.2 S01	1181	-		0.2 SFAM01.1	PPB	8/4/2023 13:45	Std
	2.5 S02	13708	-		2.5 SFAM01.1	PPB	8/4/2023 13:47	Std
	5 S03	25498	-		5 SFAM01.1	PPB	8/4/2023 13:50	Std
	7.5 S04	38343	-		7.5 SFAM01.1	PPB	8/4/2023 13:52	Std
	10 S05	53177	-		10 SFAM01.1	PPB	8/4/2023 13:54	Std
ICV004	ICV004	22004	4.2005 -		SFAM01.1	PPB	8/4/2023 13:57	SMPL
ICB004	ICB004	41	0.0018 -		SFAM01.1	PPB	8/4/2023 13:59	SMPL
CCV071	CCV071	25934	4.9518 -		SFAM01.1	PPB	8/4/2023 14:01	SMPL
CCB071	CCB071	29	-0.0005 -		SFAM01.1	PPB	8/4/2023 14:04	SMPL
PB154656BL	PBW656	78	0.0089 -		SFAM01.1	PPB	8/4/2023 14:06	SMPL
O3573-02	MDL-WATER	260	0.0437 -		SFAM01.1	PPB	8/4/2023 14:08	SMPL
CCV072	CCV072	25341	4.8385 -		SFAM01.1	PPB	8/4/2023 14:10	SMPL
CCB072	CCB072	33	0.0003 -		SFAM01.1	PPB	8/4/2023 14:13	SMPL