



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126751

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/04/2023	15:20
2	CAL2	1	Water	NA	NA	20.3	7.00	08/04/2023	15:21
3	CAL3	1	Water	NA	NA	20.4	10.02	08/04/2023	15:22
4	ICV	1	Water	NA	NA	20.4	6.98	08/04/2023	15:25
5	CCV1	1	Water	NA	NA	20.4	2.01	08/04/2023	15:30
6	O3896-01	1	Solid	20.02	20	21.1	7.95	08/04/2023	15:35
7	O3896-01DUP	1	Solid	20.03	20	21.2	7.96	08/04/2023	15:36
8	CCV2	1	Water	NA	NA	20.7	12.02	08/04/2023	15:40

WORKLIST(Hardcopy Internal Chain)

LB 118561

WorkList Name : fp2422 WorkList ID : 156582 Department : Wet-Chemistry Date : 02-04-2022 10:43:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2022	Water	N1366-01	C Flash Point	Cool 4 deg C	PSEG03	M11	N28660	02/02/2022	1010B
02/09/2022	Water	N1368-01	F Flash Point	Cool 4 deg C	PSEG03	M11	31382	02/02/2022	1010B
02/10/2022	Water	N1392-01	C Flash Point	Cool 4 deg C	PSEG03	M12	OILY-WATER-NWB-1735	02/03/2022	1010B

Date/Time 02/07/22 09:05
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: KS(wc)

Date/Time 02/07/22 11:20
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: KS(wc)