



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126752

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/04/2023	15:20
2	CAL2	1	Water	NA	NA	20.3	7.00	08/04/2023	15:21
3	CAL3	1	Water	NA	NA	20.4	10.02	08/04/2023	15:22
4	ICV	1	Water	NA	NA	20.4	6.98	08/04/2023	15:25
5	CCV1	1	Water	NA	NA	20.4	2.01	08/04/2023	15:30
6	O3896-03	1	Water	NA	NA	20.8	7.29	08/04/2023	15:44
7	O3912-01	1	Water	NA	NA	23.6	7.37	08/04/2023	15:45
8	O3912-01DUP	1	Water	NA	NA	23.7	7.38	08/04/2023	15:46
9	CCV2	1	Water	NA	NA	20.6	12.02	08/04/2023	15:50

UB 126752

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 03862 w

WorkList ID : 172493

Department : Wet-Chemistry

Date : 08-04-2023 15:00:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3896-03	50119	Water	pH	Cool 4 deg C	PSEG03	L11	08/03/2023	9040C
O3912-01	COMPOSITE-1	Water	pH	Cool 4 deg C	PSEG03	L11	08/04/2023	9040C

15:05

Date/Time 08/04/23
 Raw Sample Received by: JT (SM)
 Raw Sample Relinquished by: JT (SM)

Date/Time 08/04/23
 Raw Sample Received by: JT (SM)
 Raw Sample Relinquished by: JT (SM)