

Clear table

Instrument ID: IC-2										Analyst :		IZ		Method:		300.0 / 9056A	
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst						
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-18_	7/18/23 11:54		IZ/MA						
STD2	0.5190	0.6440	0.6590	2.1930	0.5650	1.2310	3.4630	2023-07-18_	7/18/23 12:25		IZ/MA						
STD3	0.7210	1.2220	1.2040	3.9830	1.0030	1.9630	6.0460	2023-07-18_	7/18/23 12:55		IZ/MA						
STD4	1.0180	1.5110	1.5050	5.1180	1.2620	2.5080	7.5420	2023-07-18_	7/18/23 13:26		IZ/MA						
STD5	1.9840	2.9150	2.9260	9.6980	2.4210	4.8360	14.4150	2023-07-18_	7/18/23 13:57		IZ/MA						
STD6	3.7660	5.9300	5.9290	19.6870	4.9030	9.6670	29.4280	2023-07-18_	7/18/23 14:27		IZ/MA						
STD7	5.1680	7.5780	7.5770	25.3210	6.3470	12.7960	38.1050	2023-07-18_	7/18/23 14:58		IZ/MA						
ICV	2.0250	2.9710	2.9980	9.8960	2.5030	4.9290	14.7030	2023-07-18_	7/18/23 15:29		IZ/MA						
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-18_	7/18/23 16:00		IZ/MA						
CCV	1.8800	3.0090	2.9520	9.9100	2.4550	4.9410	14.6360	2023-08-02_	8/2/23 10:04		IZ/MA						
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-08-02_	8/2/23 10:38		IZ/MA						
LB126753BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-08-02_	8/2/23 11:09		IZ/MA						
LB126753BSW	1.9340	3.0900	3.0650	10.1540	2.5200	5.0580	15.2710	2023-08-02_	8/2/23 11:39		IZ/MA						
O3881-01	1.1420	350.0070	0.0000	1.6490	0.2090	0.0000	38.9460	2023-08-02_	8/2/23 15:53		IZ/MA						
O3881-01MS	3.8230	340.7610	2.8040	11.4660	2.6040	11.8560	54.8680	2023-08-02_	8/2/23 16:23		IZ/MA						
O3881-01MSD	3.7100	340.5950	2.8660	11.4600	2.6310	11.8310	55.2100	2023-08-02_	8/2/23 16:54		IZ/MA						
O3881-02	1.1570	371.1460	0.0000	1.7790	0.0000	0.0000	40.8610	2023-08-02_	8/2/23 17:25		IZ/MA						
O3874-01X10	0.0000	194.3290	0.0000	1.1590	0.0000	1.1460	1.8440	2023-08-02_	8/2/23 17:55		IZ/MA						
O3874-01DLX300	0.0000	4.8590	0.0000	0.0000	0.0000	0.0000	0.9910	2023-08-02_	8/2/23 18:26		IZ/MA						
CCV	1.9360	3.1760	3.1600	9.8760	2.5880	5.8460	15.4190	2023-08-02_	8/2/23 18:57		IZ/MA						
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-08-02_	8/2/23 19:27		IZ/MA						

Clear table

Instrument ID: IC-2 Analyst : IZ Method: 300.0 / 9056A


ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-18_	7/18/23 11:54		IZ/MA
STD2	0.5190	0.6440	0.6590	2.1930	0.5650	1.2310	3.4630	2023-07-18_	7/18/23 12:25		IZ/MA
STD3	0.7210	1.2220	1.2040	3.9830	1.0030	1.9630	6.0460	2023-07-18_	7/18/23 12:55		IZ/MA
STD4	1.0180	1.5110	1.5050	5.1180	1.2620	2.5080	7.5420	2023-07-18_	7/18/23 13:26		IZ/MA
STD5	1.9840	2.9150	2.9260	9.6980	2.4210	4.8360	14.4150	2023-07-18_	7/18/23 13:57		IZ/MA
STD6	3.7660	5.9300	5.9290	19.6870	4.9030	9.6670	29.4280	2023-07-18_	7/18/23 14:27		IZ/MA
STD7	5.1680	7.5780	7.5770	25.3210	6.3470	12.7960	38.1050	2023-07-18_	7/18/23 14:58		IZ/MA

ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000

ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4
STD1							
STD2	29.7500	7.3333	9.8333	9.6500	13.0000	23.1000	15.4333
STD3	-9.8750	1.8333	0.3333	-0.4250	0.3000	-1.8500	0.7667
STD4	1.8000	0.7333	0.3333	2.3600	0.9600	0.3200	0.5600
STD5	-0.8000	-2.8333	-2.4667	-3.0200	-3.1600	-3.2800	-3.9000
STD6	-5.8500	-1.1667	-1.1833	-1.5650	-1.9400	-3.3300	-1.9067
STD7	3.3600	1.0400	1.0267	1.2840	1.5520	2.3680	2.9865

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0893931 \cdot A + 3.11575$
 RSD: 7.560 %
 Correlation coefficient: 0.996880




K3 = 0 K2 = 0 K1 = 0.0893931 K0 = 3.11575
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	1.069	85.37	0.4	20	6.387	
3	1.887	124.8	0.8	20	6.387	
4	2.787	195.8	1	20	6.387	
5	6.176	409	2	20	6.387	
6	12.27	807.7	4	20	6.387	
7	17.4	1121	5	20	6.387	

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.129379 \cdot A + 2.57559$
 RSD: 2.187 %
 Correlation coefficient: 0.999739




K3 = 0 K2 = 0 K1 = 0.129379 K0 = 2.57559
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	5.821	79.67	0.6	20	8.957	
3	11.94	169	1.2	20	8.957	
4	15.21	213.7	1.5	20	8.957	
5	31.35	430.7	3	20	8.957	
6	65.62	896.7	6	20	8.957	
7	84.19	1152	7.5	20	8.957	

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0571447 \cdot A + 2.7868$
 RSD: 2.142 %
 Correlation coefficient: 0.999750




K3 = 0 K2 = 0 K1 = 0.0571447 K0 = 2.7868
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	10.91	181.7	0.6	20	10.82		
3	22.02	372.5	1.2	20	10.82		
4	28.54	478.1	1.5	20	10.82		
5	58.22	975.2	3	20	10.82		
6	117	2026	6	20	10.82		
7	150.1	2603	7.5	20	10.82		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.288518 \cdot A + 10.6539$
 RSD: 2.666 %
 Correlation coefficient: 0.999612




K3 = 0 K2 = 0 K1 = 0.288518 K0 = 10.6539
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.375	115.1	2	20	13.99		
3	13.22	239.2	4	20	13.99		
4	17.19	317.8	5	20	13.99		
5	35.66	635.4	10	20	13.99		
6	74.24	1328	20	20	13.99		
7	95.06	1718	25	20	13.99		

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0499072 \cdot A + 2.72307$
 RSD: 3.120 %
 Correlation coefficient: 0.999469




K3 = 0 K2 = 0 K1 = 0.0499072 K0 = 2.72307
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.794	171.7	0.5	20	15.52		
3	15.93	347.4	1	20	15.52		
4	20.7	451.3	1.25	20	15.52		
5	42.9	915.4	2.5	20	15.52		
6	91.09	1910	5	20	15.52		
7	118.8	2489	6.25	20	15.52		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0597633 \cdot A + 12.5173$
 RSD: 4.814 %
 Correlation coefficient: 0.998736




K3 = 0 K2 = 0 K1 = 0.0597633 K0 = 12.5173
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	4.234	202.4	1	20	22.61		
3	9.564	447.6	2	20	22.61		
4	13.76	629.8	2.5	20	22.61		
5	32.81	1409	5	20	22.61		
6	75.03	3026	10	20	22.61		
7	103.9	4073	12.5	20	22.61		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.162956 \cdot A + 15.5782$
 RSD: 3.392 %
 Correlation coefficient: 0.999373



K3 = 0 K2 = 0 K1 = 0.162956 K0 = 15.5782
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0.5229		20.4	0	20		26.01
2	11.32		329.4	3	20		26.01
3	22.74		646.5	6	20		26.01
4	29.39		830	7.5	20		26.01
5	60.94		1674	15	20		26.01
6	130.3		3516	30	20		26.01
7	169.7		4581	37.5	20		26.01

Report date: 7/19/2023 11:11:53 AM
Printed by: wet


Ident: STD1
Analysis from: 7/18/2023 11:54:38 AM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34856

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:01 AM
Printed by: wet


Ident: STD2
Analysis from: 7/18/2023 12:25:16 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34857

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.32	1.167	1.03	2.17	81.326	7.00
2	9.12	0.201	5.82	12.26	79.672	6.86
3	10.98	0.233	10.91	22.98	181.715	15.65
4	14.19	0.275	6.37	13.42	115.119	9.91
5	16.21	0.326	7.79	16.41	171.712	14.78
6	21.82	0.586	4.23	8.91	202.440	17.43
7	25.12	0.434	11.32	23.83	329.435	28.36
7	28.00	0.460	47.48	99.99	1161.418	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:10 AM
Printed by: wet


Ident: STD3
Analysis from: 7/18/2023 12:55:54 PM
File: _2023-07-18_

Last save: 7/18/2023 3:32:11 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34858

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.27	1.020	1.90	1.95	126.563	5.39
2	9.13	0.202	11.94	12.27	168.983	7.19
3	10.98	0.237	22.02	22.62	372.529	15.86
4	14.17	0.274	13.22	13.59	239.159	10.18
5	16.17	0.323	15.93	16.37	347.382	14.79
6	21.86	0.580	9.56	9.83	447.587	19.06
7	25.16	0.426	22.74	23.37	646.495	27.53
7	28.00	0.437	97.31	100.00	2348.698	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:18 AM
Printed by: wet


Ident: STD4
Analysis from: 7/18/2023 1:26:31 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34859

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.35	1.020	2.76	2.16	192.844	6.19
2	9.15	0.202	15.21	11.92	213.663	6.86
3	11.01	0.235	28.54	22.38	478.126	15.36
4	14.25	0.280	17.19	13.47	317.821	10.21
5	16.26	0.322	20.70	16.23	451.305	14.49
6	22.01	0.566	13.76	10.79	629.820	20.23
7	25.33	0.424	29.39	23.04	830.037	26.66
7	28.00	0.436	127.54	99.99	3113.615	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:26 AM
Printed by: wet


Ident: STD5
Analysis from: 7/18/2023 1:57:09 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34860

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.27	0.962	6.17	2.30	409.042	6.34
2	9.07	0.198	31.35	11.70	430.714	6.68
3	10.89	0.235	58.22	21.72	975.219	15.12
4	14.05	0.269	35.66	13.30	635.355	9.85
5	15.98	0.314	42.90	16.00	915.443	14.20
6	21.75	0.533	32.81	12.24	1408.769	21.85
7	25.05	0.413	60.94	22.73	1673.615	25.95
7	28.00	0.418	268.05	100.00	6448.158	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:33 AM
Printed by: wet


Ident: STD6
Analysis from: 7/18/2023 2:27:46 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34861

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.26	0.960	12.27	2.17	807.679	5.98
2	9.07	0.196	65.62	11.60	896.723	6.64
3	10.88	0.244	117.03	20.69	2026.370	15.00
4	14.04	0.269	74.24	13.13	1327.749	9.83
5	15.93	0.308	91.08	16.11	1910.213	14.14
6	21.75	0.503	75.02	13.27	3025.506	22.39
7	25.08	0.407	130.25	23.03	3516.225	26.03
7	28.00	0.412	565.52	100.00	13510.466	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:39 AM
Printed by: wet


Ident: STD7
Analysis from: 7/18/2023 2:58:23 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34862

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.34	0.949	17.40	2.35	1121.440	6.32
2	9.09	0.196	84.19	11.39	1151.573	6.49
3	10.89	0.245	150.10	20.31	2603.211	14.68
4	14.06	0.273	95.06	12.86	1718.350	9.69
5	15.93	0.306	118.78	16.07	2488.841	14.03
6	21.80	0.491	103.93	14.06	4072.751	22.96
7	25.15	0.407	169.72	22.96	4581.095	25.83
7	28.00	0.410	739.17	100.00	17737.261	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:14:06 AM
Printed by: wet


Ident: ICV
Analysis from: 7/18/2023 3:29:09 PM
File: _2023-07-18_

Last save: 7/18/2023 3:57:09 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34863

Last save: 7/18/2023 3:26:

SAMPLE:
: IZ/MA
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.26	0.954	6.38	2.31	418.174	6.33
2	9.03	0.197	32.27	11.67	439.302	6.65
3	10.83	0.233	60.17	21.76	1000.409	15.15
4	13.96	0.265	36.92	13.35	649.049	9.83
5	15.86	0.310	44.57	16.12	948.693	14.36
6	21.71	0.528	33.75	12.20	1440.075	21.80
7	25.01	0.412	62.48	22.59	1708.970	25.88
7	28.00	0.414	276.55	100.00	6604.672	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:14:16 AM
Printed by: wet


Ident: ICB
Analysis from: 7/18/2023 4:00:04 PM
File: _2023-07-18_

Last save: 7/19/2023 8:57:46 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34864

Last save: 7/18/2023 3:26:

SAMPLE:
: IZ/MA
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:33:10 PM
Printed by: wet


Ident: CCV
Analysis from: 8/2/2023 10:04:19 AM
File: _2023-08-02_

Last save: 8/2/2023 10:35:32 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35025

Last save: 7/19/2023 10:12

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.55	1.109	5.11	2.19	385.653	5.90
2	9.69	0.232	28.56	12.24	445.313	6.81
3	11.77	0.280	52.60	22.53	984.394	15.05
4	15.42	0.334	29.20	12.51	650.068	9.94
5	17.71	0.413	33.41	14.31	929.346	14.21
6	22.21	0.622	30.14	12.91	1444.037	22.08
7	25.44	0.473	54.41	23.31	1700.730	26.01
7	28.00	0.495	233.44	100.00	6539.541	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:33:27 PM
Printed by: wet


Ident: CCB
Analysis from: 8/2/2023 10:38:22 AM
File: _2023-08-02_

Last save: 8/2/2023 11:06:23 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35026

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:49:23 PM
Printed by: wet

Ident: LB126753BLW
Analysis from: 8/2/2023 11:09:00 AM
File: _2023-08-02_


Last save: 8/4/2023 3:49:14 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35027

Last save: 8/2/2023 10:35:

SAMPLE:

: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:49:31 PM
Printed by: wet


Ident: LB126753BSW
Analysis from: 8/2/2023 11:39:38 AM
File: _2023-08-02_

Last save: 8/4/2023 3:49:14 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35028

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.50	1.116	5.28	2.16	397.908	5.88
2	9.52	0.227	29.71	12.14	457.691	6.77
3	11.54	0.275	55.13	22.53	1024.030	15.14
4	15.06	0.323	31.10	12.71	666.935	9.86
5	17.25	0.397	35.76	14.61	955.417	14.13
6	22.01	0.613	31.35	12.81	1483.128	21.93
7	25.26	0.478	56.38	23.04	1778.638	26.30
7	28.00	0.490	244.71	100.00	6763.747	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:34:56 PM
Printed by: wet


Ident: 03881-01
Analysis from: 8/2/2023 3:53:14 PM
File: _2023-08-02_

Last save: 8/2/2023 4:21:15 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35030

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.65	1.446	2.44	0.14	220.707	0.37
2	9.87	0.554	1613.94	91.04	54085.670	91.52
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.34	0.316	3.76	0.21	77.412	0.13
5	17.66	0.399	1.12	0.06	29.093	0.05
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.27	0.463	151.55	8.55	4684.361	7.93
7	28.00	0.454	1772.81	100.00	59097.243	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:37:02 PM
Printed by: wet


Ident: 03881-01MS
Analysis from: 8/2/2023 4:23:52 PM
File: _2023-08-02_

Last save: 8/2/2023 4:51:52 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35031

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.59	1.229	10.22	0.50	820.548	1.23
2	9.82	0.542	1607.70	79.43	52656.451	79.12
3	11.56	0.287	43.82	2.16	932.434	1.40
4	15.16	0.308	37.01	1.83	757.868	1.14
5	17.33	0.391	37.76	1.87	988.821	1.49
6	22.11	0.704	69.87	3.45	3758.261	5.65
7	25.26	0.461	217.56	10.75	6638.528	9.97
7	28.00	0.560	2023.95	100.00	66552.911	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:37:20 PM
Printed by: wet


Ident: 03881-01MSD
Analysis from: 8/2/2023 4:54:30 PM
File: _2023-08-02_

Last save: 8/2/2023 5:22:30 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35032

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.52	1.243	9.82	0.48	795.233	1.19
2	9.70	0.542	1610.39	79.25	52630.764	79.06
3	11.40	0.282	44.39	2.18	954.411	1.43
4	14.93	0.302	37.71	1.86	757.458	1.14
5	17.06	0.383	38.56	1.90	999.850	1.50
6	21.88	0.702	69.78	3.43	3749.863	5.63
7	25.00	0.457	221.22	10.89	6680.500	10.04
7	28.00	0.559	2031.87	100.00	66568.080	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:37:47 PM
Printed by: wet

Ident: 03881-02
Analysis from: 8/2/2023 5:25:08 PM
File: _2023-08-02_


Last save: 8/3/2023 11:50:49 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35033

Last save: 8/2/2023 10:35:

SAMPLE:

: IZ/MA
Vial number: 20
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.53	1.329	2.62	0.15	224.085	0.36
2	9.74	0.586	1622.45	90.65	57353.457	91.64
3	0.00	0.000	0.00	0.00	0.000	0.00
4	14.99	0.308	4.31	0.24	86.379	0.14
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.05	0.459	160.35	8.96	4919.321	7.86
7	28.00	0.383	1789.72	100.00	62583.242	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:38:11 PM
Printed by: wet


Ident: 03874-01X10
Analysis from: 8/2/2023 5:55:46 PM
File: _2023-08-02_

Last save: 8/2/2023 6:23:46 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35034

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 21
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.54	0.322	1415.01	99.41	30020.279	98.85
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.13	0.307	2.20	0.15	43.409	0.14
5	0.00	0.000	0.00	0.00	0.000	0.00
6	22.18	0.962	2.51	0.18	174.166	0.57
7	25.26	0.522	3.65	0.26	130.724	0.43
7	28.00	0.302	1423.37	100.00	30368.578	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:38:25 PM
Printed by: wet


Ident: O3874-01DLX300
Analysis from: 8/2/2023 6:26:23 PM
File: _2023-08-02_

Last save: 8/2/2023 6:54:23 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35035

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 22
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.35	0.218	49.07	98.86	731.169	96.56
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.09	0.675	0.56	1.13	26.031	3.44
7	28.00	0.128	49.63	99.99	757.200	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:38:46 PM
Printed by: wet


Ident: CCV
Analysis from: 8/2/2023 6:57:05 PM
File: _2023-08-02_

Last save: 8/2/2023 7:25:06 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35036

Last save: 8/2/2023 10:35:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.41	1.161	5.12	1.98	398.378	5.61
2	9.37	0.221	31.41	12.18	471.125	6.64
3	11.34	0.272	57.22	22.18	1057.190	14.89
4	14.75	0.308	32.20	12.48	647.665	9.12
5	16.87	0.384	38.09	14.76	982.404	13.84
6	21.80	0.645	35.26	13.67	1746.888	24.60
7	24.97	0.464	58.68	22.75	1796.774	25.31
7	28.00	0.494	257.99	100.00	7100.424	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/4/2023 3:39:03 PM
Printed by: wet

Ident: CCB
Analysis from: 8/2/2023 7:27:44 PM
File: _2023-08-02_


Last save: 8/2/2023 7:55:44 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35037

Last save: 8/2/2023 10:35:

SAMPLE:

: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD