



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126755

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/04/2023	15:00
2	CAL2	1	Water	NA	NA	20.7	7.00	08/04/2023	15:01
3	CAL3	1	Water	NA	NA	20.7	10.02	08/04/2023	15:02
4	ICV	1	Water	NA	NA	20.8	7.01	08/04/2023	15:05
5	CCV1	1	Water	NA	NA	20.6	2.01	08/04/2023	15:10
6	O3891-02	1	Solid	20.02	20	20.2	7.51	08/04/2023	15:11
7	O3891-04	1	Solid	20.03	20	20.7	8.48	08/04/2023	15:15
8	O3891-06	1	Solid	20.05	20	20.2	9.13	08/04/2023	15:20
9	O3895-02	1	Solid	20.02	20	20.1	5.30	08/04/2023	15:25
10	O3908-02	1	Solid	20.02	20	20.2	7.07	08/04/2023	15:30
11	O3913-02	1	Solid	20.04	20	20.2	9.30	08/04/2023	15:40
12	O3913-02DUP	1	Solid	20.03	20	20.3	9.31	08/04/2023	15:41
13	CCV2	1	Water	NA	NA	20.7	12.03	08/04/2023	15:45

VB 26755

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o3905

WorkList ID : 172484

Department : Wet-Chemistry

Date : 08-04-2023 10:42:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3891-02	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/03/2023	9045D
O3891-04	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/03/2023	9045D
O3891-06	TP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/03/2023	9045D
O3895-02	MCLEAN-TP11	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/03/2023	9045D
O3908-02	ETGI-368	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/04/2023	9045D
O3913-02	72-11933	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/04/2023	9045D

Date/Time

08/04/23 14:30

Raw Sample Received by:

JP (WCC)

Raw Sample Relinquished by:

RM SM

Date/Time

08/04/23

Raw Sample Received by:

RM SM

Raw Sample Relinquished by:

JP (WCC)