

LB126761


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

68857



A= 0.0000e+000

B= 1.4784e-004

C= -2.2117e-002

Rho= 0.9991238

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	2/01
0.0	0.000	0.085	0.085	727	0.000	727					1
0.2	0.200	0.221	0.021	1645	0.0 %	1645					11
2.5	2.500	2.381	-0.119	16256	0.0 %	16256					-5
5.0	5.000	5.135	0.135	34883	0.0 %	34883					3
7.5	7.500	7.220	-0.280	48989	0.0 %	48989					-4
10.0	10.000	10.157	0.157	68857	0.0 %	68857					2

LB126761 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Aug 2023 11:17:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	727	-	0.00007471B	PPB		07 Aug 2023 11:32:28	S	Std
0.2 - 1	50.2	1645	-	0.20007471B	PPB		07 Aug 2023 11:34:45	S	Std
2.5 - 1	52.5	16256	-	2.50007471B	PPB		07 Aug 2023 11:37:02	S	Std
5.0 - 1	55	34883	-	5.00007471B	PPB		07 Aug 2023 11:39:19	S	Std
7.5 - 1	57.5	48989	-	7.50007471B	PPB		07 Aug 2023 11:41:35	S	Std
10.0 - 1	510	68857	-	10.00007471B	PPB		07 Aug 2023 11:49:14	S	Std
ICV06 - 1	ICV06	28176	4.1433	-7471B	PPB		07 Aug 2023 11:52:02	U	SMPL
ICB06 - 1	ICB06	605	0.0673	-7471B	PPB		07 Aug 2023 11:54:17	U	SMPL
CCV75 - 1	CCV75	33387	4.9137	-7471B	PPB		07 Aug 2023 11:56:32	U	SMPL
CCB75 - 1	CCB75	605	0.0673	-7471B	PPB		07 Aug 2023 11:58:47	U	SMPL
CRA - 1	CRA	1632	0.2192	-7471B	PPB		07 Aug 2023 12:01:02	U	SMPL
HighStd - 1	HighStd	69993	10.3254	-7471B	PPB		07 Aug 2023 12:03:17	U	SMPL
ChkStd - 1	ChkStd	46906	6.9123	-7471B	PPB		07 Aug 2023 12:05:32	U	SMPL
PB154675BL - 1	PBS	733	0.0862	-7471B	PPB		07 Aug 2023 12:07:48	U	SMPL
PB154675BS - 1	LCSS	26611	3.9119	-7471B	PPB		07 Aug 2023 12:10:04	U	SMPL
O3903-01 - 1	MW-01	979	0.1226	-7471B	PPB		07 Aug 2023 12:12:20	U	SMPL
O3903-02 - 1	MW-02	922	0.1142	-7471B	PPB		07 Aug 2023 12:14:37	U	SMPL
O3903-03 - 1	MW-03	936	0.1163	-7471B	PPB		07 Aug 2023 12:16:54	U	SMPL
O3903-04 - 1	MW-04	831	0.1007	-7471B	PPB		07 Aug 2023 12:19:11	U	SMPL
O3903-05 - 1	MW-05	829	0.1004	-7471B	PPB		07 Aug 2023 12:21:26	U	SMPL
CCV76 - 1	CCV76	33668	4.9552	-7471B	PPB		07 Aug 2023 12:23:42	U	SMPL
CCB76 - 1	CCB76	624	0.0701	-7471B	PPB		07 Aug 2023 12:25:58	U	SMPL
O3903-06 - 1	MW-06	963	0.1202	-7471B	PPB		07 Aug 2023 12:28:13	U	SMPL
O3903-07 - 1	MW-07	735	0.0865	-7471B	PPB		07 Aug 2023 12:30:28	U	SMPL
O3903-08 - 1	MW-08	847	0.1031	-7471B	PPB		07 Aug 2023 12:32:44	U	SMPL
O3903-09 - 1	MW-09	1123	0.1439	-7471B	PPB		07 Aug 2023 12:34:59	U	SMPL
O3903-10 - 1	MW-10	873	0.1069	-7471B	PPB		07 Aug 2023 12:37:14	U	SMPL
O3903-11 - 1	MW-11	1411	0.1865	-7471B	PPB		07 Aug 2023 12:39:30	U	SMPL
O3903-11DUP - 1	MW-11DUP	1317	0.1726	-7471B	PPB		07 Aug 2023 12:45:16	U	SMPL
O3903-11MS - 1	MW-11MS	25275	3.7144	-7471B	PPB		07 Aug 2023 12:50:15	U	SMPL
O3903-11MSD - 1	MW-11MSD	25810	3.7935	-7471B	PPB		07 Aug 2023 12:52:31	U	SMPL
PB154676BL - 1	PBS	783	0.0936	-7471B	PPB		07 Aug 2023 12:54:46	U	SMPL
CCV77 - 1	CCV77	33569	4.9406	-7471B	PPB		07 Aug 2023 12:57:01	U	SMPL
CCB77 - 1	CCB77	383	0.0345	-7471B	PPB		07 Aug 2023 12:59:17	U	SMPL
PB154676BS - 1	LCSS	24939	3.6648	-7471B	PPB		07 Aug 2023 13:01:32	U	SMPL
O3880-01 - 1	1272	2454	0.3407	-7471B	PPB		07 Aug 2023 13:03:48	U	SMPL
O3891-01 - 1	TP-1	1256	0.1636	-7471B	PPB		07 Aug 2023 13:06:03	U	SMPL
O3891-03 - 1	TP-2	4691	0.6714	-7471B	PPB		07 Aug 2023 13:08:19	U	SMPL
O3891-05 - 1	TP-3	1491	0.1983	-7471B	PPB		07 Aug 2023 13:10:34	U	SMPL
O3895-01 - 1	MCLEAN-TP11	3774	0.5358	-7471B	PPB		07 Aug 2023 13:12:50	U	SMPL
O3895-01DUP - 1	MCLEAN-TP11DUP	4484	0.6408	-7471B	PPB		07 Aug 2023 13:15:06	U	SMPL
O3895-01MS - 1	MCLEAN-TP11MS	28744	4.2273	-7471B	PPB		07 Aug 2023 13:17:23	U	SMPL
O3895-01MSD - 1	MCLEAN-TP11MSD	29464	4.3337	-7471B	PPB		07 Aug 2023 13:22:19	U	SMPL
O3897-01 - 1	WC-1	14638	2.1419	-7471B	PPB		07 Aug 2023 13:24:34	U	SMPL
CCV78 - 1	CCV78	33939	4.9953	-7471B	PPB		07 Aug 2023 13:26:49	U	SMPL
CCB78 - 1	CCB78	647	0.0735	-7471B	PPB		07 Aug 2023 13:29:05	U	SMPL
O3898-01 - 1	TR-4-080323	3209	0.4523	-7471B	PPB		07 Aug 2023 13:31:20	U	SMPL
O3899-01 - 1	520	1156	0.1488	-7471B	PPB		07 Aug 2023 13:33:35	U	SMPL
O3899-02 - 1	522	1460	0.1937	-7471B	PPB		07 Aug 2023 13:35:51	U	SMPL
O3899-03 - 1	15017	1497	0.1992	-7471B	PPB		07 Aug 2023 13:38:06	U	SMPL
O3900-01 - 1	31464	824	0.0997	-7471B	PPB		07 Aug 2023 13:40:21	U	SMPL
O3900-02 - 1	31921	3514	0.4974	-7471B	PPB		07 Aug 2023 13:42:37	U	SMPL
O3908-01 - 1	ETGI-368	4579	0.6548	-7471B	PPB		07 Aug 2023 13:44:53	U	SMPL
O3911-01 - 1	PSCR206(164716)	8354	1.2129	-7471B	PPB		07 Aug 2023 13:47:13	U	SMPL
O3913-01 - 1	72-11933	2598	0.3620	-7471B	PPB		07 Aug 2023 13:49:29	U	SMPL
O3917-01 - 1	B-PP04AA	1969	0.2690	-7471B	PPB		07 Aug 2023 13:51:45	U	SMPL
CCV79 - 1	CCV79	34786	5.1205	-7471B	PPB		07 Aug 2023 13:54:00	U	SMPL
CCB79 - 1	CCB79	629	0.0709	-7471B	PPB		07 Aug 2023 13:56:16	U	SMPL
O3917-02 - 1	B-PP04AB	1196	0.1547	-7471B	PPB		07 Aug 2023 13:58:31	U	SMPL
O3917-03 - 1	B-PP05A	12751	1.8629	-7471B	PPB		07 Aug 2023 14:00:47	U	SMPL
O3917-04 - 1	B-PP05B	1760	0.2381	-7471B	PPB		07 Aug 2023 14:03:03	U	SMPL
O3903-11LX5 - 1		228	0.0116	-7471B	PPB		07 Aug 2023 14:05:18	U	SMPL
O3903-11A - 1		30876	4.5425	-7471B	PPB		07 Aug 2023 14:07:33	U	SMPL
O3895-01LX5 - 1		555	0.0599	-7471B	PPB		07 Aug 2023 14:09:50	U	SMPL
O3895-01A - 1		31294	4.6043	-7471B	PPB		07 Aug 2023 14:12:06	U	SMPL
CCV80 - 1	CCV80	35074	5.1631	-7471B	PPB		07 Aug 2023 14:14:23	U	SMPL
CCB80 - 1	CCB80	229	0.0117	-7471B	PPB		07 Aug 2023 14:16:39	U	SMPL