



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126769

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/07/2023	12:05
2	CAL2	1	Water	NA	NA	20.7	7.00	08/07/2023	12:06
3	CAL3	1	Water	NA	NA	20.7	10.03	08/07/2023	12:08
4	ICV	1	Water	NA	NA	20.7	6.99	08/07/2023	12:10
5	CCV1	1	Water	NA	NA	20.7	2.01	08/07/2023	12:15
6	O3919-02	1	Solid	20.02	20	21.2	6.72	08/07/2023	12:20
7	O3919-05	1	Solid	20.03	20	21.3	6.05	08/07/2023	12:25
8	O3919-05DUP	1	Solid	20.04	20	20.8	6.06	08/07/2023	12:26
9	CCV2	1	Water	NA	NA	20.7	12.03	08/07/2023	12:30

WORKLIST(Hardcopy Internal Chain)

126769

WorkList Name : corrosivity 03921

WorkList ID : 172525

Department : Wet-Chemistry

Date : 08-07-2023 11:52:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3919-02	TP-38	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/05/2023	9045D
O3919-05	TP-36	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	08/05/2023	9045D

Date/Time 08/07/23 12:00

Raw Sample Received by: JDC

Raw Sample Relinquished by: JT(Sm)

Date/Time 08/07/23 13:30

Raw Sample Received by: JT(Sm)

Raw Sample Relinquished by: JDC