



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126770

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	08/07/2023	15:30
2	CAL2	1	Water	NA	NA	20.6	7.00	08/07/2023	15:31
3	CAL3	1	Water	NA	NA	20.7	10.02	08/07/2023	15:33
4	ICV	1	Water	NA	NA	20.7	7.01	08/07/2023	15:35
5	CCV1	1	Water	NA	NA	20.6	2.02	08/07/2023	15:40
6	O3923-01	1	Water	NA	NA	22.5	4.94	08/07/2023	15:45
7	O3924-02	1	Water	NA	NA	24.0	8.03	08/07/2023	15:50
8	O3924-02DUP	1	Water	NA	NA	24.1	8.04	08/07/2023	15:51
9	CCV2	1	Water	NA	NA	20.7	12.01	08/07/2023	16:00

W3 126770

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 03923 w

WorkList ID : 172528

Department : Wet-Chemistry

Date : 08-07-2023 15:13:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3923-01	✓ POOL-WATER	Water	pH	Cool 4 deg C	STER04	K11	08/04/2023	9040C
O3924-02	✓ 08922	Water	pH	Cool 4 deg C	PSEG03	L11	08/07/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/07/23 15:25

JO (woc)

IT (9m)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/07/23

IT (5m)

JO (woc)

17:00