



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126771

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	08/07/2023	15:30
2	CAL2	1	Water	NA	NA	20.6	7.00	08/07/2023	15:31
3	CAL3	1	Water	NA	NA	20.7	10.02	08/07/2023	15:33
4	ICV	1	Water	NA	NA	20.7	7.01	08/07/2023	15:35
5	CCV1	1	Water	NA	NA	20.6	2.02	08/07/2023	15:40
6	O3924-01	1	Solid	20.02	20	22.8	4.81	08/07/2023	16:10
7	O3924-01DUP	1	Solid	20.01	20	22.8	4.82	08/07/2023	16:11
8	O3926-01	1	Solid	20.02	20	22.1	7.75	08/07/2023	16:20
9	CCV2	1	Water	NA	NA	20.7	12.02	08/07/2023	16:25

JB 126771

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 03926 WorkList ID : 172527 Department : Wet-Chemistry Date : 08-07-2023 15:12:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3924-01	1115	Solid	pH	Cool 4 deg C	PSEG03	L11	08/07/2023	9045D
O3926-01	OILY-DEBRIS	Solid	pH	Cool 4 deg C	PSEG03	L11	08/07/2023	9045D

Date/Time 08/07/23 15:25
Raw Sample Received by: JT (3m)
Raw Sample Relinquished by: JT (3m)

Date/Time 08/07/23 17:00
Raw Sample Received by: JT (3m)
Raw Sample Relinquished by: JT (3m)