

LB126775


7470A

INSTRUMENT ID: CY1

Linear

μ Abs.:

49544



A= 0.0000e+000

B= 2.0443e-004

C= 3.0796e-002

Rho= 0.9995315

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.028	0.028	-13	0.000	-13	0				5
0.2	0.200	0.211	0.011	883	0.0 %	883					3
2.5	2.500	2.577	0.077	12456	0.0 %	12456					2
5.0	5.000	4.918	-0.082	23906	0.0 %	23906					3
7.5	7.500	7.306	-0.194	35588	0.0 %	35588					2
10.0	10.000	10.159	0.159	49544	0.0 %	49544					

LB126775 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Aug 2023 10:39:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-13	-	0.0000	7470A	PPB	08 Aug 2023 10:56:38	S	Std
0.2 - 1	50.2	883	-	0.2000	7470A	PPB	08 Aug 2023 10:58:54	S	Std
2.5 - 1	52.5	12456	-	2.5000	7470A	PPB	08 Aug 2023 11:01:11	S	Std
5.0 - 1	55	23906	-	5.0000	7470A	PPB	08 Aug 2023 11:03:27	S	Std
7.5 - 1	57.5	35588	-	7.5000	7470A	PPB	08 Aug 2023 11:05:44	S	Std
10.0 - 1	510	49544	-	10.0000	7470A	PPB	08 Aug 2023 11:10:36	S	Std
ICV07 - 1	ICV07	19170	3.9498	-	7470A	PPB	08 Aug 2023 11:13:26	U	SMPL
ICB07 - 1	ICB07	21	0.0351	-	7470A	PPB	08 Aug 2023 11:18:11	U	SMPL
CCV81 - 1	CCV81	23712	4.8783	-	7470A	PPB	08 Aug 2023 11:20:26	U	SMPL
CCB81 - 1	CCB81	-29	0.0249	-	7470A	PPB	08 Aug 2023 11:22:40	U	SMPL
CRA - 1	CRA	880	0.2107	-	7470A	PPB	08 Aug 2023 11:24:55	U	SMPL
HighStd - 1	HighStd	48775	10.0020	-	7470A	PPB	08 Aug 2023 11:27:11	U	SMPL
ChkStd - 1	ChkStd	32577	6.6906	-	7470A	PPB	08 Aug 2023 11:29:27	U	SMPL
PB154679BL - 1	PBW	24	0.0357	-	7470A	PPB	08 Aug 2023 11:34:32	U	SMPL
PB154679BS - 1	LCSW	16571	3.4185	-	7470A	PPB	08 Aug 2023 11:36:48	U	SMPL
O3891-02 - 1	TP-1	248	0.0815	-	7470A	PPB	08 Aug 2023 11:42:14	U	SMPL
O3891-04 - 1	TP-2	301	0.0923	-	7470A	PPB	08 Aug 2023 11:44:30	U	SMPL
O3891-06 - 1	TP-3	240	0.0799	-	7470A	PPB	08 Aug 2023 11:46:48	U	SMPL
O3895-02 - 1	MCLEAN-TP11	403	0.1132	-	7470A	PPB	08 Aug 2023 11:49:04	U	SMPL
O3908-02 - 1	ETGI-368	441	0.1210	-	7470A	PPB	08 Aug 2023 11:51:20	U	SMPL
CCV82 - 1	CCV82	23629	4.8613	-	7470A	PPB	08 Aug 2023 11:53:35	U	SMPL
CCB82 - 1	CCB82	-36	0.0234	-	7470A	PPB	08 Aug 2023 11:55:51	U	SMPL
O3911-02 - 1	PSCR206(164716)	324	0.0970	-	7470A	PPB	08 Aug 2023 11:58:05	U	SMPL
O3913-02 - 1	72-11933	200	0.0717	-	7470A	PPB	08 Aug 2023 12:00:20	U	SMPL
O3913-02DUP - 1	72-11933DUP	126	0.0566	-	7470A	PPB	08 Aug 2023 12:02:35	U	SMPL
O3913-02MS - 1	72-11933MS	17144	3.5356	-	7470A	PPB	08 Aug 2023 12:04:51	U	SMPL
O3913-02MSD - 1	72-11933MSD	16304	3.3639	-	7470A	PPB	08 Aug 2023 12:07:06	U	SMPL
PB154654TB - 1	PB154654TB	179	0.0674	-	7470A	PPB	08 Aug 2023 12:11:31	U	SMPL
O3913-02LX5 - 1		148	0.0611	-	7470A	PPB	08 Aug 2023 12:13:46	U	SMPL
O3913-02A - 1		18962	3.9073	-	7470A	PPB	08 Aug 2023 12:16:03	U	SMPL
CCV83 - 1	CCV83	23781	4.8924	-	7470A	PPB	08 Aug 2023 12:18:20	U	SMPL
CCB83 - 1	CCB83	-43	0.0220	-	7470A	PPB	08 Aug 2023 12:20:36	U	SMPL