



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/9/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 08/08/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 10:00
Out Date: 08/09/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126776

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3929-01	B-PP19	1	1.14	8.85	9.99	8.16	79.3	
O3929-02	B-PP20	2	1.15	8.82	9.97	8.21	80.0	
O3929-03	DUP-07	3	1.18	8.78	9.96	7.76	74.9	
O3930-01	VNJ-207	11	1.18	8.41	9.59	8.5	87.0	
O3930-02	72-11935	12	1.17	8.40	9.57	8.77	90.5	
O3931-01	DAYTON-SOIL1	18	1.14	8.77	9.91	8.21	80.6	
O3931-03	DAYTON-SOIL2	19	1.13	8.74	9.87	8.36	82.7	
O3932-01	NEVINS-SB-01-Z18-20	4	1.15	8.38	9.53	8.04	82.2	
O3932-02	NEVINS-SB-01-Z20-22	5	1.19	8.52	9.71	8.11	81.2	
O3932-03	NEVINS-SB-01-Z30-32	6	1.14	8.85	9.99	8.48	82.9	
O3932-04	NEVINS-SB-01-Z67-69	7	1.11	8.88	9.99	9.16	90.7	
O3935-01	AT014P-MW01D-SB02-00G	8	1.16	8.48	9.64	8.75	89.5	
O3935-02	AT014P-MW01I-SB02-00G	9	1.18	8.56	9.74	8.63	87.0	
O3935-03	AT014P-MW01S-SB02-00G	10	1.12	8.50	9.62	8.55	87.4	
O3939-01	ORA-1745-1746-1747	13	1.19	8.42	9.61	8.77	90.0	
O3939-02	ORA-1745-1746-1747-EPH-2	14	1.19	8.51	9.7	8.83	89.8	
O3939-03	ORA-1748-1719-1750-1751-1752-1753-1754	15	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3939-06	ORA-1748-1719-1750-1751-1752-1753-1754	16	1.12	8.46	9.58	8.27	84.5	
O3939-07	ORA-1748-1719-1750-1751-1752-1753-1754-EPH-2	17	1.11	8.49	9.6	8.26	84.2	
O3941-01	DEAN-OILY-DEBRIS	20	1.12	8.54	9.66	9.36	96.5	
O3941-03	DEAN-OILY-STONE	21	1.00	1.00	2.00	2.00	100.0	oily-stone
O3941-04	DEAN-OILY-STONE-EPH-2	22	1.00	1.00	2.00	2.00	100.0	oily-stone
O3943-01	B-PP21A	23	1.17	8.53	9.7	8.56	86.6	
O3943-02	B-PP21B	24	1.19	8.60	9.79	8.71	87.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$