

LB126781


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55056



A= 0.0000e+000

B= 1.8214e-004 slope

C= 1.4446e-002 y-intercept

Rho= 0.9999386

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.028	0.028	75	0.000	75					
0.05	0.050					504	0				
0.20	0.200	0.221	0.021	1135	0.0 %	1135					
2.50	2.500	2.492	-0.008	13601	0.0 %	13601					
5.00	5.000	4.917	-0.083	26915	0.0 %	26915					
7.50	7.500	7.500	-0.000	41098	0.0 %	41098					
10.0	10.000	10.042	0.042	55056	0.0 %	55056					

100% 11/11

LB126781

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	75	-		0 SFAM01.1	PPB	8/8/2023 15:29	Std
	0.2 S01	1135	-		0.2 SFAM01.1	PPB	8/8/2023 15:31	Std
	2.5 S02	13601	-		2.5 SFAM01.1	PPB	8/8/2023 15:33	Std
	5 S03	26915	-		5 SFAM01.1	PPB	8/8/2023 15:36	Std
	7.5 S04	41098	-		7.5 SFAM01.1	PPB	8/8/2023 15:38	Std
	10 S05	55056	-		10 SFAM01.1	PPB	8/8/2023 15:40	Std
ICV008	ICV008	22449	4.1033	-	SFAM01.1	PPB	8/8/2023 15:43	SMPL
ICB008	ICB008	57	0.0248	-	SFAM01.1	PPB	8/8/2023 15:45	SMPL
CCV084	CCV084	26964	4.9256	-	SFAM01.1	PPB	8/8/2023 15:47	SMPL
CCB084	CCB084	53	0.0241	-	SFAM01.1	PPB	8/8/2023 15:50	SMPL
PB154708BL	PBS708	73	0.0277	-	SFAM01.1	PPB	8/8/2023 15:52	SMPL
03785-01	GCAx0	1062	0.2079	-	SFAM01.1	PPB	8/8/2023 15:54	SMPL
03785-02	GCAx1	1044	0.2046	-	SFAM01.1	PPB	8/8/2023 15:56	SMPL
03785-03	GCAx2	851	0.1694	-	SFAM01.1	PPB	8/8/2023 15:59	SMPL
03785-04	GCAx3	842	0.1678	-	SFAM01.1	PPB	8/8/2023 16:01	SMPL
03785-05	GCAx4	991	0.1949	-	SFAM01.1	PPB	8/8/2023 16:03	SMPL
03785-06	GCAx5	1041	0.2041	-	SFAM01.1	PPB	8/8/2023 16:05	SMPL
03785-07	GCAx6	1041	0.2041	-	SFAM01.1	PPB	8/8/2023 16:08	SMPL
03785-08	GCAx7	1093	0.2135	-	SFAM01.1	PPB	8/8/2023 16:10	SMPL
03785-09	GCAx8	980	0.1929	-	SFAM01.1	PPB	8/8/2023 16:12	SMPL
03785-10	GCAx9	1103	0.2153	-	SFAM01.1	PPB	8/8/2023 16:15	SMPL
03785-11	GCAy0	1125	0.2194	-	SFAM01.1	PPB	8/8/2023 16:17	SMPL
03785-12	GCAy0D	1119	0.2183	-	SFAM01.1	PPB	8/8/2023 16:19	SMPL
03785-13	GCAy0S	15029	2.7518	-	SFAM01.1	PPB	8/8/2023 16:21	SMPL
03785-14	GCAy1	1301	0.2514	-	SFAM01.1	PPB	8/8/2023 16:24	SMPL
03785-15	GCAy2	1702	0.3244	-	SFAM01.1	PPB	8/8/2023 16:26	SMPL
03785-16	GCAy3	1065	0.2084	-	SFAM01.1	PPB	8/8/2023 16:28	SMPL
03785-17	GCAy4	1051	0.2059	-	SFAM01.1	PPB	8/8/2023 16:30	SMPL
03785-18	GCAy5	1213	0.2354	-	SFAM01.1	PPB	8/8/2023 16:33	SMPL
03785-19	GCAy6	1177	0.2288	-	SFAM01.1	PPB	8/8/2023 16:35	SMPL
CCV085	CCV085	27170	4.9632	-	SFAM01.1	PPB	8/8/2023 16:37	SMPL
CCB085	CCB085	67	0.0266	-	SFAM01.1	PPB	8/8/2023 16:39	SMPL
03785-20	GCAy7	1209	0.2347	-	SFAM01.1	PPB	8/8/2023 16:42	SMPL
03785-21	GCAy8	1315	0.254	-	SFAM01.1	PPB	8/8/2023 16:44	SMPL
03785-22	GCAy9	1388	0.2673	-	SFAM01.1	PPB	8/8/2023 16:46	SMPL
03785-23	GCAz0	26048	4.7588	-	SFAM01.1	PPB	8/8/2023 16:48	SMPL
PB154709BL	PBS709	85	0.0299	-	SFAM01.1	PPB	8/8/2023 16:55	SMPL
03786-01	GCAz1	1990	0.3769	-	SFAM01.1	PPB	8/8/2023 16:58	SMPL
03786-02	GCAz1D	2250	0.4243	-	SFAM01.1	PPB	8/8/2023 17:00	SMPL
03786-03	GCAz1S	15846	2.9006	-	SFAM01.1	PPB	8/8/2023 17:02	SMPL
03786-04	GCAz2	965	0.1902	-	SFAM01.1	PPB	8/8/2023 17:04	SMPL
03786-05	GCAz3	1096	0.2141	-	SFAM01.1	PPB	8/8/2023 17:07	SMPL
03786-06	GCAz4	1320	0.2549	-	SFAM01.1	PPB	8/8/2023 17:09	SMPL
03786-07	GCAz5	2274	0.4286	-	SFAM01.1	PPB	8/8/2023 17:11	SMPL
03786-08	GCAz6	2007	0.38	-	SFAM01.1	PPB	8/8/2023 17:13	SMPL
03786-09	GCAz7	1686	0.3215	-	SFAM01.1	PPB	8/8/2023 17:16	SMPL

LB126781
INSTRUMENT ID : CV1

O3786-10	GCAZ8	1345	0.2594 -	SFAM01.1	PPB	8/8/2023 17:18 SMPL
O3786-11	GCAZ9	27116	4.9533 -	SFAM01.1	PPB	8/8/2023 17:25 SMPL
CCV086	CCV086	27214	4.9712 -	SFAM01.1	PPB	8/8/2023 17:27 SMPL
CCB086	CCB086	61	0.0256 -	SFAM01.1	PPB	8/8/2023 17:29 SMPL