



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126788

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/08/2023	15:10
2	CAL2	1	Water	NA	NA	20.7	7.00	08/08/2023	15:11
3	CAL3	1	Water	NA	NA	20.7	10.02	08/08/2023	15:15
4	ICV	1	Water	NA	NA	20.7	7.00	08/08/2023	15:20
5	CCV1	1	Water	NA	NA	20.8	2.01	08/08/2023	16:15
6	O3931-05	1	Water	NA	NA	22.7	6.26	08/08/2023	16:20
7	O3934-01	1	Water	NA	NA	20.6	4.96	08/08/2023	16:25
8	O3934-01DUP	1	Water	NA	NA	20.7	4.97	08/08/2023	16:26
9	CCV2	1	Water	NA	NA	20.7	12.02	08/08/2023	16:30

VB 126788

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o3941

WorkList ID : 172557

Department : Wet-Chemistry

Date : 08-08-2023 16:05:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3931-05	DAYTON-TOTES	Water	pH	Cool 4 deg C	PSEG03	L11	08/08/2023	9040C
O3934-01	POOL-WATER	Water	pH	Cool 4 deg C	STER05	L13	08/03/2023	9040C

Date/Time 08/08/23 16:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 08/08/23 17:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]