

LB126797


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

6151E



A= 0.0000e+000

B= 1.6048e-004

C= -5.3138e-002

Rho= 0.9993231

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0.1
0.0	0.000	-0.018	-0.018	216	0.000	216					-
0.2	0.200	0.181	-0.019	1462	0.0 %	1462					-10
2.5	2.500	2.378	-0.122	15147	0.0 %	15147					-5
5.0	5.000	5.218	0.218	32844	0.0 %	32844					4
7.5	7.500	7.622	0.122	47826	0.0 %	47826					2
10.0	10.000	9.819	-0.181	61518	0.0 %	61518					-2

LB126797 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Aug 2023 12:32:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	216	-	0.0000	7471B	PPB	09 Aug 2023 13:47:53	S	Std
0.2 - 1	51.2	1462	-	0.2000	7471B	PPB	09 Aug 2023 13:50:09	S	Std
2.5 - 1	52.5	15147	-	2.5000	7471B	PPB	09 Aug 2023 13:52:26	S	Std
5.0 - 1	55	32844	-	5.0000	7471B	PPB	09 Aug 2023 13:54:43	S	Std
7.5 - 1	57.5	47826	-	7.5000	7471B	PPB	09 Aug 2023 13:56:59	S	Std
10.0 - 1	510	61518	-	10.0000	7471B	PPB	09 Aug 2023 14:01:49	S	Std
ICV10 - 1	ICV10	24922	3.9464	-	7471B	PPB	09 Aug 2023 14:07:08	U	SMPL
ICB10 - 1	ICB10	200	-0.0210	-	7471B	PPB	09 Aug 2023 14:09:23	U	SMPL
CCV91 - 1	CCV91	32916	5.2293	-	7471B	PPB	09 Aug 2023 14:11:37	U	SMPL
CCB91 - 1	CCB91	174	-0.0252	-	7471B	PPB	09 Aug 2023 14:13:52	U	SMPL
CRA - 1	CRA	1432	0.1767	-	7471B	PPB	09 Aug 2023 14:16:07	U	SMPL
HighStd - 1	HighStd	62478	9.9735	-	7471B	PPB	09 Aug 2023 14:18:22	U	SMPL
ChkStd - 1	ChkStd	45428	7.2373	-	7471B	PPB	09 Aug 2023 14:20:37	U	SMPL
PB154727BL - 1	PBS	22	-0.0496	-	7471B	PPB	09 Aug 2023 14:22:54	U	SMPL
PB154727BS - 1	LCSS	21664	3.4236	-	7471B	PPB	09 Aug 2023 14:25:12	U	SMPL
O3919-01 - 1	TP-38	1563	0.1977	-	7471B	PPB	09 Aug 2023 14:27:29	U	SMPL
O3919-01DUP - 1	TP-38DUP	1781	0.2327	-	7471B	PPB	09 Aug 2023 14:29:45	U	SMPL
O3919-01MS - 1	TP-38MS	24506	3.8796	-	7471B	PPB	09 Aug 2023 14:32:02	U	SMPL
O3919-01MSD - 1	TP-38MSD	24553	3.8872	-	7471B	PPB	09 Aug 2023 14:34:18	U	SMPL
O3919-04 - 1	TP-36	478	0.0236	-	7471B	PPB	09 Aug 2023 14:36:33	U	SMPL
CCV92 - 1	CCV92	33094	5.2579	-	7471B	PPB	09 Aug 2023 14:38:49	U	SMPL
CCB92 - 1	CCB92	198	-0.0214	-	7471B	PPB	09 Aug 2023 14:41:04	U	SMPL
O3920-01 - 1	OR-3-080723	2985	0.4259	-	7471B	PPB	09 Aug 2023 14:43:20	U	SMPL
O3924-01 - 1	1115	358	0.0043	-	7471B	PPB	09 Aug 2023 14:45:35	U	SMPL
O3927-01 - 1	1127	1583	0.2009	-	7471B	PPB	09 Aug 2023 14:47:50	U	SMPL
O3929-01 - 1	B-PP19	27378	4.3406	-	7471B	PPB	09 Aug 2023 14:50:05	U	SMPL
O3929-02 - 1	B-PP20	8219	1.2659	-	7471B	PPB	09 Aug 2023 14:52:21	U	SMPL
O3929-03 - 1	DUP-07	19164	3.0224	-	7471B	PPB	09 Aug 2023 14:54:38	U	SMPL
O3932-01 - 1	NEVINS-SB-01-Z18-20	522	0.0306	-	7471B	PPB	09 Aug 2023 14:56:55	U	SMPL
O3932-02 - 1	NEVINS-SB-01-Z20-22	358	0.0043	-	7471B	PPB	09 Aug 2023 14:59:18	U	SMPL
O3932-03 - 1	NEVINS-SB-01-Z30-32	178	-0.0246	-	7471B	PPB	09 Aug 2023 15:01:38	U	SMPL
O3932-04 - 1	NEVINS-SB-01-Z67-69	458	0.0204	-	7471B	PPB	09 Aug 2023 15:03:53	U	SMPL
CCV93 - 1	CCV93	33497	5.3225	-	7471B	PPB	09 Aug 2023 15:06:08	U	SMPL
CCB93 - 1	CCB93	217	-0.0183	-	7471B	PPB	09 Aug 2023 15:08:24	U	SMPL
O3919-01LX5 - 1		224	-0.0172	-	7471B	PPB	09 Aug 2023 15:10:39	U	SMPL
O3919-01A - 1		26596	4.2151	-	7471B	PPB	09 Aug 2023 15:12:54	U	SMPL
CCV94 - 1	CCV94	33138	5.2649	-	7471B	PPB	09 Aug 2023 15:15:10	U	SMPL
CCB94 - 1	CCB94	190	-0.0226	-	7471B	PPB	09 Aug 2023 15:17:25	U	SMPL