

CB126802


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

5097E



A= 0.0000e+000
 B= 1.9686e-004 slope
 C= -2.0658e-002 y-intercept
 Rho= 0.9999753
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%I
0.00	0.000	0.016	0.016	185	0.000	185					
0.05	0.050					504	0				
0.20	0.200	0.191	-0.009	1077	0.0 %	1077					
2.50	2.500	2.478	-0.022	12690	0.0 %	12690					
5.00	5.000	5.039	0.039	25702	0.0 %	25702					
7.50	7.500	7.462	-0.038	38008	0.0 %	38008					
10.0	10.000	10.015	0.015	50976	0.0 %	50976					

0.1-1.5%

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	185	-		0 SFAM01.1	PPB	8/9/2023 15:48	Std
	0.2 S01	1077	-		0.2 SFAM01.1	PPB	8/9/2023 15:50	Std
	2.5 S02	12690	-		2.5 SFAM01.1	PPB	8/9/2023 15:52	Std
	5 S03	25702	-		5 SFAM01.1	PPB	8/9/2023 15:54	Std
	7.5 S04	38008	-		7.5 SFAM01.1	PPB	8/9/2023 15:57	Std
	10 S05	50976	-		10 SFAM01.1	PPB	8/9/2023 15:59	Std
ICV011	ICV011	19953	3.9073 -		SFAM01.1	PPB	8/9/2023 16:05	SMPL
ICB011	ICB011	142	0.0073 -		SFAM01.1	PPB	8/9/2023 16:08	SMPL
CCV095	CCV095	25685	5.0357 -		SFAM01.1	PPB	8/9/2023 16:10	SMPL
CCB095	CCB095	125	0.0039 -		SFAM01.1	PPB	8/9/2023 16:12	SMPL
PB154733BL	PBW733	105	0 -		SFAM01.1	PPB	8/9/2023 16:14	SMPL
O3803-01	GCB99	73	-0.0063 -		SFAM01.1	PPB	8/9/2023 16:17	SMPL
O3803-02	GCB99D	93	-0.0023 -		SFAM01.1	PPB	8/9/2023 16:19	SMPL
O3803-03	GCB99S	5751	1.1115 -		SFAM01.1	PPB	8/9/2023 16:21	SMPL
O3803-04	GCBA0	32485	6.3744 -		SFAM01.1	PPB	8/9/2023 16:32	SMPL
O3803-05	GCBA1	112	0.0014 -		SFAM01.1	PPB	8/9/2023 16:34	SMPL
O3803-06	GCBA2	100	-0.001 -		SFAM01.1	PPB	8/9/2023 16:37	SMPL
O3803-07	GCBA3	90	-0.0029 -		SFAM01.1	PPB	8/9/2023 16:39	SMPL
O3803-08	GCBA4	150	0.0089 -		SFAM01.1	PPB	8/9/2023 16:41	SMPL
O3803-09	GCBA5	119	0.0028 -		SFAM01.1	PPB	8/9/2023 16:43	SMPL
O3803-10	GCBA6	214	0.0215 -		SFAM01.1	PPB	8/9/2023 16:46	SMPL
O3803-11	GCBA7	143	0.0075 -		SFAM01.1	PPB	8/9/2023 16:48	SMPL
CCV096	CCV096	25540	5.0072 -		SFAM01.1	PPB	8/9/2023 16:50	SMPL
CCB096	CCB096	84	-0.0041 -		SFAM01.1	PPB	8/9/2023 16:52	SMPL
PB154734BL	PBW734	210	0.0207 -		SFAM01.1	PPB	8/9/2023 16:55	SMPL
O3804-01	GCB84	104	-0.0002 -		SFAM01.1	PPB	8/9/2023 16:57	SMPL
O3804-02	GCB84D	138	0.0065 -		SFAM01.1	PPB	8/9/2023 16:59	SMPL
O3804-03	GCB84S	5344	1.0314 -		SFAM01.1	PPB	8/9/2023 17:02	SMPL
O3804-04	GCB85	323	0.0429 -		SFAM01.1	PPB	8/9/2023 17:04	SMPL
O3804-05	GCB86	325	0.0433 -		SFAM01.1	PPB	8/9/2023 17:06	SMPL
O3804-06	GCB87	389	0.0559 -		SFAM01.1	PPB	8/9/2023 17:08	SMPL
O3804-07	GCB88	149	0.0087 -		SFAM01.1	PPB	8/9/2023 17:11	SMPL
O3804-08	GCB89	254	0.0293 -		SFAM01.1	PPB	8/9/2023 17:13	SMPL
O3804-09	GCBC0	275	0.0335 -		SFAM01.1	PPB	8/9/2023 17:15	SMPL
O3804-10	GCBC1	172	0.0132 -		SFAM01.1	PPB	8/9/2023 17:17	SMPL
PB154735BL	PBW735	90	-0.0029 -		SFAM01.1	PPB	8/9/2023 17:20	SMPL
O3825-01	MEYF85	77	-0.0055 -		SFAM01.1	PPB	8/9/2023 17:22	SMPL
O3825-03	MEYF87	59	-0.009 -		SFAM01.1	PPB	8/9/2023 17:24	SMPL
O3825-04	MEYF87D	101	-0.0008 -		SFAM01.1	PPB	8/9/2023 17:26	SMPL
O3825-05	MEYF87S	5410	1.0444 -		SFAM01.1	PPB	8/9/2023 17:29	SMPL
CCV097	CCV097	25593	5.0176 -		SFAM01.1	PPB	8/9/2023 17:31	SMPL
CCB097	CCB097	129	0.0047 -		SFAM01.1	PPB	8/9/2023 17:34	SMPL