



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126808

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	08/10/2023	08:20
2	CAL2	1	Water	NA	NA	20.6	7.00	08/10/2023	08:21
3	CAL3	1	Water	NA	NA	20.6	10.02	08/10/2023	08:24
4	ICV	1	Water	NA	NA	20.7	6.98	08/10/2023	08:25
5	CCV1	1	Water	NA	NA	20.7	2.01	08/10/2023	08:30
6	O3947-03	1	Water	NA	NA	20.7	6.51	08/10/2023	08:33
7	O3947-03DUP	1	Water	NA	NA	20.8	6.52	08/10/2023	08:34
8	O3955-22	1	Water	NA	NA	22.7	6.73	08/10/2023	08:40
9	CCV2	1	Water	NA	NA	20.7	12.02	08/10/2023	08:45