

LB126810


7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

43177



A= 0.0000e+000

B= 2.3265e-004

C= -4.1828e-002

Rho= 0.9998426

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O.D
0.0	0.000	0.017	0.017	252	0.000	252					0.0
0.2	0.200	0.222	0.022	1133	0.0 %	1133	0				11
2.5	2.500	2.518	0.018	11004	0.0 %	11004					1
5.0	5.000	4.863	-0.137	21083	0.0 %	21083					3
7.5	7.500	7.577	0.077	32747	0.0 %	32747					1
10.0	10.000	10.003	0.003	43177	0.0 %	43177					0

LB126810 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Aug 2023 11:31:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	252	-	0.0000	7470A	PPB	10 Aug 2023 11:39:06	S	Std
0.2 - 1	50.2	1133	-	0.2000	7470A	PPB	10 Aug 2023 11:41:23	S	Std
2.5 - 1	52.5	11004	-	2.5000	7470A	PPB	10 Aug 2023 11:43:39	S	Std
5.0 - 1	55	21083	-	5.0000	7470A	PPB	10 Aug 2023 11:45:56	S	Std
7.5 - 1	57.5	32747	-	7.5000	7470A	PPB	10 Aug 2023 11:50:45	S	Std
10.0 - 1	510	43177	-	10.0000	7470A	PPB	10 Aug 2023 11:55:32	S	Std
ICV12 - 1	ICV12	17887	4.1196	-	7470A	PPB	10 Aug 2023 11:58:21	U	SMPL
ICB12 - 1	ICB12	210	0.0070	-	7470A	PPB	10 Aug 2023 12:00:36	U	SMPL
CCV01 - 1	CCV01	23591	5.4466	-	7470A	PPB	10 Aug 2023 12:02:51	U	SMPL
CCB01 - 1	CCB01	191	0.0026	-	7470A	PPB	10 Aug 2023 12:05:06	U	SMPL
CRA - 1	CRA	1055	0.2036	-	7470A	PPB	10 Aug 2023 12:07:21	U	SMPL
HighStd - 1	HighStd	42737	9.9009	-	7470A	PPB	10 Aug 2023 12:09:36	U	SMPL
ChkStd - 1	ChkStd	30242	6.9940	-	7470A	PPB	10 Aug 2023 12:11:51	U	SMPL
PB154738BL - 1	PBW	161	-0.0044	-	7470A	PPB	10 Aug 2023 12:14:07	U	SMPL
PB154738BS - 1	LCSW	17669	4.0689	-	7470A	PPB	10 Aug 2023 12:16:22	U	SMPL
O3905-02 - 1	FLY-ASH	221	0.0096	-	7470A	PPB	10 Aug 2023 12:18:38	U	SMPL
O3919-02 - 1	TP-38	324	0.0336	-	7470A	PPB	10 Aug 2023 12:20:55	U	SMPL
O3919-05 - 1	TP-36	311	0.0305	-	7470A	PPB	10 Aug 2023 12:23:11	U	SMPL
O3921-01 - 1	WC-1	306	0.0294	-	7470A	PPB	10 Aug 2023 12:25:28	U	SMPL
O3925-01 - 1	SEW-30103	303	0.0287	-	7470A	PPB	10 Aug 2023 12:27:43	U	SMPL
CCV02 - 1	CCV02	23324	5.3845	-	7470A	PPB	10 Aug 2023 12:29:58	U	SMPL
CCB02 - 1	CCB02	234	0.0126	-	7470A	PPB	10 Aug 2023 12:32:13	U	SMPL
O3931-02 - 1	DAYTON-SOIL1	321	0.0329	-	7470A	PPB	10 Aug 2023 12:34:28	U	SMPL
O3931-04 - 1	DAYTON-SOIL2	290	0.0256	-	7470A	PPB	10 Aug 2023 12:36:44	U	SMPL
O3942-01 - 1	CENTRAL-PAINT-CHIPS	362	0.0424	-	7470A	PPB	10 Aug 2023 12:38:59	U	SMPL
O3942-01DUP - 1	CENTRAL-PAINT-CHIPSDUP	380	0.0466	-	7470A	PPB	10 Aug 2023 12:41:15	U	SMPL
O3942-01MS - 1	CENTRAL-PAINT-CHIPSMS	18148	4.1803	-	7470A	PPB	10 Aug 2023 12:43:30	U	SMPL
O3942-01MSD - 1	CENTRAL-PAINT-CHIPSMSD	18383	4.2350	-	7470A	PPB	10 Aug 2023 12:45:46	U	SMPL
PB154748BL - 1	PBW	278	0.0228	-	7470A	PPB	10 Aug 2023 12:48:03	U	SMPL
PB154748BS - 1	LCSW	17097	3.9358	-	7470A	PPB	10 Aug 2023 12:50:19	U	SMPL
O3931-05 - 1	DAYTON-TOTES	608	0.0996	-	7470A	PPB	10 Aug 2023 12:52:33	U	SMPL
O3931-05DUP - 1	DAYTON-TOTESDUP	609	0.0999	-	7470A	PPB	10 Aug 2023 12:54:48	U	SMPL
CCV03 - 1	CCV03	23136	5.3408	-	7470A	PPB	10 Aug 2023 13:01:53	U	SMPL
CCB03 - 1	CCB03	197	0.0040	-	7470A	PPB	10 Aug 2023 13:04:08	U	SMPL
O3931-05MS - 1	DAYTON-TOTESMS	18663	4.3001	-	7470A	PPB	10 Aug 2023 13:06:24	U	SMPL
O3955-22 - 1	TW-1	2769	0.6024	-	7470A	PPB	10 Aug 2023 13:10:54	U	SMPL
O3957-01 - 1	MW-17-0809	319	0.0324	-	7470A	PPB	10 Aug 2023 13:13:09	U	SMPL
O3957-02 - 1	MW-15-0809	370	0.0443	-	7470A	PPB	10 Aug 2023 13:15:25	U	SMPL
O3957-03 - 1	MW-16-0809	286	0.0247	-	7470A	PPB	10 Aug 2023 13:17:42	U	SMPL
O3957-04 - 1	MW-14-0809	460	0.0652	-	7470A	PPB	10 Aug 2023 13:19:58	U	SMPL
CCV04 - 1	CCV04	23248	5.3668	-	7470A	PPB	10 Aug 2023 13:22:14	U	SMPL
CCB04 - 1	CCB04	122	-0.0134	-	7470A	PPB	10 Aug 2023 13:24:31	U	SMPL
PB154703TB - 1	PB154703TB	339	0.0370	-	7470A	PPB	10 Aug 2023 13:26:47	U	SMPL
O3942-01LX5 - 1		79	-0.0234	-	7470A	PPB	10 Aug 2023 13:29:02	U	SMPL
O3942-01A - 1		17649	4.0642	-	7470A	PPB	10 Aug 2023 13:31:17	U	SMPL
O3931-05LX5 - 1		102	-0.0181	-	7470A	PPB	10 Aug 2023 13:33:33	U	SMPL
O3931-05A - 1		20417	4.7082	-	7470A	PPB	10 Aug 2023 13:35:48	U	SMPL
O3931-05MSD - 1	DAYTON-TOTESMSD	18877	4.3499	-	7470A	PPB	10 Aug 2023 13:38:03	U	SMPL
CCV05 - 1	CCV05	22293	5.1446	-	7470A	PPB	10 Aug 2023 13:40:19	U	SMPL
CCB05 - 1	CCB05	182	0.0005	-	7470A	PPB	10 Aug 2023 13:42:35	U	SMPL