



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126819

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	08/10/2023	13:30
2	O3931-05		1	12	100.00	>212.0	>212.0	08/10/2023	14:00
3	O3931-05DUP		1	11	100.00	>212.0	>212.0	08/10/2023	14:30
4	O3947-03		1	13	100.00	>212.0	>212.0	08/10/2023	15:00
5	O3955-22		1	12	100.00	>212.0	>212.0	08/10/2023	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6126819

WorkList Name : fp-8-9

WorkList ID : 172560

Department : Wet-Chemistry

Date : 08-09-2023 07:50:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3931-05	DAYTON-TOTES	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/08/2023	1010B
O3947-03	NICH-OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/09/2023	1010B
O3955-22	TW-1	Water	Flash Point	Cool 4 deg C	PSEG03	L11	08/09/2023	1010B

Date/Time 08/10/2023 13:25
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: ZP(CWC)

Date/Time 08/10/2023 15:35
 Raw Sample Received by: ZP(CWC)
 Raw Sample Relinquished by: RM CWC