

LB126821


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57483



A= 0.0000e+000

B= 1.7428e-004

C= -8.6293e-003

Rho= 0.9999845

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.039	0.039	272	0.000	272					9.0
0.2	0.200	0.177	-0.023	1063	0.0 %	1063					12
2.5	2.500	2.490	-0.010	14335	0.0 %	14335					0
5.0	5.000	4.983	-0.017	28642	0.0 %	28642					0
7.5	7.500	7.503	0.003	43099	0.0 %	43099					0
10.0	10.000	10.009	0.009	57483	0.0 %	57483					0

LB126821 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Aug 2023 14:31:25

Sample ID	Extended ID	µAbs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	272	-	0.000	7471B	PPB	10 Aug 2023 14:37:22	S	Std
0.2 - 1	50.2	1063	-	0.200	7471B	PPB	10 Aug 2023 14:39:39	S	Std
2.5 - 1	52.5	14335	-	2.500	7471B	PPB	10 Aug 2023 14:41:56	S	Std
5.0 - 1	55	28642	-	5.000	7471B	PPB	10 Aug 2023 14:44:13	S	Std
7.5 - 1	57.5	43099	-	7.500	7471B	PPB	10 Aug 2023 14:46:30	S	Std
10.0 - 1	57.5	57483	-	10.000	7471B	PPB	10 Aug 2023 14:48:48	S	Std
ICV13 - 1	ICV13	22932	3.9879	-	7471B	PPB	10 Aug 2023 14:54:06	U	SMPL
ICB13 - 1	ICB13	290	0.0430	-	7471B	PPB	10 Aug 2023 14:56:21	U	SMPL
CCV06 - 1	CCV06	28491	4.9567	-	7471B	PPB	10 Aug 2023 14:58:36	U	SMPL
CCB06 - 1	CCB06	268	0.0381	-	7471B	PPB	10 Aug 2023 15:00:51	U	SMPL
CRA - 1	CRA	1059	0.1759	-	7471B	PPB	10 Aug 2023 15:03:06	U	SMPL
HighStd - 1	HighStd	58324	10.1558	-	7471B	PPB	10 Aug 2023 15:05:21	U	SMPL
ChkStd - 1	ChkStd	41335	7.1951	-	7471B	PPB	10 Aug 2023 15:07:35	U	SMPL
PB154749BL - 1	PBS	-93	-0.0248	-	7471B	PPB	10 Aug 2023 15:09:53	U	SMPL
PB154749BS - 1	LCSS	23106	4.0162	-	7471B	PPB	10 Aug 2023 15:12:06	U	SMPL
O3930-01 - 1	VNJ-207	2504	0.4278	-	7471B	PPB	10 Aug 2023 15:19:32	U	SMPL
O3930-01DUP - 1	VNJ-207DUP	2407	0.4109	-	7471B	PPB	10 Aug 2023 15:21:48	U	SMPL
O3930-01MS - 1	VNJ-207MS	26470	4.6045	-	7471B	PPB	10 Aug 2023 15:24:05	U	SMPL
O3930-01MSD - 1	VNJ-207MSD	26270	4.5690	-	7471B	PPB	10 Aug 2023 15:26:22	U	SMPL
O3930-02 - 1	72-11935	4905	0.8462	-	7471B	PPB	10 Aug 2023 15:28:38	U	SMPL
CCV07 - 1	CCV07	28907	5.0292	-	7471B	PPB	10 Aug 2023 15:30:53	U	SMPL
CCB07 - 1	CCB07	315	0.0463	-	7471B	PPB	10 Aug 2023 15:33:08	U	SMPL
O3931-01 - 1	DAYTON-SOIL1	1513	0.2551	-	7471B	PPB	10 Aug 2023 15:35:23	U	SMPL
O3931-03 - 1	DAYTON-SOIL2	2393	0.4064	-	7471B	PPB	10 Aug 2023 15:37:39	U	SMPL
O3939-01 - 1	ORA-1745-1746-1747	1240	0.2075	-	7471B	PPB	10 Aug 2023 15:39:54	U	SMPL
O3939-03 - 1	ORA-1748-1719-1750-1751-1752-1753-1754	41313	7.1913	-	7471B	PPB	10 Aug 2023 15:42:09	U	SMPL
O3939-06 - 1	ORA-1748-1719-1750-1751-1752-1753-1754	4947	0.8535	-	7471B	PPB	10 Aug 2023 15:44:25	U	SMPL
O3941-01 - 1	DEAN-OILY-DEBRIS	629	0.1010	-	7471B	PPB	10 Aug 2023 15:46:41	U	SMPL
O3941-03 - 1	DEAN-OILY-STONE	45	-0.0008	-	7471B	PPB	10 Aug 2023 15:48:56	U	SMPL
O3943-01 - 1	B-PP21A	7923	1.3722	-	7471B	PPB	10 Aug 2023 15:51:14	U	SMPL
O3943-02 - 1	B-PP21B	1173	0.1958	-	7471B	PPB	10 Aug 2023 15:53:31	U	SMPL
O3945-01 - 1	NEVINS-SB01-Z88-90	287	0.0414	-	7471B	PPB	10 Aug 2023 15:55:46	U	SMPL
CCV08 - 1	CCV08	28515	4.9609	-	7471B	PPB	10 Aug 2023 15:58:02	U	SMPL
CCB08 - 1	CCB08	315	0.0463	-	7471B	PPB	10 Aug 2023 16:00:17	U	SMPL
O3945-02 - 1	NEVINS-SB01-Z78-80	203	0.0267	-	7471B	PPB	10 Aug 2023 16:02:33	U	SMPL
O3930-01LX5 - 1		414	0.0635	-	7471B	PPB	10 Aug 2023 16:04:48	U	SMPL
O3930-01A - 1		25133	4.3715	-	7471B	PPB	10 Aug 2023 16:07:03	U	SMPL
CCV09 - 1	CCV09	26466	4.6038	-	7471B	PPB	10 Aug 2023 16:09:18	U	SMPL
CCB09 - 1	CCB09	160	0.0193	-	7471B	PPB	10 Aug 2023 16:11:34	U	SMPL