

LB126830


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46835



A= 0.0000e+000

B= 2.1473e-004

C= -4.1931e-003

Rho= 0.9999237

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.036	0.036	189	0.000	189					
0.2	0.200	0.187	-0.013	892	0.0 %	892					-7
2.5	2.500	2.478	-0.022	11561	0.0 %	11561					-1
5.0	5.000	5.029	0.029	23439	0.0 %	23439					-1
7.5	7.500	7.417	-0.083	34560	0.0 %	34560					-1
10.0	10.000	10.053	0.053	46835	0.0 %	46835					1

LB126830 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Aug 2023 09:08:35

Sample ID	Extended ID	μ Abs.	Conc.	Strd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	189	-	0.0000	7470A	PPB	11 Aug 2023 10:16:54	S	Std
0.2 - 1	50.2	892	-	0.2000	7470A	PPB	11 Aug 2023 10:19:11	S	Std
2.5 - 1	52.5	11561	-	2.5000	7470A	PPB	11 Aug 2023 10:21:28	S	Std
5.0 - 1	55	23439	-	5.0000	7470A	PPB	11 Aug 2023 10:23:45	S	Std
7.5 - 1	57.5	34560	-	7.5000	7470A	PPB	11 Aug 2023 10:26:01	S	Std
10.0 - 1	510	46835	-	10.0000	7470A	PPB	11 Aug 2023 10:28:17	S	Std
ICV14 - 1	ICV14	18338	3.9335	-	7470A	PPB	11 Aug 2023 10:31:26	U	SMPL
ICB14 - 1	ICB14	99	0.0171	-	7470A	PPB	11 Aug 2023 10:33:40	U	SMPL
CCV10 - 1	CCV10	22803	4.8922	-	7470A	PPB	11 Aug 2023 10:35:55	U	SMPL
CCB10 - 1	CCB10	98	0.0169	-	7470A	PPB	11 Aug 2023 10:38:10	U	SMPL
CRA - 1	CRA	818	0.1715	-	7470A	PPB	11 Aug 2023 10:40:24	U	SMPL
HighStd - 1	HighStd	46180	9.9119	-	7470A	PPB	11 Aug 2023 10:42:39	U	SMPL
ChkStd - 1	ChkStd	30782	6.6055	-	7470A	PPB	11 Aug 2023 10:44:54	U	SMPL
PB154759BL - 1	PBW	-51	-0.0151	-	7470A	PPB	11 Aug 2023 10:47:10	U	SMPL
PB154759BS - 1	LCSW	19357	4.1523	-	7470A	PPB	11 Aug 2023 10:49:28	U	SMPL
O3944-01 - 1	IDW-S-20230808	82	0.0134	-	7470A	PPB	11 Aug 2023 10:51:44	U	SMPL
O3954-02 - 1	WC-15	225	0.0441	-	7470A	PPB	11 Aug 2023 10:54:00	U	SMPL
O3954-05 - 1	WC-16	141	0.0261	-	7470A	PPB	11 Aug 2023 10:56:16	U	SMPL
O3954-08 - 1	WC-17	161	0.0304	-	7470A	PPB	11 Aug 2023 10:58:33	U	SMPL
O3954-11 - 1	WC-18	168	0.0319	-	7470A	PPB	11 Aug 2023 11:00:48	U	SMPL
CCV11 - 1	CCV11	22684	4.8667	-	7470A	PPB	11 Aug 2023 11:03:03	U	SMPL
CCB11 - 1	CCB11	117	0.0209	-	7470A	PPB	11 Aug 2023 11:05:18	U	SMPL
O3954-14 - 1	WC-19	784	0.1642	-	7470A	PPB	11 Aug 2023 11:07:33	U	SMPL
O3954-17 - 1	WC-20	168	0.0319	-	7470A	PPB	11 Aug 2023 11:09:49	U	SMPL
O3954-20 - 1	WC-21	189	0.0364	-	7470A	PPB	11 Aug 2023 11:12:04	U	SMPL
O3955-02 - 1	WC-22	313	0.0630	-	7470A	PPB	11 Aug 2023 11:14:19	U	SMPL
O3955-05 - 1	WC-23	192	0.0370	-	7470A	PPB	11 Aug 2023 11:16:35	U	SMPL
O3955-08 - 1	WC-24	184	0.0353	-	7470A	PPB	11 Aug 2023 11:18:51	U	SMPL
O3955-11 - 1	WC-25	141	0.0261	-	7470A	PPB	11 Aug 2023 11:21:08	U	SMPL
O3955-14 - 1	WC-26	159	0.0299	-	7470A	PPB	11 Aug 2023 11:23:25	U	SMPL
O3955-17 - 1	WC-27	166	0.0315	-	7470A	PPB	11 Aug 2023 11:25:42	U	SMPL
O3955-20 - 1	WC-28	173	0.0330	-	7470A	PPB	11 Aug 2023 11:27:57	U	SMPL
CCV12 - 1	CCV12	22722	4.8748	-	7470A	PPB	11 Aug 2023 11:30:13	U	SMPL
CCB12 - 1	CCB12	119	0.0214	-	7470A	PPB	11 Aug 2023 11:32:28	U	SMPL
O3955-20DUP - 1	WC-28DUP	141	0.0261	-	7470A	PPB	11 Aug 2023 11:34:44	U	SMPL
O3955-20MS - 1	WC-28MS	17805	3.8190	-	7470A	PPB	11 Aug 2023 11:36:59	U	SMPL
O3955-20MSD - 1	WC-28MSD	18765	4.0251	-	7470A	PPB	11 Aug 2023 11:39:14	U	SMPL
PB154730TB - 1	PB154730TB	88	0.0147	-	7470A	PPB	11 Aug 2023 11:41:29	U	SMPL
O3955-20LX5 - 1		46	0.0057	-	7470A	PPB	11 Aug 2023 11:43:45	U	SMPL
O3955-20A - 1		18361	3.9384	-	7470A	PPB	11 Aug 2023 11:46:01	U	SMPL
CCV13 - 1	CCV13	22002	4.7202	-	7470A	PPB	11 Aug 2023 11:48:17	U	SMPL
CCB13 - 1	CCB13	102	0.0177	-	7470A	PPB	11 Aug 2023 11:50:34	U	SMPL