

LB126885


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

61866



A= 0.0000e+000

B= 1.6127e-004

C= -2.5265e-002

Rho= 0.9998850

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	315	0.000	315				
0.2	0.200	0.179	-0.021	1267	0.0 %	1267				
2.5	2.500	2.505	0.005	15690	0.0 %	15690				
5.0	5.000	4.934	-0.066	30750	0.0 %	30750				
7.5	7.500	7.605	0.105	47312	0.0 %	47312				
10.0	10.000	9.952	-0.048	61866	0.0 %	61866				

yes
-11
-10
-10
-10

LB126835 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Aug 2023 12:53:02

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	315	-	0.0007471B	PPB		11 Aug 2023 13:25:59	S	Std
0.2 - 1	50.2	1267	-	0.2007471B	PPB		11 Aug 2023 13:28:16	S	Std
2.5 - 1	52.5	15690	-	2.5007471B	PPB		11 Aug 2023 13:30:33	S	Std
5.0 - 1	55	30750	-	5.0007471B	PPB		11 Aug 2023 13:32:50	S	Std
7.5 - 1	57.5	47312	-	7.5007471B	PPB		11 Aug 2023 13:35:06	S	Std
10.0 - 1	510	61866	-	10.0007471B	PPB		11 Aug 2023 13:40:00	S	Std
ICV15 - 1	ICV15	26746	4.2881	-7471B	PPB		11 Aug 2023 13:42:48	U	SMPL
ICB15 - 1	ICB15	321	0.0265	-7471B	PPB		11 Aug 2023 13:45:03	U	SMPL
CCV14 - 1	CCV14	29904	4.7973	-7471B	PPB		11 Aug 2023 13:47:19	U	SMPL
CCB14 - 1	CCB14	328	0.0276	-7471B	PPB		11 Aug 2023 13:49:34	U	SMPL
CRA - 1	CRA	1198	0.1676	-7471B	PPB		11 Aug 2023 13:51:48	U	SMPL
HighStd - 1	HighStd	61028	9.8167	-7471B	PPB		11 Aug 2023 13:54:03	U	SMPL
ChkStd - 1	ChkStd	42648	6.8526	-7471B	PPB		11 Aug 2023 13:56:18	U	SMPL
PB154760BL - 1	PBS	76	-0.0130	-7471B	PPB		11 Aug 2023 14:01:15	U	SMPL
PB154760BS - 1	LCSS	22673	3.6312	-7471B	PPB		11 Aug 2023 14:03:30	U	SMPL
O3952-01 - 1	OK-1-080923	3718	0.5743	-7471B	PPB		11 Aug 2023 14:05:46	U	SMPL
O3954-01 - 1	WC-15	1629	0.2374	-7471B	PPB		11 Aug 2023 14:08:02	U	SMPL
O3954-01DUP - 1	WC-15DUP	1519	0.2197	-7471B	PPB		11 Aug 2023 14:10:18	U	SMPL
O3954-01MS - 1	WC-15MS	27262	4.3713	-7471B	PPB		11 Aug 2023 14:12:36	U	SMPL
O3954-01MSD - 1	WC-15MSD	26572	4.2600	-7471B	PPB		11 Aug 2023 14:14:51	U	SMPL
CCV15 - 1	CCV15	30451	4.8856	-7471B	PPB		11 Aug 2023 14:16:06	U	SMPL
CCB15 - 1	CCB15	373	0.0349	-7471B	PPB		11 Aug 2023 14:18:20	U	SMPL
O3954-04 - 1	WC-16	1560	0.2263	-7471B	PPB		11 Aug 2023 14:20:36	U	SMPL
O3954-07 - 1	WC-17	5609	0.8793	-7471B	PPB		11 Aug 2023 14:22:51	U	SMPL
O3954-10 - 1	WC-18	1665	0.2432	-7471B	PPB		11 Aug 2023 14:25:06	U	SMPL
O3954-13 - 1	WC-19	283710	45.7286	-7471B	PPB		11 Aug 2023 14:27:21	U	SMPL
O3954-16 - 1	WC-20	7817	1.2354	-7471B	PPB		11 Aug 2023 14:29:37	U	SMPL
O3954-19 - 1	WC-21	84295	13.5690	-7471B	PPB		11 Aug 2023 14:31:56	U	SMPL
O3955-01 - 1	WC-22	320903	51.7267	-7471B	PPB		11 Aug 2023 14:34:13	U	SMPL
O3955-04 - 1	WC-23	31474	5.0505	-7471B	PPB		11 Aug 2023 14:36:31	U	SMPL
O3955-07 - 1	WC-24	6974	1.0994	-7471B	PPB		11 Aug 2023 14:38:51	U	SMPL
O3955-10 - 1	WC-25	6145	0.9657	-7471B	PPB		11 Aug 2023 14:41:07	U	SMPL
CCV16 - 1	CCV16	30392	4.8760	-7471B	PPB		11 Aug 2023 14:43:22	U	SMPL
CCB16 - 1	CCB16	364	0.0334	-7471B	PPB		11 Aug 2023 14:45:37	U	SMPL
O3955-13 - 1	WC-26	1399	0.2004	-7471B	PPB		11 Aug 2023 14:47:53	U	SMPL
O3955-16 - 1	WC-27	33689	5.4077	-7471B	PPB		11 Aug 2023 14:50:09	U	SMPL
O3955-19 - 1	WC-28	12182	1.9393	-7471B	PPB		11 Aug 2023 14:52:24	U	SMPL
O3956-01 - 1	B-PP22	22386	3.5849	-7471B	PPB		11 Aug 2023 14:54:39	U	SMPL
O3956-02 - 1	B-PP23A	8529	1.3502	-7471B	PPB		11 Aug 2023 14:56:55	U	SMPL
O3956-03 - 1	B-PP23B	1808	0.2660	-7471B	PPB		11 Aug 2023 14:59:12	U	SMPL
PB154761BL - 1	PBS	161	0.0007	-7471B	PPB		11 Aug 2023 15:01:28	U	SMPL
PB154761BS - 1	LCSS	27104	4.3458	-7471B	PPB		11 Aug 2023 15:03:45	U	SMPL
O3958-01 - 1	NEVINS-SB02-Z106-107	2817	0.4290	-7471B	PPB		11 Aug 2023 15:06:02	U	SMPL
O3958-01DUP - 1	NEVINS-SB02-Z106-107DUP	2469	0.3729	-7471B	PPB		11 Aug 2023 15:08:17	U	SMPL
CCV17 - 1	CCV17	29752	4.7728	-7471B	PPB		11 Aug 2023 15:10:33	U	SMPL
CCB17 - 1	CCB17	367	0.0339	-7471B	PPB		11 Aug 2023 15:12:48	U	SMPL
O3958-01MS - 1	NEVINS-SB02-Z106-107MS	27634	4.4313	-7471B	PPB		11 Aug 2023 15:15:03	U	SMPL
O3958-01MSD - 1	NEVINS-SB02-Z106-107MSD	25942	4.1584	-7471B	PPB		11 Aug 2023 15:17:18	U	SMPL
O3958-02 - 1	NEVINS-SB02-Z93-94	1240	0.1747	-7471B	PPB		11 Aug 2023 15:19:34	U	SMPL
O3958-03 - 1	NEVINS-SB02-Z64-65	2550	0.3860	-7471B	PPB		11 Aug 2023 15:21:49	U	SMPL
O3958-04 - 1	NEVINS-SB02-Z46-47	2350	0.3537	-7471B	PPB		11 Aug 2023 15:24:05	U	SMPL
O3958-05 - 1	NEVINS-SB02-Z37-39	2190	0.3279	-7471B	PPB		11 Aug 2023 15:26:21	U	SMPL
O3958-06 - 1	NEVINS-SB02-Z21-23	2441	0.3684	-7471B	PPB		11 Aug 2023 15:28:37	U	SMPL
O3958-07 - 1	NEVINS-SB02-Z10-12	5604	0.8785	-7471B	PPB		11 Aug 2023 15:30:53	U	SMPL
O3961-01 - 1	ETGI-339	13031	2.0762	-7471B	PPB		11 Aug 2023 15:33:10	U	SMPL
O3961-03 - 1	20071	1502	0.2170	-7471B	PPB		11 Aug 2023 15:35:26	U	SMPL
CCV18 - 1	CCV18	30090	4.8273	-7471B	PPB		11 Aug 2023 15:37:41	U	SMPL
CCB18 - 1	CCB18	365	0.0336	-7471B	PPB		11 Aug 2023 15:39:57	U	SMPL
O3971-01 - 1	B-PP02A	9885	1.5689	-7471B	PPB		11 Aug 2023 15:42:12	U	SMPL
O3971-02 - 1	B-PP01A	3344	0.5140	-7471B	PPB		11 Aug 2023 15:44:27	U	SMPL
O3971-04 - 1	B-PP02B	2510	0.3795	-7471B	PPB		11 Aug 2023 15:46:43	U	SMPL
O3971-05 - 1	B-PP041A	2305	0.3465	-7471B	PPB		11 Aug 2023 15:48:58	U	SMPL
O3971-06 - 1	B-PP04B	1720	0.2521	-7471B	PPB		11 Aug 2023 15:51:15	U	SMPL
O3971-07 - 1	DUP-05	2490	0.3763	-7471B	PPB		11 Aug 2023 15:53:31	U	SMPL
O3971-08 - 1	B-PP03A	13305	2.1204	-7471B	PPB		11 Aug 2023 15:55:48	U	SMPL
O3971-09 - 1	B-PP08A	2250	0.3376	-7471B	PPB		11 Aug 2023 15:58:04	U	SMPL
O3971-10 - 1	B-PP08B	1582	0.2299	-7471B	PPB		11 Aug 2023 16:00:22	U	SMPL
O3972-01 - 1	367	9914	1.5736	-7471B	PPB		11 Aug 2023 16:02:38	U	SMPL
CCV19 - 1	CCV19	29928	4.8009	-7471B	PPB		11 Aug 2023 16:04:54	U	SMPL
CCB19 - 1	CCB19	366	0.0338	-7471B	PPB		11 Aug 2023 16:07:09	U	SMPL
O3972-03 - 1	356	3668	0.5663	-7471B	PPB		11 Aug 2023 16:09:25	U	SMPL
O3954-01LX5 - 1		247	0.0146	-7471B	PPB		11 Aug 2023 16:11:42	U	SMPL
O3954-01A - 1		26248	4.2077	-7471B	PPB		11 Aug 2023 16:13:57	U	SMPL
O3958-01LX5 - 1		531	0.0604	-7471B	PPB		11 Aug 2023 16:16:13	U	SMPL
O3958-01A - 1		26866	4.3074	-7471B	PPB		11 Aug 2023 16:18:29	U	SMPL
O3954-13DLX10 - 1		30707	4.9268	-7471B	PPB		11 Aug 2023 16:20:45	U	SMPL
O3954-19DLX5 - 1		18025	2.8816	-7471B	PPB		11 Aug 2023 16:23:02	U	SMPL
O3955-01DLX10 - 1		34424	5.5263	-7471B	PPB		11 Aug 2023 16:25:18	U	SMPL
CCV20 - 1	CCV20	30665	4.9201	-7471B	PPB		11 Aug 2023 16:27:35	U	SMPL
CCB20 - 1	CCB20	373	0.0349	-7471B	PPB		11 Aug 2023 16:29:52	U	SMPL