



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/14/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 08/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 08/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126839

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3967-01	A48K2	1	1.14	8.74	9.88	7.3	70.5	
O3967-02	A48K2MS	2	1.14	8.74	9.88	7.3	70.5	
O3967-03	A48K2MSD	3	1.14	8.74	9.88	7.3	70.5	
O3967-04	A48K4	4	1.15	8.81	9.96	8.17	79.7	
O3967-05	A48K5	5	1.18	8.61	9.79	7.82	77.1	
O3967-06	VHBLK001	6	1.00	1.00	2.00	2.00	100.0	VHBLK
O3967-09	A48L7	7	1.14	8.81	9.95	9.11	90.5	
O3967-10	A48L8	8	1.19	8.62	9.81	8.7	87.1	
O3967-11	A48L9	9	1.18	8.42	9.6	8.85	91.1	
O3967-12	A48M0	10	1.15	8.82	9.97	7.94	77.0	
O3967-13	A48M1	11	1.17	8.60	9.77	7.34	71.7	
O3967-14	A48M2	12	1.12	8.76	9.88	9.38	94.3	
O3967-15	A48M6	13	1.16	8.80	9.96	8.94	88.4	
O3967-16	A48X5	14	1.00	1.00	2.00	2.00	100.0	T.B.
O3967-17	A48M3	15	1.13	8.76	9.89	8.5	84.1	
O3967-18	A48M4	16	1.18	8.81	9.99	7.83	75.5	
O3967-19	A48M5	17	1.19	8.51	9.7	6.32	60.3	
O3967-20	A48M8	18	1.15	8.50	9.65	8.87	90.8	
O3967-21	A48M9	19	1.17	8.58	9.75	7.07	68.8	
O3967-22	A48N1	20	1.15	8.76	9.91	8.47	83.6	
O3967-23	A48N4	21	1.19	8.46	9.65	8.66	88.3	
O3967-24	A48N6	22	1.19	8.63	9.82	8.75	87.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$