

LB126846

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

63412

A= 0.0000e+000

B= 1.6012e-004

C= -7.4572e-002

Rho= 0.9996951

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/01)
0.0	0.000	-0.005	-0.005	435	0.000	435					7
0.2	0.200	0.215	0.015	1811	0.0 %	1811					1
2.5	2.500	2.481	-0.019	15960	0.0 %	15960					2
5.0	5.000	5.107	0.107	32359	0.0 %	32359					2
7.5	7.500	7.322	-0.178	46196	0.0 %	46196					1
10.0	10.000	10.079	0.079	63412	0.0 %	63412					

LB126846 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Aug 2023 12:32:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	435	-	0.0000	7471B	PPB	14 Aug 2023 13:03:20	S	Std
0.2 - 1	50.2	1811	-	0.2000	7471B	PPB	14 Aug 2023 13:05:37	S	Std
2.5 - 1	52.5	15960	-	2.5000	7471B	PPB	14 Aug 2023 13:07:54	S	Std
5.0 - 1	55	32359	-	5.0000	7471B	PPB	14 Aug 2023 13:10:11	S	Std
7.5 - 1	57.5	46196	-	7.5000	7471B	PPB	14 Aug 2023 13:12:27	S	Std
10.0 - 1	510	63412	-	10.0000	7471B	PPB	14 Aug 2023 13:14:42	S	Std
ICV16 - 1	ICV16	26422	4.1562	-	7471B	PPB	14 Aug 2023 13:17:32	U	SMPL
ICB16 - 1	ICB16	388	-0.0124	-	7471B	PPB	14 Aug 2023 13:19:47	U	SMPL
CCV21 - 1	CCV21	32232	5.0865	-	7471B	PPB	14 Aug 2023 13:22:01	U	SMPL
CCB21 - 1	CCB21	374	-0.0147	-	7471B	PPB	14 Aug 2023 13:24:21	U	SMPL
CRA - 1	CRA	1824	0.2175	-	7471B	PPB	14 Aug 2023 13:26:36	U	SMPL
HighStd - 1	HighStd	64075	10.1853	-	7471B	PPB	14 Aug 2023 13:28:51	U	SMPL
ChkStd - 1	ChkStd	43734	6.9283	-	7471B	PPB	14 Aug 2023 13:31:07	U	SMPL
PB154798BL - 1	PB154798BL	451	-0.0024	-	7471B	PPB	14 Aug 2023 13:33:23	U	SMPL
PB154798BS - 1	PB154798BS	22731	3.5652	-	7471B	PPB	14 Aug 2023 13:35:40	U	SMPL
O3974-01 - 1	EO-01-081023	3535	0.4915	-	7471B	PPB	14 Aug 2023 13:37:56	U	SMPL
O3975-01 - 1	NEVINS-SB03-Z10-12	2601	0.3419	-	7471B	PPB	14 Aug 2023 13:40:13	U	SMPL
O3975-02 - 1	NEVINS-SB03-Z10-12-Q	2057	0.2548	-	7471B	PPB	14 Aug 2023 13:42:29	U	SMPL
O3975-03 - 1	NEVINS-SB03-Z46-47	669	0.0326	-	7471B	PPB	14 Aug 2023 13:44:46	U	SMPL
O3975-04 - 1	NEVINS-SB03-Z82-83	1221	0.1209	-	7471B	PPB	14 Aug 2023 13:47:01	U	SMPL
CCV22 - 1	CCV22	32313	5.0995	-	7471B	PPB	14 Aug 2023 13:49:16	U	SMPL
CCB22 - 1	CCB22	424	-0.0067	-	7471B	PPB	14 Aug 2023 13:51:32	U	SMPL
O3975-05 - 1	NEVINS-SB03-Z96-98	1094	0.1006	-	7471B	PPB	14 Aug 2023 13:53:47	U	SMPL
O3975-05DUP - 1	NEVINS-SB03-Z96-98DUP	1079	0.0982	-	7471B	PPB	14 Aug 2023 13:56:01	U	SMPL
O3975-06 - 1	O3975-05MS	21020	3.2912	-	7471B	PPB	14 Aug 2023 13:58:17	U	SMPL
O3975-07 - 1	O3975-05MSD	22007	3.4493	-	7471B	PPB	14 Aug 2023 14:00:32	U	SMPL
O3989-01 - 1	B-PP25	9784	1.4921	-	7471B	PPB	14 Aug 2023 14:02:48	U	SMPL
O3989-02 - 1	B-PP24A	3476	0.4820	-	7471B	PPB	14 Aug 2023 14:05:04	U	SMPL
O3989-03 - 1	B-PP24B	616	0.0241	-	7471B	PPB	14 Aug 2023 14:07:21	U	SMPL
O3989-04 - 1	B-PP26	2220	0.2809	-	7471B	PPB	14 Aug 2023 14:09:38	U	SMPL
O3989-05 - 1	B-PP16	3161	0.4316	-	7471B	PPB	14 Aug 2023 14:11:55	U	SMPL
CCV23 - 1	CCV23	31968	5.0443	-	7471B	PPB	14 Aug 2023 14:14:10	U	SMPL
CCB23 - 1	CCB23	441	-0.0040	-	7471B	PPB	14 Aug 2023 14:16:25	U	SMPL
O3975-05LX5 - 1		172	-0.0470	-	7471B	PPB	14 Aug 2023 14:18:41	U	SMPL
O3975-05A - 1		26559	4.1781	-	7471B	PPB	14 Aug 2023 14:20:56	U	SMPL
CCV24 - 1	CCV24	32634	5.1509	-	7471B	PPB	14 Aug 2023 14:23:12	U	SMPL
CCB24 - 1	CCB24	365	-0.0161	-	7471B	PPB	14 Aug 2023 14:25:27	U	SMPL