



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/15/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:55
In Date: 08/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 08/15/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126847

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3968-01	A48H4	1	1.15	8.81	9.96	8.02	78.0	
O3968-02	A48H4MS	2	1.15	8.81	9.96	8.02	78.0	
O3968-03	A48H4MSD	3	1.15	8.81	9.96	8.02	78.0	
O3968-04	VHBLK001	4	1.00	1.00	2.00	2.00	100.0	VHBLK
O3968-05	A48N7	5	1.14	8.53	9.67	6.52	63.1	
O3968-06	A48X6	6	1.00	1.00	2.00	2.00	100.0	T.B.
O3968-23	A48N8	7	1.15	8.83	9.98	9.38	93.2	
O3968-24	A48P0	8	1.17	8.40	9.57	8.7	89.6	
O3968-25	A48P2	9	1.18	8.72	9.9	8.31	81.8	
O3968-26	A48P3	10	1.15	8.82	9.97	8.83	87.1	
O3968-27	A48P4	11	1.16	8.50	9.66	8.22	83.1	
O3968-28	A48P7	12	1.14	8.82	9.96	8.45	82.9	
O3968-29	A48Z3	13	1.00	1.00	2.00	2.00	100.0	T.B.
O3968-30	A48P8	14	1.12	8.77	9.89	8.64	85.7	
O3968-31	A48P9	15	1.17	8.60	9.77	7.08	68.7	
O3968-32	A48Q0	16	1.15	8.82	9.97	7.75	74.8	
O3968-33	A48Q1	17	1.14	8.75	9.89	8.65	85.8	
O3968-34	A48Q2	18	1.1	8.68	9.78	7.98	79.3	
O3968-35	A48Q3	19	1.15	8.37	9.52	7.95	81.2	
O3968-36	A48Z4	20	1.00	1.00	2.00	2.00	100.0	T.B.

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$