



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126848

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/14/2023	12:45
2	CAL2	1	Water	NA	NA	20.7	7.00	08/14/2023	12:46
3	CAL3	1	Water	NA	NA	20.7	10.03	08/14/2023	12:47
4	ICV	1	Water	NA	NA	20.7	7.01	08/14/2023	12:50
5	CCV1	1	Water	NA	NA	20.7	2.01	08/14/2023	13:00
6	O3944-01	1	Solid	20.02	20	20.2	9.44	08/14/2023	13:02
7	O3975-01	1	Solid	20.02	20	22.5	8.34	08/14/2023	13:05
8	O3975-02	1	Solid	20.03	20	22.3	8.40	08/14/2023	13:10
9	O3975-03	1	Solid	20.04	20	22.4	8.76	08/14/2023	13:15
10	O3975-04	1	Solid	20.02	20	22.6	9.43	08/14/2023	13:20
11	O3975-05	1	Solid	20.02	20	22.4	9.60	08/14/2023	13:22
12	O3981-01	1	Solid	20.03	20	20.7	8.26	08/14/2023	13:25
13	O3981-01DUP	1	Solid	20.04	20	20.8	8.27	08/14/2023	13:30
14	CCV2	1	Water	NA	NA	20.7	12.02	08/14/2023	13:33

VB 126848

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 03981		WorkList ID : 172688		Department : Wet-Chemistry		Date : 08-14-2023 12:31:10		
Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3944-01	IDW-S-20230808	Solid	pH	Cool 4 deg C	ECOL03	L11	08/08/2023	9045D
O3975-01	NEVINS-SB03-Z10-12	Solid	pH	Cool 4 deg C	ECOL03	N11	08/10/2023	9045D
O3975-02	NEVINS-SB03-Z10-12-Q	Solid	pH	Cool 4 deg C	ECOL03	N11	08/10/2023	9045D
O3975-03	NEVINS-SB03-Z46-47	Solid	pH	Cool 4 deg C	ECOL03	N11	08/10/2023	9045D
O3975-04	NEVINS-SB03-Z82-83	Solid	pH	Cool 4 deg C	ECOL03	N11	08/10/2023	9045D
O3975-05	NEVINS-SB03-Z96-98	Solid	pH	Cool 4 deg C	ECOL03	N11	08/10/2023	9045D
O3981-01	WC081023	Solid	pH	Cool 4 deg C	WALS01	N11	08/10/2023	9045D

Date/Time 08-14-23 12:40
Raw Sample Received by: JH/col
Raw Sample Relinquished by: CH SM

Date/Time 08-14-23 15:00
Raw Sample Received by: CH SM
Raw Sample Relinquished by: JH/col