



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126858

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/14/2023	09:10
2	O3969-06		1	8	41.00	105.8	105.8	08/14/2023	09:40
3	O3983-01		1	3	8.00	46.4	46.4	08/14/2023	10:10
4	O3983-01DUP		1	3	8.00	46.4	46.4	08/14/2023	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6126858

WorkList Name : FP S-8-11

WorkList ID : 172632

Department : Wet-Chemistry

Date : 08-11-2023 11:13:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3969-06 F	SOM-23-00053	Solid	Flash Point	Cool 4 deg C	PSEG03	N11	08/10/2023	1010B
O3983-01	1630	Solid	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B

Date/Time 08/11/2023 09:00
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: TA Leacy

Date/Time 08/11/2023 19:45
 Raw Sample Received by: JHalt
 Raw Sample Relinquished by: RM Cans