



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126859

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/14/2023	11:20
2	O3983-04		1	12	100.00	>212.0	>212.0	08/14/2023	11:50
3	O3983-04DUP		1	14	100.00	>212.0	>212.0	08/14/2023	12:20
4	O3983-05		1	12	100.00	>212.0	>212.0	08/14/2023	12:50
5	O3983-06		1	11	100.00	>212.0	>212.0	08/14/2023	13:20
6	O3983-07		1	12	100.00	>212.0	>212.0	08/14/2023	13:50
7	O3983-08		1	14	100.00	>212.0	>212.0	08/14/2023	14:20
8	O3983-09		1	14	100.00	>212.0	>212.0	08/14/2023	14:50
9	O3983-10		1	12	100.00	>212.0	>212.0	08/14/2023	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126859

WorkList Name : FP 8-11

WorkList ID : 172655

Department : Wet-Chemistry

Date : 08-11-2023 16:32:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3983-04 C	713-E	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3983-05 C	713-F	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3983-06 A	1624	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3983-07 A	1683	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3983-08 C	1729	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3983-09 C	1730	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3983-10 E	715	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B

Date/Time 08/11/2023 11:15
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: 70/Cole

Date/Time 08/11/2023 16:06
 Raw Sample Received by: JWC
 Raw Sample Relinquished by: RM CWC