

LB126863


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46553



A= 0.0000e+000

B= 2.1417e-004

C= 1.2361e-002

Rho= 0.9999355

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	52	0.000	52					
0.2	0.200	0.218	0.018	958	0.0 %	958					
2.5	2.500	2.459	-0.041	11424	0.0 %	11424					
5.0	5.000	4.947	-0.053	23041	0.0 %	23041					
7.5	7.500	7.570	0.070	35288	0.0 %	35288					
10.0	10.000	9.983	-0.017	46553	0.0 %	46553					

9
-2
-1
-0

LB126863 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Aug 2023 09:40:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	7470A	PPB	15 Aug 2023 09:49:15	S	Std
0.2 - 1	50.2	958	-	0.2000	7470A	PPB	15 Aug 2023 09:51:32	S	Std
2.5 - 1	52.5	11424	-	2.5000	7470A	PPB	15 Aug 2023 09:53:49	S	Std
5.0 - 1	55	23041	-	5.0000	7470A	PPB	15 Aug 2023 09:56:05	S	Std
7.5 - 1	57.5	35288	-	7.5000	7470A	PPB	15 Aug 2023 09:58:21	S	Std
10.0 - 1	58	46553	-	10.0000	7470A	PPB	15 Aug 2023 10:00:37	S	Std
ICV18 - 1	ICV18	18801	4.0390	-7470A	PPB	15 Aug 2023 10:03:21	U	SMPL	
ICB18 - 1	ICB18	-7	0.0109	-7470A	PPB	15 Aug 2023 10:05:36	U	SMPL	
CCV27 - 1	CCV27	23342	5.0116	-7470A	PPB	15 Aug 2023 10:07:50	U	SMPL	
CCB27 - 1	CCB27	-34	0.0051	-7470A	PPB	15 Aug 2023 10:10:05	U	SMPL	
CRA - 1	CRA	936	0.2128	-7470A	PPB	15 Aug 2023 10:12:20	U	SMPL	
HighStd - 1	HighStd	45778	9.8167	-7470A	PPB	15 Aug 2023 10:14:35	U	SMPL	
ChkStd - 1	ChkStd	31123	6.6780	-7470A	PPB	15 Aug 2023 10:19:28	U	SMPL	
PB154823BL - 1	PBW	-8	0.0106	-7470A	PPB	15 Aug 2023 10:21:43	U	SMPL	
PB154823BS - 1	LCSW	19255	4.1362	-7470A	PPB	15 Aug 2023 10:26:33	U	SMPL	
O3969-01 - 1	COMP-1	26	0.0179	-7470A	PPB	15 Aug 2023 10:28:49	U	SMPL	
O3969-02 - 1	COMP-2	63	0.0259	-7470A	PPB	15 Aug 2023 10:31:05	U	SMPL	
O3969-03 - 1	SOM-23-00072	83	0.0301	-7470A	PPB	15 Aug 2023 10:33:21	U	SMPL	
O3969-06 - 1	SOM-23-00053	1008	0.2282	-7470A	PPB	15 Aug 2023 10:35:38	U	SMPL	
O3971-03 - 1	WC-PP01	0	0.0124	-7470A	PPB	15 Aug 2023 10:37:53	U	SMPL	
CCV28 - 1	CCV28	23325	5.0079	-7470A	PPB	15 Aug 2023 10:40:11	U	SMPL	
CCB28 - 1	CCB28	-80	-0.0048	-7470A	PPB	15 Aug 2023 10:42:26	U	SMPL	
O3972-05 - 1	2876-2879	321	0.0811	-7470A	PPB	15 Aug 2023 10:44:42	U	SMPL	
O3973-02 - 1	SB-01-COMP	404	0.0989	-7470A	PPB	15 Aug 2023 10:46:57	U	SMPL	
O3973-04 - 1	SB-02-COMP	492	0.1177	-7470A	PPB	15 Aug 2023 10:49:12	U	SMPL	
O3973-06 - 1	SB-03-COMP	385	0.0948	-7470A	PPB	15 Aug 2023 10:51:27	U	SMPL	
O3973-08 - 1	SB-04-COMP	137	0.0417	-7470A	PPB	15 Aug 2023 10:53:43	U	SMPL	
O3973-10 - 1	SB-05-COMP	102	0.0342	-7470A	PPB	15 Aug 2023 10:55:59	U	SMPL	
O3973-12 - 1	SB-06-COMP	128	0.0398	-7470A	PPB	15 Aug 2023 10:58:15	U	SMPL	
O3973-14 - 1	SB-07-COMP	137	0.0417	-7470A	PPB	15 Aug 2023 11:00:32	U	SMPL	
O3981-02 - 1	WC081023	110	0.0359	-7470A	PPB	15 Aug 2023 11:02:49	U	SMPL	
O3985-02 - 1	PNNL-APM2-0811-1	176	0.0501	-7470A	PPB	15 Aug 2023 11:05:04	U	SMPL	
CCV29 - 1	CCV29	24046	5.1623	-7470A	PPB	15 Aug 2023 11:07:20	U	SMPL	
CCB29 - 1	CCB29	-1	0.0121	-7470A	PPB	15 Aug 2023 11:09:35	U	SMPL	
O3985-04 - 1	PNNL-APM2-0811-2	46	0.0222	-7470A	PPB	15 Aug 2023 11:11:51	U	SMPL	
O3985-04DUP - 1	PNNL-APM2-0811-2DUP	37	0.0203	-7470A	PPB	15 Aug 2023 11:14:06	U	SMPL	
O3985-04MS - 1	PNNL-APM2-0811-2MS	17583	3.7782	-7470A	PPB	15 Aug 2023 11:16:21	U	SMPL	
O3985-04MSD - 1	PNNL-APM2-0811-2MSD	16041	3.4479	-7470A	PPB	15 Aug 2023 11:18:37	U	SMPL	
PB154827BL - 1	PBW	-19	0.0083	-7470A	PPB	15 Aug 2023 11:20:53	U	SMPL	
PB154827BS - 1	LCSW	16946	3.6417	-7470A	PPB	15 Aug 2023 11:23:09	U	SMPL	
O3983-08 - 1	1729	1790	0.3957	-7470A	PPB	15 Aug 2023 11:25:25	U	SMPL	
O3983-09 - 1	1730	32	0.0192	-7470A	PPB	15 Aug 2023 11:27:42	U	SMPL	
O3983-09DUP - 1	1730DUP	45	0.0220	-7470A	PPB	15 Aug 2023 11:30:00	U	SMPL	
O3983-09MS - 1	1730MS	16846	3.6203	-7470A	PPB	15 Aug 2023 11:35:07	U	SMPL	
CCV30 - 1	CCV30	23644	5.0763	-7470A	PPB	15 Aug 2023 11:37:22	U	SMPL	
CCB30 - 1	CCB30	-7	0.0109	-7470A	PPB	15 Aug 2023 11:39:38	U	SMPL	
O3983-09MSD - 1	1730MSD	16561	3.5593	-7470A	PPB	15 Aug 2023 11:44:33	U	SMPL	
O3983-10 - 1	715	851	0.1946	-7470A	PPB	15 Aug 2023 11:46:48	U	SMPL	
O3996-01 - 1	EB-20230811	-48	0.0021	-7470A	PPB	15 Aug 2023 11:49:04	U	SMPL	
PB154763TB - 1	PB154763TB	41	0.0211	-7470A	PPB	15 Aug 2023 11:51:20	U	SMPL	
O3985-04LX5 - 1		-11	0.0100	-7470A	PPB	15 Aug 2023 11:53:35	U	SMPL	
O3985-04A - 1		19418	4.1712	-7470A	PPB	15 Aug 2023 11:55:51	U	SMPL	
O3983-09LX5 - 1		0	0.0124	-7470A	PPB	15 Aug 2023 11:58:07	U	SMPL	
O3983-09A - 1		21001	4.5102	-7470A	PPB	15 Aug 2023 12:00:24	U	SMPL	
CCV31 - 1	CCV31	21435	4.6031	-7470A	PPB	15 Aug 2023 12:02:41	U	SMPL	
CCB31 - 1	CCB31	-24	0.0072	-7470A	PPB	15 Aug 2023 12:04:56	U	SMPL	