



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126865

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.0 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	08/15/2023	08:00
2	CAL2	1	Water	NA	NA	20.6	7.00	08/15/2023	08:01
3	CAL3	1	Water	NA	NA	20.6	10.03	08/15/2023	08:02
4	ICV	1	Water	NA	NA	20.7	7.02	08/15/2023	08:10
5	CCV1	1	Water	NA	NA	20.6	2.01	08/15/2023	08:14
6	O4000-03	1	Water	NA	NA	21.5	7.03	08/15/2023	08:20
7	O4004-01	1	Water	NA	NA	21.5	4.88	08/15/2023	08:25
8	O4005-01	1	Water	NA	NA	21.3	4.80	08/15/2023	08:30
9	O4005-02	1	Water	NA	NA	21.2	5.03	08/15/2023	08:35
10	O4008-01	1	Water	NA	NA	22.4	9.07	08/15/2023	08:40
11	O4008-01DUP	1	Water	NA	NA	22.5	9.08	08/15/2023	08:41
12	CCV2	1	Water	NA	NA	20.6	12.02	08/15/2023	08:45

WORKLIST(Hardcopy Internal Chain)

VB 126865

WorkList Name : ph o4008

WorkList ID : 172699

Department : Wet-Chemistry

Date : 08-15-2023 07:38:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4000-03	COMP-1	Water	pH	Cool 4 deg C	PSEG03	N11	08/14/2023	9040C
O4004-01	GAMMA-JARINU	Water	pH	Cool 4 deg C	STER18	N11	08/07/2023	9040C
O4005-01	JS9600	Water	pH	Cool 4 deg C	STER18	N11	08/08/2023	9040C
O4005-02	JS7500	Water	pH	Cool 4 deg C	STER18	N11	08/08/2023	9040C
O4008-01	P1632	Water	pH	Cool 4 deg C	PSEG03	N11	08/14/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08-15-23 07:45

20(woc)

CP SM

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08-15-23 10:30

CP SM

20(woc)