

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
CCV1	0.8272	0.2451	29.63	TOC	
CCB1	0.2314	0.0194	8.37	TOC	
LB126866BLW	0.2506	0.0236	9.41	TOC	
LB126866BSW.....	9.6965...	0.6495..	6.70...	TOC	..
03869-01X10	12.4465	0.4012	3.22	TOC	
03869-01MSX10	13.3874	0.4309	3.22	TOC	
03869-01MSDX10.....	13.1982...	0.2870..	2.17...	TOC	..
CCV2	9.8304	0.4085	4.16	TOC	
CCB2	0.3198	0.0289	9.03	TOC	

Method ID		Sample Type	Vial		Timestamp	Message
TOC 0 - 20 ppmC		Sample	11	2023/08/15	10:22	
TOC 0 - 20 ppmC		Sample	12	2023/08/15	11:28	
TOC 0 - 20 ppmC		Sample	13	2023/08/15	11:51	
TOC 0 - 20 ppmC	...	Sample	.. 14..	2023/08/15	12:57	..
TOC 0 - 20 ppmC		Sample	15	2023/08/15	13:28	
TOC 0 - 20 ppmC		Sample	16	2023/08/15	13:58	
TOC 0 - 20 ppmC	...	Sample	.. 16..	2023/08/15	14:29	..
TOC 0 - 20 ppmC		Sample	11	2023/08/15	15:02	
TOC 0 - 20 ppmC		Sample	12	2023/08/15	15:49	

```
=====
Sample ID:   CCV1                      Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   08150959
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/08/15 10:22
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1.0361	0.5180	45438	-1.132	-0.933	135
2	0.7705	0.3853	33793	-1.008	-0.810	126
3	0.5043	0.2521	22114	-0.993	-0.793	119
4	0.9978	0.4989	43759	-1.104	-0.904	130

```
-----
<<<Statistics>>>   Mean:    0.8272   Std Dev:   0.2451   RSD: 29.63
=====
```

```
=====
Sample ID:   CCB1                      Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   08151106
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/08/15 11:28
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.2479	0.1239	10871	-0.979	-0.780	107
2	0.2464	0.1232	10806	-0.964	-0.765	108
3	0.2238	0.1119	9815	-0.918	-0.718	107
4	0.2075	0.1037	9099	-0.911	-0.712	110

```
-----
<<<Statistics>>>   Mean:    0.2314   Std Dev:   0.0194   RSD: 8.37
=====
```

```
=====
Sample ID:   LB126866BLW              Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   08151106
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/08/15 11:51
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.2827	0.1414	12399	-1.034	-0.836	108
2	0.2434	0.1217	10675	-1.008	-0.809	106
3	0.2500	0.1250	10966	-0.986	-0.787	112
4	0.2264	0.1132	9930	-0.974	-0.775	111

```
-----
<<<Statistics>>>   Mean:    0.2506   Std Dev:   0.0236   RSD: 9.41
=====
```

```
=====
Sample ID:   LB126866BSW              Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   08151227
Cal. Curve:  TOC WATER 0-20PPM        Timestamp: 2023/08/15 12:57
Operator ID: IZ MA                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	10.5498	5.2749	462671	-1.007	-0.809	232
2	9.3905	4.6952	411829	-0.982	-0.782	236
3	9.0395	4.5198	396437	-0.992	-0.793	250
4	9.8062	4.9031	430062	-0.992	-0.793	242

```
-----
<<<Statistics>>>   Mean:    9.6965   Std Dev:   0.6495   RSD: 6.70
=====
```

```
=====
Sample ID:   03869-01X10              Mode:      TOC
=====
```

Sample ID:	CCB2	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08151227
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2023/08/15 15:49
Operator ID:	IZ MA	Sample Type:	Sample

```
<<<Statistics>>>      Mean:      0.3198      Std Dev:      0.0289      RSD: 9.03
=====
```

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
0.0PPM	18992	1774	9.34	TOC	
0.5PPM	35626	280	0.79	TOC	
1.0PPM	54215	2353	4.34	TOC	
2.0PPM.....	96855...	1479..	1.53...	TOC	..
5.0PPM	238603	8685	3.64	TOC	
10.0PPM	464148	7495	1.61	TOC	
20.0PPM.....	887196...	8607..	0.97...	TOC	..
ICV1	10.8505	0.1776	1.64	TOC	
ICB1	0.3357	0.0506	15.08	TOC	
IC-20.....	0.2612...	0.0549..	21.01...	TOC	..
IC-R	0.2995	0.0616	20.56	TOC	

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2023/08/10 12:04	
TOC 0 - 20 ppmC	TOC Standard	2	2023/08/10 12:28	
TOC 0 - 20 ppmC	TOC Standard	3	2023/08/10 12:52	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2023/08/10 13:17	..
TOC 0 - 20 ppmC	TOC Standard	5	2023/08/10 13:45	
TOC 0 - 20 ppmC	TOC Standard	6	2023/08/10 14:13	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2023/08/10 14:44	..
TOC 0 - 20 ppmC	Sample	14	2023/08/10 15:13	
TOC 0 - 20 ppmC	Sample	9	2023/08/10 15:59	
TOC 0 - 20 ppmC	...Sample	.. 10..	2023/08/10 16:44	..
TOC 0 - 20 ppmC	Sample	10	2023/08/10 17:06	

```
=====
Sample ID:   0.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08101120
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/08/10 12:04
Operator ID: IZ MA                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			17310	-0.764	-0.564	111
2			18777	-0.740	-0.541	114
3			18397	-0.759	-0.560	114
4			21485	-0.778	-0.579	116

```
-----
<<<Statistics>>>   Mean:    18992   Std Dev:    1774   RSD: 9.34
=====
```

```
=====
Sample ID:   0.5PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08101120
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/08/10 12:28
Operator ID: IZ MA                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			35597	-0.824	-0.624	121
2			35875	-0.801	-0.602	123
3			35244	-0.771	-0.571	122
4			35787	-0.796	-0.597	121

```
-----
<<<Statistics>>>   Mean:    35626   Std Dev:     280   RSD: 0.79
=====
```

```
=====
Sample ID:   1.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08101120
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/08/10 12:52
Operator ID: IZ MA                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			54598	-0.858	-0.661	123
2			52571	-0.872	-0.673	128
3			52300	-0.825	-0.627	125
4			57391	-0.853	-0.654	131

```
-----
<<<Statistics>>>   Mean:    54215   Std Dev:    2353   RSD: 4.34
=====
```

```
=====
Sample ID:   2.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   08101120
Cal. Curve:  TOC WATER 0-20PPM          Timestamp: 2023/08/10 13:17
Operator ID: IZ MA                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			95957	-0.880	-0.681	136
2			96043	-0.843	-0.643	132
3			96361	-0.822	-0.623	133
4			99058	-0.826	-0.627	137

```
-----
<<<Statistics>>>   Mean:    96855   Std Dev:    1479   RSD: 1.53
=====
```

```
=====
Sample ID:   5.0PPM                      Mode:      TOC
=====
```

Sample ID:	ICB1	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	08101120
Cal. Curve:	TOC WATER 0-20PPM	Timestamp:	2023/08/10 15:59
Operator ID:	IZ MA	Sample Type:	Sample

```
<<<Statistics>>>      Mean:      0.3357      Std Dev:      0.0506      RSD: 15.08
```

```
Mode:          TOC
Filename:      08101120
Timestamp:     2023/08/10 16:44
Sample Type:   Sample
```

```
<<<Statistics>>>      Mean:      0.2612      Std Dev:      0.0549      RSD: 21.01
```

```
Mode:          TOC
Filename:      08101120
```

Sample Type: Sample

```
<<<Statistics>>>      Mean:      0.2995      Std Dev:      0.0616      RSD: 20.56
```

Cal. Curve ID: TOC WATER 0-20PPM
 Created: 2023/08/10 14:44
 Calibration Factor (m): 8.771e+04
 Y Intercept (b): 15311
 r-squared: 0.99963

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message	Date & Time
0.0PPM	18992	0.000	0.042	—	2023/08/10 12:04
0.5PPM	35626	0.250	0.232	-7.2	2023/08/10 12:28
1.0PPM	54215	0.500	0.444	-11.2	2023/08/10 12:52
2.0PPM	96855	1.000	0.930	-7.0	2023/08/10 13:17
5.0PPM	238603	2.500	2.546	1.8	2023/08/10 13:45
10.0PPM	464148	5.000	5.117	2.3	2023/08/10 14:13
20.0PPM	887196	10.000	9.940	-0.6	2023/08/10 14:44