



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126868

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	08/15/2023	10:45
2	O3986-01		1	14	100.00	>212.0	>212.0	08/15/2023	11:15
3	O3986-01DUP		1	14	100.00	>212.0	>212.0	08/15/2023	11:45
4	O3986-02		1	12	100.00	>212.0	>212.0	08/15/2023	12:15
5	O3986-03		1	14	100.00	>212.0	>212.0	08/15/2023	12:45
6	O3986-04		1	12	100.00	>212.0	>212.0	08/15/2023	13:15
7	O4000-03		1	14	100.00	>212.0	>212.0	08/15/2023	13:45
8	O4008-01		1	12	100.00	>212.0	>212.0	08/15/2023	14:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6126868

WorkList Name : FP W 8-15

WorkList ID : 172701

Department : Wet-Chemistry

Date : 08-15-2023 07:42:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3986-01 C	1770	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3986-02 C	1771	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3986-03 C	1773	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O3986-04 A	1772	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/11/2023	1010B
O4000-03 B	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/14/2023	1010B
O4008-01 F	P1632	Water	Flash Point	Cool 4 deg C	PSEG03	N11	08/14/2023	1010B

Date/Time 08/15/2023 10:00
 Raw Sample Received by: Rm CMC2
 Raw Sample Relinquished by: M. Carr

Date/Time 08/15/2023 14:50
 Raw Sample Received by: M. Carr
 Raw Sample Relinquished by: Rm CMC2