

LB126871


245.1

INSTRUMENT ID: CVJ

Linear

μ Abs.:

39033



A= 0.0000e+000

B= 2.5760e-004

C= -4.5307e-002

Rho= 0.9999500

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.010	-0.010	138	0.000	138					1
0.05	0.050	0.038	-0.012	322	0.0 %	322	0				-24
0.2	0.200	0.187	-0.013	900	0.0 %	900					-7
2.5	2.500	2.583	0.083	10202	0.0 %	10202					3
5.0	5.000	4.953	-0.047	19405	0.0 %	19405					1
7.5	7.500	7.490	-0.010	29251	0.0 %	29251					0
10.0	10.000	10.010	0.010	39033	0.0 %	39033					0

LB126871 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Aug 2023 14:14:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	138	-	0.0000	245.1	PPB	15 Aug 2023 15:32:53	S	Std
0.05 - 1	50.05	322	-	0.0500	245.1	PPB	15 Aug 2023 15:35:10	S	Std
0.2 - 1	50.2	900	-	0.2000	245.1	PPB	15 Aug 2023 15:37:27	S	Std
2.5 - 1	52.5	10202	-	2.5000	245.1	PPB	15 Aug 2023 15:39:44	S	Std
5.0 - 1	55	19405	-	5.0000	245.1	PPB	15 Aug 2023 15:42:00	S	Std
7.5 - 1	57.5	29251	-	7.5000	245.1	PPB	15 Aug 2023 15:44:16	S	Std
10.0 - 1	510	39033	-	10.0000	245.1	PPB	15 Aug 2023 15:46:32	S	Std
ICV19 - 1	ICV19	15249	3.8828	-245.1		PPB	15 Aug 2023 15:52:00	U	SMPL
ICB19 - 1	ICB19	127	-0.0126	-245.1		PPB	15 Aug 2023 15:54:15	U	SMPL
CCV32 - 1	CCV32	19707	5.0312	-245.1		PPB	15 Aug 2023 15:56:29	U	SMPL
CCB32 - 1	CCB32	131	-0.0116	-245.1		PPB	15 Aug 2023 15:58:45	U	SMPL
CRA - 1	CRA	322	0.0376	-245.1		PPB	15 Aug 2023 16:08:50	U	SMPL
HighStd - 1	HighStd	39625	10.1621	-245.1		PPB	15 Aug 2023 16:11:05	U	SMPL
ChkStd - 1	ChkStd	27059	6.9251	-245.1		PPB	15 Aug 2023 16:13:20	U	SMPL
PB154818BL - 1	PBW	81	-0.0244	-245.1		PPB	15 Aug 2023 16:15:37	U	SMPL
PB154818BS - 1	LCSW	16188	4.1247	-245.1		PPB	15 Aug 2023 16:17:53	U	SMPL
O3587-05 - 1	PT-HG-WS	45351	11.6372	-245.1		PPB	15 Aug 2023 16:28:49	U	SMPL
CCV33 - 1	CCV33	19558	4.9929	-245.1		PPB	15 Aug 2023 16:31:04	U	SMPL
CCB33 - 1	CCB33	85	-0.0234	-245.1		PPB	15 Aug 2023 16:33:20	U	SMPL
O3937-01 - 1	MANHOLE	132	-0.0113	-245.1		PPB	15 Aug 2023 16:35:35	U	SMPL
O3937-01DUP - 1	MANHOLEDUP	192	0.0042	-245.1		PPB	15 Aug 2023 16:37:51	U	SMPL
O3937-01MS - 1	MANHOLEMS	16378	4.1737	-245.1		PPB	15 Aug 2023 16:40:06	U	SMPL
O3937-01MSD - 1	MANHOLEMSD	16316	4.1577	-245.1		PPB	15 Aug 2023 16:43:31	U	SMPL
O3937-01LX5 - 1		-18	-0.0499	-245.1		PPB	15 Aug 2023 16:45:47	U	SMPL
O3937-01A - 1		16176	4.1216	-245.1		PPB	15 Aug 2023 16:48:03	U	SMPL
O3587-05DLX2 - 1	PT-HG-WS	22872	5.8465	-245.1		PPB	15 Aug 2023 16:56:48	U	SMPL
CCV34 - 1	CCV34	19745	5.0410	-245.1		PPB	15 Aug 2023 17:01:20	U	SMPL
CCB34 - 1	CCB34	42	-0.0345	-245.1		PPB	15 Aug 2023 17:03:35	U	SMPL