

LB126876


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

50319



A= 0.0000e+000

B= 1.9924e-004

C= -3.6012e-002

Rho= 0.9998896

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.017	-0.017	96	0.000	96	0				-
0.2	0.200	0.171	-0.029	1040	0.0 %	1040					-15
2.5	2.500	2.510	0.010	12779	0.0 %	12779					0
5.0	5.000	5.111	0.111	25835	0.0 %	25835					2
7.5	7.500	7.434	-0.066	37494	0.0 %	37494					-1
10.0	10.000	9.990	-0.010	50319	0.0 %	50319					0

LB126876 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Aug 2023 09:47:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	96	-	0.0000	7470A	PPB	16 Aug 2023 09:55:36	S	Std
0.2 - 1	50.2	1040	-	0.2000	7470A	PPB	16 Aug 2023 09:57:53	S	Std
2.5 - 1	52.5	12779	-	2.5000	7470A	PPB	16 Aug 2023 10:00:11	S	Std
5.0 - 1	55	25835	-	5.0000	7470A	PPB	16 Aug 2023 10:02:27	S	Std
7.5 - 1	53.5	37494	-	7.5000	7470A	PPB	16 Aug 2023 10:04:43	S	Std
10.0 - 1	500	50319	-	10.0000	7470A	PPB	16 Aug 2023 10:15:05	S	Std
ICV20 - 1	ICV20	19661	3.8813	-	7470A	PPB	16 Aug 2023 10:17:47	U	SMPL
ICB20 - 1	ICB20	-64	-0.0488	-	7470A	PPB	16 Aug 2023 10:20:01	U	SMPL
CCV35 - 1	CCV35	27452	5.4336	-	7470A	PPB	16 Aug 2023 10:29:23	U	SMPL
CCB35 - 1	CCB35	-29	-0.0418	-	7470A	PPB	16 Aug 2023 10:31:38	U	SMPL
CRA - 1	CRA	1157	0.1945	-	7470A	PPB	16 Aug 2023 10:37:08	U	SMPL
HighStd - 1	HighStd	49522	9.8309	-	7470A	PPB	16 Aug 2023 10:39:23	U	SMPL
ChkStd - 1	ChkStd	33210	6.5809	-	7470A	PPB	16 Aug 2023 10:41:39	U	SMPL
PB154849BL - 1	PBW	-96	-0.0551	-	7470A	PPB	16 Aug 2023 10:43:55	U	SMPL
PB154849BS - 1	LCSW	18828	3.7153	-	7470A	PPB	16 Aug 2023 10:48:46	U	SMPL
O3984-01 - 1	BX5-1-23	-27	-0.0414	-	7470A	PPB	16 Aug 2023 10:51:01	U	SMPL
O3984-02 - 1	BX5-2-23	266	0.0170	-	7470A	PPB	16 Aug 2023 10:53:18	U	SMPL
O3984-03 - 1	BX6-1-23	19	-0.0322	-	7470A	PPB	16 Aug 2023 10:55:35	U	SMPL
O3984-04 - 1	BX6-2-23	37	-0.0286	-	7470A	PPB	16 Aug 2023 10:57:52	U	SMPL
O3984-05 - 1	BX7-1-23	31	-0.0298	-	7470A	PPB	16 Aug 2023 11:00:07	U	SMPL
CCV36 - 1	CCV36	25768	5.0981	-	7470A	PPB	16 Aug 2023 11:04:55	U	SMPL
CCB36 - 1	CCB36	-49	-0.0458	-	7470A	PPB	16 Aug 2023 11:07:10	U	SMPL
O3984-06 - 1	BX7-2-23	20	-0.0320	-	7470A	PPB	16 Aug 2023 11:09:25	U	SMPL
O4000-01 - 1	COMP-2	24	-0.0312	-	7470A	PPB	16 Aug 2023 11:11:40	U	SMPL
O4000-02 - 1	5811	9	-0.0342	-	7470A	PPB	16 Aug 2023 11:13:56	U	SMPL
O4002-01 - 1	RB21147	52	-0.0257	-	7470A	PPB	16 Aug 2023 11:16:11	U	SMPL
O4006-03 - 1	E-1-EPH	34	-0.0292	-	7470A	PPB	16 Aug 2023 11:18:27	U	SMPL
O4009-02 - 1	CONC-COMP	29	-0.0302	-	7470A	PPB	16 Aug 2023 11:20:43	U	SMPL
O4009-02DUP - 1	CONC-COMPDUP	46	-0.0268	-	7470A	PPB	16 Aug 2023 11:23:00	U	SMPL
O4009-02MS - 1	CONC-COMPMS	20222	3.9931	-	7470A	PPB	16 Aug 2023 11:27:50	U	SMPL
O4009-02MSD - 1	CONC-COMPMSD	20367	4.0220	-	7470A	PPB	16 Aug 2023 11:30:07	U	SMPL
PB154850BL - 1	PBW	-50	-0.0460	-	7470A	PPB	16 Aug 2023 11:32:23	U	SMPL
CCV37 - 1	CCV37	25152	4.9754	-	7470A	PPB	16 Aug 2023 11:34:38	U	SMPL
CCB37 - 1	CCB37	-62	-0.0484	-	7470A	PPB	16 Aug 2023 11:36:53	U	SMPL
PB154850BS - 1	LCSW	18562	3.6623	-	7470A	PPB	16 Aug 2023 11:41:55	U	SMPL
O4000-03 - 1	COMP-1	-32	-0.0424	-	7470A	PPB	16 Aug 2023 11:44:10	U	SMPL
O4000-03DUP - 1	COMP-1DUP	26	-0.0308	-	7470A	PPB	16 Aug 2023 11:46:25	U	SMPL
O4000-03MS - 1	COMP-1MS	15560	3.0642	-	7470A	PPB	16 Aug 2023 11:48:41	U	SMPL
O4000-03MSD - 1	COMP-1MSD	15709	3.0939	-	7470A	PPB	16 Aug 2023 11:50:57	U	SMPL
PB154851BL - 1	PBW	-60	-0.0480	-	7470A	PPB	16 Aug 2023 11:53:13	U	SMPL
PB154851BS - 1	LCSW	18157	3.5816	-	7470A	PPB	16 Aug 2023 11:58:12	U	SMPL
O4008-01 - 1	P1632	-75	-0.0510	-	7470A	PPB	16 Aug 2023 12:00:29	U	SMPL
O4008-01DUP - 1	P1632DUP	-18	-0.0396	-	7470A	PPB	16 Aug 2023 12:02:46	U	SMPL
O4008-01MS - 1	P1632MS	17166	3.3842	-	7470A	PPB	16 Aug 2023 12:07:47	U	SMPL
CCV38 - 1	CCV38	24806	4.9064	-	7470A	PPB	16 Aug 2023 12:10:02	U	SMPL
CCB38 - 1	CCB38	-82	-0.0523	-	7470A	PPB	16 Aug 2023 12:12:17	U	SMPL
O4008-01MSD - 1	P1632MSD	17431	3.4370	-	7470A	PPB	16 Aug 2023 12:14:33	U	SMPL
O4018-01 - 1	ST-SUMPWATER	360	0.0357	-	7470A	PPB	16 Aug 2023 12:16:48	U	SMPL
PB154822TB - 1	PB154822TB	-19	-0.0398	-	7470A	PPB	16 Aug 2023 12:19:03	U	SMPL
PB154828TB - 1	PB154828TB	4	-0.0352	-	7470A	PPB	16 Aug 2023 12:21:18	U	SMPL
O4009-02LX5 - 1		2	-0.0356	-	7470A	PPB	16 Aug 2023 12:23:34	U	SMPL
O4009-02A - 1		22632	4.4733	-	7470A	PPB	16 Aug 2023 12:25:50	U	SMPL
O4000-03LX5 - 1		-39	-0.0438	-	7470A	PPB	16 Aug 2023 12:28:09	U	SMPL
O4000-03A - 1		17514	3.4535	-	7470A	PPB	16 Aug 2023 12:46:35	U	SMPL
O4008-01LX5 - 1		-321	-0.1000	-	7470A	PPB	16 Aug 2023 12:48:51	U	SMPL
O4008-01A - 1		17551	3.4609	-	7470A	PPB	16 Aug 2023 12:51:09	U	SMPL
CCV39 - 1	CCV39	23873	4.7205	-	7470A	PPB	16 Aug 2023 12:53:24	U	SMPL
CCB39 - 1	CCB39	-132	-0.0623	-	7470A	PPB	16 Aug 2023 12:55:41	U	SMPL